

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2012 SAS No.: Q2012 SDG NO.: Q2012

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/15/2025

Lab File ID: BN037025.D Time Analyzed: 09:55

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2015	7.618	5524	10.39	3171	14.27
UPPER LIMIT	4030	8.118	11048	10.893	6342	14.766
LOWER LIMIT	1007.5	7.118	2762	9.893	1585.5	13.766
EPA SAMPLE NO.						
01 RW5-SP100-20250509DL	2107	7.62	5709	10.39	3269	14.27

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.