

LAB CHRONICLE

OrderID:	Q2013	OrderDate:	5/12/2025 10:15:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2013-01	BP-TB-20250508	Water	VOCMS Group1	8260-Low	05/08/25		05/13/25	05/12/25
Q2013-02	BP-TT192D2-GW-202 50508	Water	VOCMS Group1	8260-Low	05/08/25		05/13/25	05/12/25
Q2013-03	BP-TT192D1-GW-202 50509	Water	VOCMS Group1	8260-Low	05/09/25		05/13/25	05/12/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2013
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:				0					
Total Voc :									
Total Concentration:									



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/08/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/12/25	
Client Sample ID:	BP-TB-20250508		SDG No.:	Q2013	
Lab Sample ID:	Q2013-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046162.D	1		05/13/25 13:38	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/08/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/12/25	
Client Sample ID:	BP-TB-20250508		SDG No.:	Q2013	
Lab Sample ID:	Q2013-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046162.D	1		05/13/25 13:38	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.2		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	50.6		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.6		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	61600	5.55				
540-36-3	1,4-Difluorobenzene	121000	6.757				
3114-55-4	Chlorobenzene-d5	115000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	45200	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/08/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/12/25	
Client Sample ID:	BP-TB-20250508		SDG No.:	Q2013	
Lab Sample ID:	Q2013-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046162.D	1		05/13/25 13:38	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/08/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/12/25	
Client Sample ID:	BP-TT192D2-GW-20250508		SDG No.:	Q2013	
Lab Sample ID:	Q2013-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046160.D	1		05/13/25 12:51	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/08/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/12/25	
Client Sample ID:	BP-TT192D2-GW-20250508		SDG No.:	Q2013	
Lab Sample ID:	Q2013-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046160.D	1		05/13/25 12:51	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.9		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.4		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	62300	5.55				
540-36-3	1,4-Difluorobenzene	125000	6.757				
3114-55-4	Chlorobenzene-d5	118000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	49400	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/08/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/12/25	
Client Sample ID:	BP-TT192D2-GW-20250508		SDG No.:	Q2013	
Lab Sample ID:	Q2013-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046160.D	1		05/13/25 12:51	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/09/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/12/25	
Client Sample ID:	BP-TT192D1-GW-20250509		SDG No.:	Q2013	
Lab Sample ID:	Q2013-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046161.D	1		05/13/25 13:15	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/09/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/12/25	
Client Sample ID:	BP-TT192D1-GW-20250509		SDG No.:	Q2013	
Lab Sample ID:	Q2013-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046161.D	1		05/13/25 13:15	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.1		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	52.7		80 - 119		105%	SPK: 50
2037-26-5	Toluene-d8	50.7		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	62700	5.55				
540-36-3	1,4-Difluorobenzene	124000	6.757				
3114-55-4	Chlorobenzene-d5	118000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	49100	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/09/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/12/25	
Client Sample ID:	BP-TT192D1-GW-20250509		SDG No.:	Q2013	
Lab Sample ID:	Q2013-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046161.D	1		05/13/25 13:15	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: Q2013

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2013-01	BP-TB-20250508	1,2-Dichloroethane-d4	50	54.2	108	81	118
		Dibromofluoromethane	50	51.0	102	80	119
		Toluene-d8	50	50.6	101	89	112
		4-Bromofluorobenzene	50	49.6	99	85	114
Q2013-02	BP-TT192D2-GW-20250508	1,2-Dichloroethane-d4	50	53.9	108	81	118
		Dibromofluoromethane	50	49.8	100	80	119
		Toluene-d8	50	49.4	99	89	112
		4-Bromofluorobenzene	50	50.4	101	85	114
Q2013-03	BP-TT192D1-GW-20250509	1,2-Dichloroethane-d4	50	54.1	108	81	118
		Dibromofluoromethane	50	52.7	105	80	119
		Toluene-d8	50	50.7	101	89	112
		4-Bromofluorobenzene	50	51.1	102	85	114
VX0513WBL01	VX0513WBL01	1,2-Dichloroethane-d4	50	53.8	108	81	118
		Dibromofluoromethane	50	51.5	103	80	119
		Toluene-d8	50	50.4	101	89	112
		4-Bromofluorobenzene	50	50.2	100	85	114
VX0513WBS01	VX0513WBS01	1,2-Dichloroethane-d4	50	51.5	103	81	118
		Dibromofluoromethane	50	50.7	101	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	49.3	99	85	114
VX0513WBSD0	VX0513WBSD01	1,2-Dichloroethane-d4	50	52.2	104	81	118
		Dibromofluoromethane	50	51.9	104	80	119
		Toluene-d8	50	51.1	102	89	112
		4-Bromofluorobenzene	50	50.4	101	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2013

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX046158.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0513WBS01	Chloromethane	20	20.0	ug/L	100			50	139	
	Vinyl chloride	20	19.0	ug/L	95			58	137	
	Bromomethane	20	19.1	ug/L	96			53	141	
	Chloroethane	20	21.0	ug/L	105			60	138	
	Trichlorofluoromethane	20	20.1	ug/L	101			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.6	ug/L	103			70	136	
	1,1-Dichloroethene	20	18.9	ug/L	95			71	131	
	Acetone	100	99.7	ug/L	100			39	160	
	Carbon disulfide	20	17.6	ug/L	88			64	133	
	Methyl tert-butyl Ether	20	19.9	ug/L	100			71	124	
	Methylene Chloride	20	18.9	ug/L	95			74	124	
	trans-1,2-Dichloroethene	20	19.6	ug/L	98			75	124	
	1,1-Dichloroethane	20	20.7	ug/L	104			77	125	
	2-Butanone	100	100	ug/L	100			56	143	
	Carbon Tetrachloride	20	19.4	ug/L	97			72	136	
	cis-1,2-Dichloroethene	20	20.0	ug/L	100			78	123	
	Chloroform	20	21.0	ug/L	105			79	124	
	1,1,1-Trichloroethane	20	20.2	ug/L	101			74	131	
	Methylcyclohexane	20	18.1	ug/L	91			72	132	
	Benzene	20	20.0	ug/L	100			79	120	
	1,2-Dichloroethane	20	20.2	ug/L	101			73	128	
	Trichloroethene	20	19.0	ug/L	95			79	123	
	1,2-Dichloropropane	20	20.1	ug/L	101			78	122	
	Bromodichloromethane	20	20.2	ug/L	101			79	125	
	4-Methyl-2-Pentanone	100	99.5	ug/L	100			67	130	
	Toluene	20	19.8	ug/L	99			80	121	
	t-1,3-Dichloropropene	20	19.0	ug/L	95			73	127	
	cis-1,3-Dichloropropene	20	19.2	ug/L	96			75	124	
	1,1,2-Trichloroethane	20	20.1	ug/L	101			80	119	
	2-Hexanone	100	100	ug/L	100			57	139	
	Dibromochloromethane	20	19.9	ug/L	100			74	126	
	Tetrachloroethene	20	20.1	ug/L	101			74	129	
	Chlorobenzene	20	19.7	ug/L	99			82	118	
	Ethyl Benzene	20	19.5	ug/L	98			79	121	
	m/p-Xylenes	40	38.9	ug/L	97			80	121	
	o-Xylene	20	19.6	ug/L	98			78	122	
	Styrene	20	20.1	ug/L	101			78	123	
	Bromoform	20	19.2	ug/L	96			66	130	
	Isopropylbenzene	20	20.4	ug/L	102			72	131	
	1,1,2,2-Tetrachloroethane	20	20.0	ug/L	100			71	121	
	1,3-Dichlorobenzene	20	19.7	ug/L	99			80	119	
	1,4-Dichlorobenzene	20	19.4	ug/L	97			79	118	
	1,2-Dichlorobenzene	20	20.5	ug/L	103			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2013

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX046159.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0513WBSD01	Chloromethane	20	18.9	ug/L	95	5		50	139	20
	Vinyl chloride	20	18.1	ug/L	91	4		58	137	20
	Bromomethane	20	18.6	ug/L	93	3		53	141	20
	Chloroethane	20	19.3	ug/L	97	8		60	138	20
	Trichlorofluoromethane	20	19.8	ug/L	99	2		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	19.5	ug/L	98	5		70	136	20
	1,1-Dichloroethene	20	18.8	ug/L	94	1		71	131	20
	Acetone	100	100	ug/L	100	0		39	160	20
	Carbon disulfide	20	17.0	ug/L	85	3		64	133	20
	Methyl tert-butyl Ether	20	20.3	ug/L	102	2		71	124	20
	Methylene Chloride	20	18.9	ug/L	95	0		74	124	20
	trans-1,2-Dichloroethene	20	19.2	ug/L	96	2		75	124	20
	1,1-Dichloroethane	20	20.0	ug/L	100	4		77	125	20
	2-Butanone	100	110	ug/L	110	10		56	143	20
	Carbon Tetrachloride	20	19.9	ug/L	100	3		72	136	20
	cis-1,2-Dichloroethene	20	20.0	ug/L	100	0		78	123	20
	Chloroform	20	20.4	ug/L	102	3		79	124	20
	1,1,1-Trichloroethane	20	19.4	ug/L	97	4		74	131	20
	Methylcyclohexane	20	18.7	ug/L	94	3		72	132	20
	Benzene	20	19.8	ug/L	99	1		79	120	20
	1,2-Dichloroethane	20	20.8	ug/L	104	3		73	128	20
	Trichloroethene	20	20.0	ug/L	100	5		79	123	20
	1,2-Dichloropropane	20	20.7	ug/L	104	3		78	122	20
	Bromodichloromethane	20	20.9	ug/L	104	3		79	125	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	10		67	130	20
	Toluene	20	20.0	ug/L	100	1		80	121	20
	t-1,3-Dichloropropene	20	19.9	ug/L	100	5		73	127	20
	cis-1,3-Dichloropropene	20	20.0	ug/L	100	4		75	124	20
	1,1,2-Trichloroethane	20	21.7	ug/L	109	8		80	119	20
	2-Hexanone	100	110	ug/L	110	10		57	139	20
	Dibromochloromethane	20	21.3	ug/L	106	6		74	126	20
	Tetrachloroethene	20	19.7	ug/L	99	2		74	129	20
	Chlorobenzene	20	19.2	ug/L	96	3		82	118	20
	Ethyl Benzene	20	19.5	ug/L	98	0		79	121	20
	m/p-Xylenes	40	39.3	ug/L	98	1		80	121	20
	o-Xylene	20	19.8	ug/L	99	1		78	122	20
	Styrene	20	20.4	ug/L	102	1		78	123	20
	Bromoform	20	18.8	ug/L	94	2		66	130	20
	Isopropylbenzene	20	20.0	ug/L	100	2		72	131	20
	1,1,2,2-Tetrachloroethane	20	20.4	ug/L	102	2		71	121	20
	1,3-Dichlorobenzene	20	19.8	ug/L	99	0		80	119	20
	1,4-Dichlorobenzene	20	19.5	ug/L	98	1		79	118	20
	1,2-Dichlorobenzene	20	20.1	ug/L	101	2		80	119	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0513WBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q2013

SAS No.: Q2013 SDG NO.: Q2013

Lab File ID: VX046155.D

Lab Sample ID: VX0513WBL01

Date Analyzed: 05/13/2025

Time Analyzed: 10:52

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0513WBS01	VX0513WBS01	VX046158.D	05/13/2025
VX0513WBSD01	VX0513WBSD01	VX046159.D	05/13/2025
BP-TT192D2-GW-20250508	Q2013-02	VX046160.D	05/13/2025
BP-TT192D1-GW-20250509	Q2013-03	VX046161.D	05/13/2025
BP-TB-20250508	Q2013-01	VX046162.D	05/13/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG NO.: Q2013
Lab File ID: VX046038.D BFB Injection Date: 05/05/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	68.8
175	5.0 - 9.0% of mass 174	5 (7.3) 1
176	95.0 - 101.0% of mass 174	66.7 (97) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC020	VSTDIC020	VX046041.D	05/05/2025	11:35
VSTDIC050	VSTDIC050	VX046042.D	05/05/2025	11:58
VSTDIC100	VSTDIC100	VX046043.D	05/05/2025	12:21
VSTDIC150	VSTDIC150	VX046044.D	05/05/2025	12:45
VSTDIC005	VSTDIC005	VX046046.D	05/05/2025	16:04
VSTDIC001	VSTDIC001	VX046047.D	05/05/2025	16:27

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG NO.: Q2013
Lab File ID: VX046152.D BFB Injection Date: 05/13/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:29
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.8
75	30.0 - 60.0% of mass 95	58.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.3) 1
174	50.0 - 100.0% of mass 95	68.6
175	5.0 - 9.0% of mass 174	4.8 (7) 1
176	95.0 - 101.0% of mass 174	66.3 (96.7) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046153.D	05/13/2025	10:00
VX0513WBL01	VX0513WBL01	VX046155.D	05/13/2025	10:52
VX0513WBS01	VX0513WBS01	VX046158.D	05/13/2025	12:02
VX0513WBSD01	VX0513WBSD01	VX046159.D	05/13/2025	12:28
BP-TT192D2-GW-20250508	Q2013-02	VX046160.D	05/13/2025	12:51
BP-TT192D1-GW-20250509	Q2013-03	VX046161.D	05/13/2025	13:15
BP-TB-20250508	Q2013-01	VX046162.D	05/13/2025	13:38
VSTDCCC050EC	VSTDCCC050	VX046179.D	05/13/2025	20:15

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG NO.: Q2013
Lab File ID: VX046153.D Date Analyzed: 05/13/2025
Instrument ID: MSVOA_X Time Analyzed: 10:00
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	98502	5.54	167387	6.75	146749	10.05
UPPER LIMIT	197004	6.043	334774	7.25	293498	10.549
LOWER LIMIT	49251	5.043	83693.5	6.25	73374.5	9.549
EPA SAMPLE NO.						
BP-TB-20250508	61629	5.55	121469	6.76	115380	10.05
BP-TT192D2-GW-20250508	62262	5.55	125239	6.76	118445	10.05
BP-TT192D1-GW-20250509	62654	5.55	124070	6.76	117714	10.05
VX0513WBL01	67277	5.55	133345	6.76	124650	10.05
VX0513WBS01	89387	5.54	160585	6.76	139963	10.06
VX0513WBSD01	85900	5.54	148923	6.76	132995	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG NO.: Q2013
Lab File ID: VX046153.D Date Analyzed: 05/13/2025
Instrument ID: MSVOA_X Time Analyzed: 10:00
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	70437	12.018				
UPPER LIMIT	140874	12.518				
LOWER LIMIT	35218.5	11.518				
EPA SAMPLE NO.						
BP-TB-20250508	45238	12.02				
BP-TT192D2-GW-20250508	49417	12.02				
BP-TT192D1-GW-20250509	49112	12.02				
VX0513WBL01	52712	12.02				
VX0513WBS01	64268	12.02				
VX0513WBSD01	62827	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0513WBL01		SDG No.:	Q2013
Lab Sample ID:	VX0513WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046155.D	1		05/13/25 10:52	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0513WBL01		SDG No.:	Q2013
Lab Sample ID:	VX0513WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046155.D	1		05/13/25 10:52	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.8		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.5		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	50.4		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.2		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	67300	5.55				
540-36-3	1,4-Difluorobenzene	133000	6.757				
3114-55-4	Chlorobenzene-d5	125000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	52700	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0513WBS01		SDG No.:	Q2013
Lab Sample ID:	VX0513WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046158.D	1		05/13/25 12:02	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	20.0		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	19.0		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	19.1		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	21.0		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	20.1		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.9		0.23	0.75	1.00	ug/L
67-64-1	Acetone	99.7		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.6		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.9		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	18.9		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.6		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.7		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	100		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.4		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.0		0.19	0.75	1.00	ug/L
67-66-3	Chloroform	21.0		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.2		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.1		0.16	0.50	1.00	ug/L
71-43-2	Benzene	20.0		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.2		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.0		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.1		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.2		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	99.5		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.8		0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	19.0		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.2		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.1		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	100		0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0513WBS01		SDG No.:	Q2013
Lab Sample ID:	VX0513WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046158.D	1		05/13/25 12:02	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	19.9		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	20.1		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.7		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.5		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	38.9		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.6		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.1		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	19.2		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.4		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.0		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.7		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.4		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.5		0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.5		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	50.7		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	50.0		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.3		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	89400	5.544				
540-36-3	1,4-Difluorobenzene	161000	6.757				
3114-55-4	Chlorobenzene-d5	140000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	64300	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0513WBSD01		SDG No.:	Q2013
Lab Sample ID:	VX0513WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046159.D	1		05/13/25 12:28	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	18.9		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.1		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	18.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.3		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.8		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.5		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.8		0.23	0.75	1.00	ug/L
67-64-1	Acetone	100		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.3		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	18.9		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.2		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.0		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.0		0.19	0.75	1.00	ug/L
67-66-3	Chloroform	20.4		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.4		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.7		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.8		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.8		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	20.0		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.7		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.9		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	20.0		0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	19.9		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.0		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.7		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0513WBSD01		SDG No.:	Q2013
Lab Sample ID:	VX0513WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046159.D	1		05/13/25 12:28	VX051325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	21.3		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.2		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.5		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.3		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.8		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.4		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	18.8		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.0		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.8		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.5		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.1		0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.2		81 - 118		104%	SPK: 50
1868-53-7	Dibromofluoromethane	51.9		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	51.1		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	85900	5.544				
540-36-3	1,4-Difluorobenzene	149000	6.757				
3114-55-4	Chlorobenzene-d5	133000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	62800	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2013</u>
Instrument ID:	<u>MSVOA_X</u>	SAS No.:	<u>Q2013</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>Q2013</u>
GC Column:	<u>DB-624UI</u>	Calibration Date(s):	<u>05/05/2025</u>
ID:	<u>0.18 (mm)</u>	Calibration Time(s):	<u>11:35</u>
			<u>16:27</u>

LAB FILE ID:	RRF020 = VX046041.D	RRF050 = VX046042.D	RRF100 = VX046043.D	RRF150 = VX046044.D	RRF005 = VX046046.D	RRF001 = VX046047.D		
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD
Chloromethane	0.727	0.775	0.787	0.791	0.679	0.694	0.742	6.6
Vinyl Chloride	0.660	0.710	0.727	0.755	0.619	0.673	0.691	7.2
Bromomethane	0.296	0.326	0.340	0.334	0.305		0.320	5.8
Chloroethane	0.354	0.378	0.329	0.317	0.368	0.467	0.369	14.4
Trichlorofluoromethane	1.035	1.068	0.983	0.985	0.990	1.064	1.021	3.9
1,1,2-Trichlorotrifluoroethane	0.628	0.641	0.629	0.648	0.610	0.633	0.632	2.1
1,1-Dichloroethene	0.565	0.601	0.607	0.625	0.567	0.594	0.593	3.9
Acetone	0.361	0.362	0.361	0.370	0.408	0.380	0.374	4.9
Carbon Disulfide	1.295	1.455	1.522	1.597	1.141	1.423	1.406	11.7
Methyl tert-butyl Ether	2.044	2.160	2.172	2.239	1.908	1.949	2.079	6.4
Methylene Chloride	0.689	0.684	0.691	0.691	0.689	0.853	0.716	9.4
trans-1,2-Dichloroethene	0.573	0.610	0.612	0.622	0.557	0.604	0.596	4.3
1,1-Dichloroethane	1.233	1.263	1.263	1.286	1.154	1.116	1.219	5.6
2-Butanone	0.540	0.555	0.558	0.569	0.539	0.495	0.543	4.8
Carbon Tetrachloride	0.528	0.558	0.552	0.577	0.505	0.541	0.544	4.6
cis-1,2-Dichloroethene	0.716	0.737	0.738	0.755	0.642	0.719	0.718	5.5
Chloroform	1.287	1.296	1.277	1.300	1.199	1.265	1.271	3
1,1,1-Trichloroethane	1.106	1.131	1.155	1.188	1.013	1.015	1.101	6.6
Methylcyclohexane	0.596	0.641	0.627	0.658	0.587	0.627	0.623	4.3
Benzene	1.426	1.474	1.441	1.477	1.337	1.348	1.417	4.3
1,2-Dichloroethane	0.632	0.627	0.611	0.625	0.594	0.579	0.612	3.5
Trichloroethene	0.344	0.355	0.345	0.362	0.315	0.324	0.341	5.3
1,2-Dichloropropane	0.356	0.371	0.368	0.378	0.324	0.317	0.352	7.4
Bromodichloromethane	0.557	0.577	0.573	0.594	0.498	0.485	0.547	8.2
4-Methyl-2-Pentanone	0.620	0.634	0.630	0.631	0.555	0.561	0.605	6
Toluene	0.884	0.898	0.885	0.904	0.838	0.803	0.869	4.5
t-1,3-Dichloropropene	0.468	0.528	0.555	0.591	0.406	0.371	0.487	17.9
cis-1,3-Dichloropropene	0.531	0.578	0.602	0.623	0.469	0.423	0.538	14.6
1,1,2-Trichloroethane	0.349	0.354	0.351	0.356	0.337	0.308	0.343	5.3
2-Hexanone	0.466	0.473	0.477	0.473	0.414	0.385	0.448	8.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG No.: Q2013

Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D			
		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D			
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD	
Dibromochloromethane	0.378	0.400	0.415	0.431	0.326	0.306	0.376	13.3	
Tetrachloroethene	0.390	0.375	0.345	0.344	0.323	0.347	0.354	6.8	
Chlorobenzene	1.093	1.098	1.085	1.114	1.046	1.131	1.094	2.7	
Ethyl Benzene	1.919	2.022	1.979	2.036	1.816	1.803	1.929	5.2	
m/p-Xylenes	0.706	0.740	0.721	0.740	0.678	0.648	0.706	5.2	
o-Xylene	0.688	0.727	0.706	0.726	0.639	0.642	0.688	5.7	
Styrene	1.135	1.219	1.214	1.230	1.012	0.951	1.127	10.6	
Bromoform	0.270	0.304	0.312	0.327	0.236	0.234	0.281	14.2	
Isopropylbenzene	3.843	4.130	3.876	4.156	3.562	3.789	3.893	5.7	
1,1,2,2-Tetrachloroethane	1.315	1.338	1.284	1.345	1.350	1.552	1.364	7	
1,3-Dichlorobenzene	1.633	1.701	1.656	1.730	1.558	1.619	1.649	3.7	
1,4-Dichlorobenzene	1.629	1.693	1.639	1.722	1.606	1.817	1.684	4.6	
1,2-Dichlorobenzene	1.613	1.696	1.634	1.702	1.577	1.710	1.655	3.3	
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6	
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7	
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2	
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG No.: Q2013

Instrument ID: MSVOA_X Calibration Date/Time: 05/13/2025 10:00

Lab File ID: VX046153.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.751	0.1	1.21	20
Vinyl Chloride	0.691	0.696		0.72	20
Bromomethane	0.320	0.304		-5	20
Chloroethane	0.369	0.375		1.63	20
Trichlorofluoromethane	1.021	1.066		4.41	20
1,1,2-Trichlorotrifluoroethane	0.632	0.653		3.32	20
1,1-Dichloroethene	0.593	0.578		-2.53	20
Acetone	0.374	0.410		9.63	20
Carbon Disulfide	1.406	1.375		-2.2	20
Methyl tert-butyl Ether	2.079	2.132		2.55	20
Methylene Chloride	0.716	0.668		-6.7	20
trans-1,2-Dichloroethene	0.596	0.589		-1.17	20
1,1-Dichloroethane	1.219	1.231	0.1	0.98	20
2-Butanone	0.543	0.561		3.32	20
Carbon Tetrachloride	0.544	0.574		5.51	20
cis-1,2-Dichloroethene	0.718	0.715		-0.42	20
Chloroform	1.271	1.300		2.28	20
1,1,1-Trichloroethane	1.101	1.117		1.45	20
Methylcyclohexane	0.623	0.621		-0.32	20
Benzene	1.417	1.440		1.62	20
1,2-Dichloroethane	0.612	0.625		2.12	20
Trichloroethene	0.341	0.344		0.88	20
1,2-Dichloropropane	0.352	0.370		5.11	20
Bromodichloromethane	0.547	0.588		7.49	20
4-Methyl-2-Pentanone	0.605	0.632		4.46	20
Toluene	0.869	0.884		1.73	20
t-1,3-Dichloropropene	0.487	0.532		9.24	20
cis-1,3-Dichloropropene	0.538	0.581		7.99	20
1,1,2-Trichloroethane	0.343	0.351		2.33	20
2-Hexanone	0.448	0.467		4.24	20
Dibromochloromethane	0.376	0.418		11.17	20
Tetrachloroethene	0.354	0.368		3.95	20
Chlorobenzene	1.094	1.091	0.3	-0.27	20
Ethyl Benzene	1.929	1.981		2.7	20
m/p-Xylenes	0.706	0.736		4.25	20
o-Xylene	0.688	0.715		3.92	20
Styrene	1.127	1.209		7.28	20
Bromoform	0.281	0.307	0.1	9.25	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG No.: Q2013

Instrument ID: MSVOA_X Calibration Date/Time: 05/13/2025 10:00

Lab File ID: VX046153.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.893	4.009		2.98	20
1,1,2,2-Tetrachloroethane	1.364	1.283	0.3	-5.94	20
1,3-Dichlorobenzene	1.649	1.673		1.46	20
1,4-Dichlorobenzene	1.684	1.657		-1.6	20
1,2-Dichlorobenzene	1.655	1.637		-1.09	20
1,2-Dichloroethane-d4	0.932	0.899		-3.54	20
Dibromofluoromethane	0.360	0.371		3.06	20
Toluene-d8	1.246	1.236		-0.8	20
4-Bromofluorobenzene	0.478	0.484		1.25	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG No.: Q2013

Instrument ID: MSVOA_X Calibration Date/Time: 05/13/2025 20:15

Lab File ID: VX046179.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.834	0.1	12.4	50
Vinyl Chloride	0.691	0.739		6.95	50
Bromomethane	0.320	0.310		-3.13	50
Chloroethane	0.369	0.409		10.84	50
Trichlorofluoromethane	1.021	1.112		8.91	50
1,1,2-Trichlorotrifluoroethane	0.632	0.671		6.17	50
1,1-Dichloroethene	0.593	0.631		6.41	50
Acetone	0.374	0.410		9.63	50
Carbon Disulfide	1.406	1.446		2.85	50
Methyl tert-butyl Ether	2.079	2.375		14.24	50
Methylene Chloride	0.716	0.739		3.21	50
trans-1,2-Dichloroethene	0.596	0.632		6.04	50
1,1-Dichloroethane	1.219	1.342	0.1	10.09	50
2-Butanone	0.543	0.622		14.55	50
Carbon Tetrachloride	0.544	0.571		4.96	50
cis-1,2-Dichloroethene	0.718	0.790		10.03	50
Chloroform	1.271	1.395		9.76	50
1,1,1-Trichloroethane	1.101	1.237		12.35	50
Methylcyclohexane	0.623	0.637		2.25	50
Benzene	1.417	1.515		6.92	50
1,2-Dichloroethane	0.612	0.643		5.07	50
Trichloroethene	0.341	0.356		4.4	50
1,2-Dichloropropane	0.352	0.385		9.38	50
Bromodichloromethane	0.547	0.591		8.04	50
4-Methyl-2-Pentanone	0.605	0.681		12.56	50
Toluene	0.869	0.937		7.82	50
t-1,3-Dichloropropene	0.487	0.529		8.62	50
cis-1,3-Dichloropropene	0.538	0.588		9.29	50
1,1,2-Trichloroethane	0.343	0.377		9.91	50
2-Hexanone	0.448	0.514		14.73	50
Dibromochloromethane	0.376	0.425		13.03	50
Tetrachloroethene	0.354	0.332		-6.22	50
Chlorobenzene	1.094	1.133	0.3	3.57	50
Ethyl Benzene	1.929	2.060		6.79	50
m/p-Xylenes	0.706	0.748		5.95	50
o-Xylene	0.688	0.747		8.58	50
Styrene	1.127	1.257		11.53	50
Bromoform	0.281	0.299	0.1	6.41	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG No.: Q2013

Instrument ID: MSVOA_X Calibration Date/Time: 05/13/2025 20:15

Lab File ID: VX046179.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.893	4.228		8.6	50
1,1,2,2-Tetrachloroethane	1.364	1.421	0.3	4.18	50
1,3-Dichlorobenzene	1.649	1.766		7.09	50
1,4-Dichlorobenzene	1.684	1.744		3.56	50
1,2-Dichlorobenzene	1.655	1.753		5.92	50
1,2-Dichloroethane-d4	0.932	0.935		0.32	50
Dibromofluoromethane	0.360	0.356		-1.11	50
Toluene-d8	1.246	1.240		-0.48	50
4-Bromofluorobenzene	0.478	0.491		2.72	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	Q2013	OrderDate:	5/12/2025 10:15:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2013-02	BP-TT192D2-GW-202 50508	Water			05/08/25			05/12/25
			SVOC-SIMGroup1	8270-Modified		05/13/25	05/14/25	
Q2013-03	BP-TT192D1-GW-202 50509	Water			05/09/25			05/12/25
			SVOC-SIMGroup1	8270-Modified		05/13/25	05/14/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q2013

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : BP-TT192D2-GW-20250508								
Q2013-02	BP-TT192D2-GW-20250 WATER	1,4-Dioxane	1.200		0.09	0.26	0.26	ug/L
		Total Svoc :		1.20				
		Total Concentration:		1.20				
Client ID : BP-TT192D1-GW-20250509								
Q2013-03	BP-TT192D1-GW-20250 WATER	1,4-Dioxane	1.500		0.08	0.24	0.24	ug/L
		Total Svoc :		1.50				
		Total Concentration:		1.50				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/08/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/12/25
Client Sample ID:	BP-TT192D2-GW-20250508	SDG No.:	Q2013
Lab Sample ID:	Q2013-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	770 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037019.D	1	05/13/25 08:49	05/14/25 19:12	PB167952

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	1.20		0.090	0.26	0.26	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		79%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 - 150		96%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		78%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		53 - 106		86%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.69	*	58 - 132		172%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2100	7.618				
1146-65-2	Naphthalene-d8	5680	10.394				
15067-26-2	Acenaphthene-d10	3260	14.267				
1517-22-2	Phenanthrene-d10	6420	17.009				
1719-03-5	Chrysene-d12	5770	21.207				
1520-96-3	Perylene-d12	5630	23.409				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/09/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/12/25
Client Sample ID:	BP-TT192D1-GW-20250509	SDG No.:	Q2013
Lab Sample ID:	Q2013-03	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	820 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037020.D	1	05/13/25 08:49	05/14/25 19:48	PB167952

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	1.50		0.080	0.24	0.24	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		79%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 - 150		94%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		74%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.31		53 - 106		76%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.81	*	58 - 132		203%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2080		7.618			
1146-65-2	Naphthalene-d8	5600		10.393			
15067-26-2	Acenaphthene-d10	3310		14.266			
1517-22-2	Phenanthrene-d10	6610		17.008			
1719-03-5	Chrysene-d12	5760		21.206			
1520-96-3	Perylene-d12	5190		23.409			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2013

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167952BL	PB167952BL	2-Methylnaphthalene-d10	0.4	0.36	90		30	150
		Fluoranthene-d10	0.4	0.37	91		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.37	93		53	106
		Terphenyl-d14	0.4	0.39	97		58	132
PB167952BS	PB167952BS	2-Methylnaphthalene-d10	0.4	0.40	100		30	150
		Fluoranthene-d10	0.4	0.34	85		30	150
		Nitrobenzene-d5	0.4	0.35	87		55	111
		2-Fluorobiphenyl	0.4	0.37	92		53	106
		Terphenyl-d14	0.4	0.40	100		58	132
PB167952BSD	PB167952BSD	2-Methylnaphthalene-d10	0.4	0.40	100		30	150
		Fluoranthene-d10	0.4	0.39	96		30	150
		Nitrobenzene-d5	0.4	0.38	94		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.41	103		58	132
Q2013-02	BP-TT192D2-GW-20250508	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.38	96		30	150
		Nitrobenzene-d5	0.4	0.31	78		55	111
		2-Fluorobiphenyl	0.4	0.35	86		53	106
		Terphenyl-d14	0.4	0.69	172	*	58	132
Q2013-03	BP-TT192D1-GW-20250509	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.30	74		55	111
		2-Fluorobiphenyl	0.4	0.31	76		53	106
		Terphenyl-d14	0.4	0.81	203	*	58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2013

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified **DataFile:** BN037021.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167952BS	1,4-Dioxane	0.4	0.36	ug/L	90				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2013

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified **DataFile:** BN037022.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167952BSD	1,4-Dioxane	0.4	0.40	ug/L	100	11			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167952BL

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q2013

SAS No.: Q2013 SDG NO.: Q2013

Lab File ID: BN037010.D

Lab Sample ID: PB167952BL

Instrument ID: BNA_N

Date Extracted: 05/13/2025

Matrix: (soil/water) Water

Date Analyzed: 05/14/2025

Level: (low/med) LOW

Time Analyzed: 11:20

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167952BS	PB167952BS	BN037021.D	05/14/2025
BP-TT192D2-GW-20250508	Q2013-02	BN037019.D	05/14/2025
BP-TT192D1-GW-20250509	Q2013-03	BN037020.D	05/14/2025
PB167952BSD	PB167952BSD	BN037022.D	05/14/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2013

SDG NO.: Q2013

Lab File ID: BN036998.D

DFTPP Injection Date: 05/13/2025

Instrument ID: BNA_N

DFTPP Injection Time: 17:02

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	62.8
68	Less than 2.0% of mass 69	0.8 (1.4) 1
69	Mass 69 relative abundance	55.6
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	52.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	8.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.4 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036999.D	05/13/2025	17:41
SSTDICC0.2	SSTDICC0.2	BN037000.D	05/13/2025	18:17
SSTDICCC0.4	SSTDICCC0.4	BN037001.D	05/13/2025	18:53
SSTDICC0.8	SSTDICC0.8	BN037002.D	05/13/2025	19:29
SSTDICC1.6	SSTDICC1.6	BN037003.D	05/13/2025	20:05
SSTDICC3.2	SSTDICC3.2	BN037004.D	05/13/2025	20:41
SSTDICC5.0	SSTDICC5.0	BN037005.D	05/13/2025	21:17

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2013 SDG NO.: Q2013

Lab File ID: BN037008.D

DFTPP Injection Date: 05/14/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:37

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	71.5
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	59.2
70	Less than 2.0% of mass 69	0.4 (0.6) 1
127	10.0 - 80.0% of mass 198	55.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	24.4
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	8.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.2 (18.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037009.D	05/14/2025	10:31
PB167952BL	PB167952BL	BN037010.D	05/14/2025	11:20
BP-TT192D2-GW-20250508	Q2013-02	BN037019.D	05/14/2025	19:12
BP-TT192D1-GW-20250509	Q2013-03	BN037020.D	05/14/2025	19:48
PB167952BS	PB167952BS	BN037021.D	05/14/2025	20:24
PB167952BSD	PB167952BSD	BN037022.D	05/14/2025	21:00
SSTDCCC0.4EC	SSTDCCC0.4	BN037023.D	05/14/2025	21:36

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG NO.: Q2013

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/14/2025

Lab File ID: BN037009.D Time Analyzed: 10:31

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1633	7.618	4300	10.39	2444	14.27
UPPER LIMIT	3266	8.118	8600	10.894	4888	14.767
LOWER LIMIT	816.5	7.118	2150	9.894	1222	13.767
EPA SAMPLE NO.						
01 PB167952BL	1697	7.62	4346	10.40	2391	14.27
02 PB167952BS	2111	7.62	5540	10.39	3182	14.27
03 PB167952BSD	1677	7.62	4591	10.39	2680	14.27
04 BP-TT192D2-GW-20250508	2098	7.62	5678	10.39	3261	14.27
05 BP-TT192D1-GW-20250509	2077	7.62	5596	10.39	3313	14.27

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG NO.: Q2013

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/14/2025

Lab File ID: BN037009.D Time Analyzed: 10:31

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	4780	17.009	4137	21.215	4043	23.421
UPPER LIMIT	9560	17.509	8274	21.715	8086	23.921
LOWER LIMIT	2390	16.509	2068.5	20.715	2021.5	22.921
EPA SAMPLE NO.						
01 PB167952BL	4921	17.02	4198	21.22	4092	23.42
02 PB167952BS	6367	17.01	4959	21.21	4122	23.41
03 PB167952BSD	5142	17.01	4350	21.21	3999	23.41
04 BP-TT192D2-GW-20250508	6421	17.01	5769	21.21	5632	23.41
05 BP-TT192D1-GW-20250509	6611	17.01	5758	21.21	5194	23.41

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB167952BL		SDG No.:	Q2013
Lab Sample ID:	PB167952BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037010.D	1	05/13/25 08:49	05/14/25 11:20	PB167952

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		90%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		91%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		93%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.39		58 - 132		97%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1700	7.618				
1146-65-2	Naphthalene-d8	4350	10.404				
15067-26-2	Acenaphthene-d10	2390	14.267				
1517-22-2	Phenanthrene-d10	4920	17.021				
1719-03-5	Chrysene-d12	4200	21.215				
1520-96-3	Perylene-d12	4090	23.418				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB167952BS		SDG No.:	Q2013
Lab Sample ID:	PB167952BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037021.D	1	05/13/25 08:49	05/14/25 20:24	PB167952

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.36		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.40		30 - 150		100%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		85%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.35		55 - 111		87%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		92%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.40		58 - 132		100%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2110	7.618				
1146-65-2	Naphthalene-d8	5540	10.394				
15067-26-2	Acenaphthene-d10	3180	14.267				
1517-22-2	Phenanthrene-d10	6370	17.009				
1719-03-5	Chrysene-d12	4960	21.207				
1520-96-3	Perylene-d12	4120	23.41				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB167952BSD		SDG No.:	Q2013
Lab Sample ID:	PB167952BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037022.D	1	05/13/25 08:49	05/14/25 21:00	PB167952

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.40		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.40		30 - 150		100%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		96%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.38		55 - 111		94%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		98%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.41		58 - 132		103%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1680	7.618				
1146-65-2	Naphthalene-d8	4590	10.394				
15067-26-2	Acenaphthene-d10	2680	14.267				
1517-22-2	Phenanthrene-d10	5140	17.009				
1719-03-5	Chrysene-d12	4350	21.207				
1520-96-3	Perylene-d12	4000	23.407				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN051425.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 14 11:26:32 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN036999.D 0.2 =BN037000.D 0.4 =BN037001.D 0.8 =BN037002.D 1.6 =BN037003.D 3.2 =BN037004.D 5.0 =BN037005.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.510	0.512	0.487	0.514	0.467	0.454	0.491	5.25	
3)	n-Nitrosodimet...	1.465	0.974	0.980	0.971	1.075	0.967	0.950	1.054	17.59
4) S	2-Fluorophenol	1.101	1.134	1.024	1.093	0.964	0.971	1.048	6.87	
5) S	Phenol-d6	1.304	1.385	1.236	1.392	1.259	1.292	1.311	4.91	
6)	bis(2-Chloroet...	1.441	1.163	1.153	1.135	1.240	1.168	1.148	1.207	9.02
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.546	0.383	0.398	0.400	0.452	0.426	0.442	0.436	12.60
9)	Naphthalene	1.326	1.140	1.144	1.122	1.226	1.152	1.165	1.182	6.05
10)	Hexachlorobuta...	0.286	0.248	0.244	0.235	0.256	0.236	0.233	0.248	7.47
11) SURR2	Methylnaphth...	0.529	0.547	0.552	0.548	0.603	0.574	0.588	0.563	4.65
12)	2-Methylnaphth...	0.754	0.724	0.736	0.733	0.814	0.770	0.790	0.760	4.34
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.174	0.168	0.178	0.160	0.186	0.175	0.189	0.176	5.77
15) S	2-Fluorobiphenyl	1.912	1.801	1.901	1.802	1.927	1.672	1.807	1.832	4.90
16)	Acenaphthylene	1.906	1.838	1.894	1.849	2.071	1.997	2.075	1.947	5.14
17)	Acenaphthene	1.255	1.229	1.243	1.217	1.350	1.298	1.315	1.272	3.89
18)	Fluorene	1.602	1.581	1.635	1.611	1.779	1.721	1.752	1.669	4.80
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.060	0.073	0.079	0.102	0.103	0.124	0.090	26.02	
21)	4-Bromophenyl-...	0.243	0.246	0.250	0.247	0.262	0.261	0.259	0.253	3.13
22)	Hexachlorobenzene	0.267	0.269	0.281	0.259	0.281	0.270	0.267	0.270	3.03
23)	Atrazine	0.199	0.207	0.211	0.213	0.237	0.234	0.242	0.220	7.64
24)	Pentachlorophenol	0.133	0.134	0.141	0.137	0.159	0.162	0.177	0.149	11.45
25)	Phenanthrene	1.259	1.272	1.292	1.263	1.367	1.337	1.361	1.307	3.56
26)	Anthracene	1.099	1.104	1.166	1.130	1.269	1.259	1.300	1.190	7.13
27) SURR	Fluoranthene-d10	1.033	1.033	1.078	1.042	1.153	1.161	1.178	1.097	5.95
28)	Fluoranthene	1.461	1.439	1.500	1.496	1.670	1.672	1.693	1.562	7.13
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.744	1.708	1.727	1.656	1.790	1.641	1.711	1.711	2.96
31) S	Terphenyl-d14	0.897	0.844	0.871	0.822	0.891	0.816	0.848	0.856	3.73
32)	Benzo(a)anthra...	1.463	1.432	1.485	1.438	1.594	1.521	1.609	1.506	4.77
33)	Chrysene	1.655	1.559	1.616	1.532	1.653	1.560	1.576	1.593	3.05
34)	Bis(2-ethylhex...	0.955	0.919	0.906	0.855	0.941	0.903	1.011	0.927	5.27
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN051425.M

36)	Indeno(1,2,3-c...	1.511	1.613	1.645	1.568	1.687	1.732	1.680	1.634	4.65
37)	Benzo(b)fluora...	1.631	1.570	1.602	1.599	1.749	1.698	1.765	1.659	4.71
38)	Benzo(k)fluora...	1.539	1.538	1.642	1.601	1.770	1.661	1.719	1.639	5.34
39) C	Benzo(a)pyrene	1.380	1.343	1.381	1.331	1.486	1.444	1.486	1.407	4.59
40)	Dibenzo(a,h)an...	1.116	1.232	1.273	1.237	1.340	1.376	1.334	1.272	6.90
41)	Benzo(g,h,i)pe...	1.299	1.407	1.424	1.330	1.403	1.439	1.376	1.383	3.72

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG No.: Q2013

Instrument ID: BNA_N Calibration Date/Time: 05/14/2025 10:31

Lab File ID: BN037009.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.552		-2.0	20.0
Fluoranthene-d10	1.097	1.065		-2.9	20.0
2-Fluorophenol	1.048	1.078		2.9	20.0
Phenol-d6	1.311	1.303		-0.6	20.0
Nitrobenzene-d5	0.436	0.407		-6.7	20.0
2-Fluorobiphenyl	1.832	1.850		1.0	20.0
2,4,6-Tribromophenol	0.176	0.176		0.0	20.0
Terphenyl-d14	0.856	0.892		4.2	20.0
1,4-Dioxane	0.491	0.448		-8.8	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG No.: Q2013

Instrument ID: BNA_N Calibration Date/Time: 05/14/2025 21:36

Lab File ID: BN037023.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 17:41 21:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.563	0.559		-0.7	50.0
Fluoranthene-d10	1.097	1.047		-4.6	50.0
2-Fluorophenol	1.048	1.129		7.7	50.0
Phenol-d6	1.311	1.364		4.0	50.0
Nitrobenzene-d5	0.436	0.406		-6.9	50.0
2-Fluorobiphenyl	1.832	1.894		3.3	50.0
2,4,6-Tribromophenol	0.176	0.178		1.1	50.0
Terphenyl-d14	0.856	0.900		5.1	50.0
1,4-Dioxane	0.491	0.508		3.5	50.0

All other compounds must meet a minimum RRF of 0.010.