

Surrogate Summary

SW-846

SDG No.: Q2013

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167952BL	PB167952BL	2-Methylnaphthalene-d10	0.4	0.36	90		30	150
		Fluoranthene-d10	0.4	0.37	91		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.37	93		53	106
		Terphenyl-d14	0.4	0.39	97		58	132
PB167952BS	PB167952BS	2-Methylnaphthalene-d10	0.4	0.40	100		30	150
		Fluoranthene-d10	0.4	0.34	85		30	150
		Nitrobenzene-d5	0.4	0.35	87		55	111
		2-Fluorobiphenyl	0.4	0.37	92		53	106
		Terphenyl-d14	0.4	0.40	100		58	132
PB167952BSD	PB167952BSD	2-Methylnaphthalene-d10	0.4	0.40	100		30	150
		Fluoranthene-d10	0.4	0.39	96		30	150
		Nitrobenzene-d5	0.4	0.38	94		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.41	103		58	132
Q2013-02	BP-TT192D2-GW-20250508	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.38	96		30	150
		Nitrobenzene-d5	0.4	0.31	78		55	111
		2-Fluorobiphenyl	0.4	0.35	86		53	106
		Terphenyl-d14	0.4	0.69	172	*	58	132
Q2013-03	BP-TT192D1-GW-20250509	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.30	74		55	111
		2-Fluorobiphenyl	0.4	0.31	76		53	106
		Terphenyl-d14	0.4	0.81	203	*	58	132