


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX051325\  
Data File : VX046152.D  
Acq On : 13 May 2025 09:29  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X050525W.M  
Title : SW846 8260  
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw   | Result    |
|--------|---------|--------|--------|-------|-------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn   | Pass/Fail |
| 50     | 95      | 15     | 40     | 22.8  | 7764  | PASS      |
| 75     | 95      | 30     | 60     | 58.5  | 19919 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 34040 | PASS      |
| 96     | 95      | 5      | 9      | 6.8   | 2313  | PASS      |
| 173    | 174     | 0.00   | 2      | 1.3   | 300   | PASS      |
| 174    | 95      | 50     | 100    | 68.6  | 23352 | PASS      |
| 175    | 174     | 5      | 9      | 7.0   | 1623  | PASS      |
| 176    | 174     | 95     | 101    | 96.7  | 22576 | PASS      |
| 177    | 176     | 5      | 9      | 6.9   | 1554  | PASS      |