

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG NO.: Q2013
Lab File ID: VX046153.D Date Analyzed: 05/13/2025
Instrument ID: MSVOA_X Time Analyzed: 10:00
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	98502	5.54	167387	6.75	146749	10.05
UPPER LIMIT	197004	6.043	334774	7.25	293498	10.549
LOWER LIMIT	49251	5.043	83693.5	6.25	73374.5	9.549
EPA SAMPLE NO.						
BP-TB-20250508	61629	5.55	121469	6.76	115380	10.05
BP-TT192D2-GW-20250508	62262	5.55	125239	6.76	118445	10.05
BP-TT192D1-GW-20250509	62654	5.55	124070	6.76	117714	10.05
VX0513WBL01	67277	5.55	133345	6.76	124650	10.05
VX0513WBS01	89387	5.54	160585	6.76	139963	10.06
VX0513WBSD01	85900	5.54	148923	6.76	132995	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.