

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06  
 Lab Code: CHEM Case No.: Q2013 SAS No.: Q2013 SDG NO.: Q2013  
 Lab File ID: VX046153.D Date Analyzed: 05/13/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 10:00  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                        | IS4<br>AREA # | RT #   |  |  |  |  |
|------------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD            | 70437         | 12.018 |  |  |  |  |
| UPPER LIMIT            | 140874        | 12.518 |  |  |  |  |
| LOWER LIMIT            | 35218.5       | 11.518 |  |  |  |  |
| EPA SAMPLE NO.         |               |        |  |  |  |  |
| BP-TB-20250508         | 45238         | 12.02  |  |  |  |  |
| BP-TT192D2-GW-20250508 | 49417         | 12.02  |  |  |  |  |
| BP-TT192D1-GW-20250509 | 49112         | 12.02  |  |  |  |  |
| VX0513WBL01            | 52712         | 12.02  |  |  |  |  |
| VX0513WBS01            | 64268         | 12.02  |  |  |  |  |
| VX0513WBSD01           | 62827         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.