

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2018 SAS No.: Q2018 SDG NO.: Q2018
Lab File ID: VX046181.D Date Analyzed: 05/14/2025
Instrument ID: MSVOA_X Time Analyzed: 09:59
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	92522	5.54	155707	6.76	140519	10.05
UPPER LIMIT	185044	6.043	311414	7.257	281038	10.549
LOWER LIMIT	46261	5.043	77853.5	6.257	70259.5	9.549
EPA SAMPLE NO.						
MW2DL	71196	5.55	139917	6.76	132344	10.06
VX0514WBL01	66143	5.55	133003	6.76	124085	10.05
VX0514WBS01	87880	5.55	155746	6.76	138245	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.