

Continuing Calibration Report for SequenceID : FG051625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015862.D

Aliphatic C9-C12	6919996.000	60.000	3.136	6.775	115333.267	113522.354	-1.595
Aliphatic C12-C16	4638382.000	40.000	6.776	10.226	115959.550	116725.144	0.656
Aliphatic C16-C21	7015333.000	60.000	10.227	13.601	116922.217	122199.347	4.318
Aliphatic C21-C28	8940416.000	80.000	13.602	17.272	111755.200	120144.859	6.983
Aliphatic C28-C40	12361219.000	120.000	17.273	22.179	103010.158	110433.346	6.722
Aliphatic EPH	39875346.000	360.000	3.136	22.179	110764.850	115766.384	4.320