

SOIL TPH GC SURROGATE RECOVERY

Lab Name: Chemtech Client: Sciacca General Contractors, LLC
Lab Code: CHEM Case No.: Q2021 SAS No.: Q2021 SDG No.: Q2021

EPA SAMPLE NO.	S1 TETRACOSANE-d50	S2	S3	S4	TOT OUT
PIBLK-FF015850.D	85				0
PIBLK-FF015859.D	92				0
PB168042BL	81				0
PB168042BS	87				0
WASTEMS	61				0
WASTEMSD	58				0
WASTE	57				0

QC LIMITS

TETRACOSANE-d50

For Water : 29-130
For Soil : 37-130

Column to be used to flag recovery values
* Values outside of contract required QC limits
D Surrogate Diluted Out