



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q2025

**Order ID:** Q2025

**Client:** Tully Environmental, Inc

**Project ID:** Transfer Station-SPDES

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
<b>Client ID : 001-WILLETS-PT-BLVD(APR)</b>								
Q2025-01	001-WILLETS-PT-BLVD(APR)	Water	Copper	47.7		1.89	10.0	ug/L
Q2025-01	001-WILLETS-PT-BLVD(APR)	Water	Iron	9600		18.0	50.0	ug/L
Q2025-01	001-WILLETS-PT-BLVD(APR)	Water	Lead	52.0		1.21	6.00	ug/L
<b>Client ID : 002-35TH-AVE(APR)</b>								
Q2025-02	002-35TH-AVE(APR)	Water	Copper	41.5		1.89	10.0	ug/L
Q2025-02	002-35TH-AVE(APR)	Water	Iron	8730		18.0	50.0	ug/L
Q2025-02	002-35TH-AVE(APR)	Water	Lead	46.8		1.21	6.00	ug/L



# SAMPLE DATA

A

B

C

D

## Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	05/12/25
Project:	Transfer Station-SPDES	Date Received:	05/13/25
Client Sample ID:	001-WILLETS-PT-BLVD(APR)	SDG No.:	Q2025
Lab Sample ID:	Q2025-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	47.7		1	1.89	10.0	ug/L	05/13/25 15:05	05/14/25 15:01	EPA 200.7	
7439-89-6	Iron	9600	N	1	18.0	50.0	ug/L	05/13/25 15:05	05/14/25 15:01	EPA 200.7	
7439-92-1	Lead	52.0		1	1.21	6.00	ug/L	05/13/25 15:05	05/14/25 15:01	EPA 200.7	

Color Before:	Colorless	Clarity Before:	Cloudy	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group 10			

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits

## Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	05/12/25
Project:	Transfer Station-SPDES	Date Received:	05/13/25
Client Sample ID:	002-35TH-AVE(APR)	SDG No.:	Q2025
Lab Sample ID:	Q2025-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	41.5		1	1.89	10.0	ug/L	05/13/25 15:05	05/14/25 15:05	EPA 200.7	
7439-89-6	Iron	8730	N	1	18.0	50.0	ug/L	05/13/25 15:05	05/14/25 15:05	EPA 200.7	
7439-92-1	Lead	46.8		1	1.21	6.00	ug/L	05/13/25 15:05	05/14/25 15:05	EPA 200.7	

Color Before:	Colorless	Clarity Before:	Cloudy	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group 10			

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q2025	OrderDate:	5/13/2025 12:52:00 PM
Client:	Tully Environmental, Inc	Project:	Transfer Station-SPDES
Contact:	Dean Devoe	Location:	L41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2025-01	001-WILLETS-PT-BLV D(APR)	Water			05/12/25			05/13/25
			Metals Group 10	200.7		05/13/25	05/14/25	
Q2025-02	002-35TH-AVE(APR)	Water			05/12/25			05/13/25
			Metals Group 10	200.7		05/13/25	05/14/25	