

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2025
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2025 **SAS No.:** Q2025
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	504	510	99	95 - 105	P	05/14/2025	13:24	LB135778
	Iron	10300	10000	104	95 - 105	P	05/14/2025	13:24	LB135778
	Lead	962	1000	96	95 - 105	P	05/14/2025	13:24	LB135778

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2025
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2025 **SAS No.:** Q2025
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.6	20.0	113	80 - 120	P	05/14/2025	13:34	LB135778
	Iron	93.7	100	94	80 - 120	P	05/14/2025	13:34	LB135778
	Lead	12.3	12.0	103	80 - 120	P	05/14/2025	13:34	LB135778

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc SDG No.: Q2025
Contract: TULL01 Lab Code: CHEM Case No.: Q2025 SAS No.: Q2025
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1230	1250	99	90 - 110	P	05/14/2025	14:51	LB135778
	Iron	4840	5000	97	90 - 110	P	05/14/2025	14:51	LB135778
	Lead	4840	5000	97	90 - 110	P	05/14/2025	14:51	LB135778
CCV02	Copper	1250	1250	100	90 - 110	P	05/14/2025	15:53	LB135778
	Iron	5010	5000	100	90 - 110	P	05/14/2025	15:53	LB135778
	Lead	4880	5000	98	90 - 110	P	05/14/2025	15:53	LB135778
CCV03	Copper	1220	1250	98	90 - 110	P	05/14/2025	16:44	LB135778
	Iron	4820	5000	96	90 - 110	P	05/14/2025	16:44	LB135778
	Lead	4800	5000	96	90 - 110	P	05/14/2025	16:44	LB135778
CCV04	Copper	1220	1250	98	90 - 110	P	05/14/2025	18:08	LB135778
	Iron	4990	5000	100	90 - 110	P	05/14/2025	18:08	LB135778
	Lead	4840	5000	97	90 - 110	P	05/14/2025	18:08	LB135778
CCV05	Copper	1200	1250	96	90 - 110	P	05/14/2025	19:21	LB135778
	Iron	4830	5000	97	90 - 110	P	05/14/2025	19:21	LB135778
	Lead	4830	5000	96	90 - 110	P	05/14/2025	19:21	LB135778

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2025
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2025 **SAS No.:** Q2025
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	508	510	100	95 - 105	P	05/15/2025	13:37	LB135794
	Iron	9530	10000	95	95 - 105	P	05/15/2025	13:37	LB135794
	Lead	989	1000	99	95 - 105	P	05/15/2025	13:37	LB135794

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2025
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2025 **SAS No.:** Q2025
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.6	20.0	113	80 - 120	P	05/15/2025	13:42	LB135794
	Iron	98.5	100	98	80 - 120	P	05/15/2025	13:42	LB135794
	Lead	12.2	12.0	102	80 - 120	P	05/15/2025	13:42	LB135794

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2025
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2025 **SAS No.:** Q2025
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	05/15/2025	14:18	LB135794
	Iron	4870	5000	97	90 - 110	P	05/15/2025	14:18	LB135794
	Lead	4990	5000	100	90 - 110	P	05/15/2025	14:18	LB135794
CCV02	Copper	1260	1250	101	90 - 110	P	05/15/2025	15:05	LB135794
	Iron	5010	5000	100	90 - 110	P	05/15/2025	15:05	LB135794
	Lead	5030	5000	100	90 - 110	P	05/15/2025	15:05	LB135794
CCV03	Copper	1240	1250	99	90 - 110	P	05/15/2025	15:53	LB135794
	Iron	4540	5000	91	90 - 110	P	05/15/2025	15:53	LB135794
	Lead	4930	5000	99	90 - 110	P	05/15/2025	15:53	LB135794
CCV04	Copper	1240	1250	99	90 - 110	P	05/15/2025	16:39	LB135794
	Iron	4680	5000	94	90 - 110	P	05/15/2025	16:39	LB135794
	Lead	4960	5000	99	90 - 110	P	05/15/2025	16:39	LB135794
CCV05	Copper	1240	1250	99	90 - 110	P	05/15/2025	17:27	LB135794
	Iron	4850	5000	97	90 - 110	P	05/15/2025	17:27	LB135794
	Lead	5580	5000	112	90 - 110	P	05/15/2025	17:27	LB135794
CCV06	Copper	1230	1250	98	90 - 110	P	05/15/2025	17:41	LB135794
	Iron	4860	5000	97	90 - 110	P	05/15/2025	17:41	LB135794
	Lead	5000	5000	100	90 - 110	P	05/15/2025	17:41	LB135794
CCV07	Copper	1210	1250	97	90 - 110	P	05/15/2025	18:47	LB135794
	Iron	4790	5000	96	90 - 110	P	05/15/2025	18:47	LB135794
	Lead	4940	5000	99	90 - 110	P	05/15/2025	18:47	LB135794