

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q2025
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2025 **SAS No.:** Q2025
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	0.58	2.0	29	-18	22	05/14/2025	14:09	LB135778
	Iron	93300	101000	92	85600	116500	05/14/2025	14:09	LB135778
	Lead	0.26			-12	12	05/14/2025	14:09	LB135778
ICSAB01	Copper	468	511	92	434	588	05/14/2025	14:16	LB135778
	Iron	94500	99300	95	84400	114500	05/14/2025	14:16	LB135778
	Lead	46.1	49.0	94	37	61	05/14/2025	14:16	LB135778
ICSA	Copper	2.79	2.0	140	-18	22	05/14/2025	14:43	LB135778
	Iron	88000	101000	87	85600	116500	05/14/2025	14:43	LB135778
	Lead	4.80			-12	12	05/14/2025	14:43	LB135778
ICSAB	Copper	516	511	101	434	588	05/14/2025	14:47	LB135778
	Iron	98200	99300	99	84400	114500	05/14/2025	14:47	LB135778
	Lead	23.8	49.0	49	37	61	05/14/2025	14:47	LB135778
ICSA01	Copper	-2.10	2.0	105	-18	22	05/15/2025	14:01	LB135794
	Iron	93300	101000	92	85600	116500	05/15/2025	14:01	LB135794
	Lead	1.30			-12	12	05/15/2025	14:01	LB135794
ICSAB01	Copper	453	511	89	434	588	05/15/2025	14:05	LB135794
	Iron	93800	99300	94	84400	114500	05/15/2025	14:05	LB135794
	Lead	45.4	49.0	93	37	61	05/15/2025	14:05	LB135794
ICSA	Copper	-3.98	2.0	199	-18	22	05/15/2025	14:09	LB135794
	Iron	86000	101000	85	85600	116500	05/15/2025	14:09	LB135794
	Lead	-2.62			-12	12	05/15/2025	14:09	LB135794
ICSAB	Copper	471	511	92	434	588	05/15/2025	14:14	LB135794
	Iron	85300	99300	86	84400	114500	05/15/2025	14:14	LB135794
	Lead	65.2	49.0	133	37	61	05/15/2025	14:14	LB135794