

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2025</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2025</u> sas no.: <u>Q2025</u>
matrix: <u>Water</u>	sample id: <u>Q2006-02</u>	client id: <u>EFF-WWMS</u>
Percent Solids for Sample: NA	Spiked ID: Q2006-02MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	254		90.1		150	109		P
Iron	ug/L	75 - 125	2340		397		1500	129	N	P
Lead	ug/L	75 - 125	563		1.77	J	500	112		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2025</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2025</u> sas no.: <u>Q2025</u>
matrix: <u>Water</u>	sample id: <u>Q2006-02</u>	client id: <u>EFF-WWMSD</u>
Percent Solids for Sample: NA	Spiked ID: Q2006-02MSD	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	243		90.1		150	102		P
Iron	ug/L	75 - 125	2010		397		1500	107		P
Lead	ug/L	75 - 125	529		1.77	J	500	105		P