



- 7 -

Client:	<u>Tully Environmental, Inc</u>	SDG No.:	<u>Q2025</u>				
Contract:	TULL01	Lab Code:	CHEM	Case No.:	Q2025	SAS No.:	Q2025

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167985BS							
Copper	ug/L	150	141		94	85 - 115	P
Iron	ug/L	1500	1410		94	85 - 115	P
Lead	ug/L	500	459		92	85 - 115	P