



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q2026
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2026 **SAS No.:** Q2026
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	22.3	20.0	112	65 - 135	P	05/14/2025	13:43	LB135778
	Iron	93.3	100	93	65 - 135	P	05/14/2025	13:43	LB135778
	Lead	11.1	12.0	92	65 - 135	P	05/14/2025	13:43	LB135778
CRI01	Copper	22.2	20.0	111	65 - 135	P	05/15/2025	13:51	LB135794
	Iron	104	100	104	65 - 135	P	05/15/2025	13:51	LB135794
	Lead	12.3	12.0	102	65 - 135	P	05/15/2025	13:51	LB135794