

**metals**  
**- 5a -**  
**MATRIX SPIKE SUMMARY**

|                                                |                                   |                                                            |
|------------------------------------------------|-----------------------------------|------------------------------------------------------------|
| <b>client:</b> <u>Tully Environmental, Inc</u> | <b>level:</b> <u>low</u>          | <b>sdg no.:</b> <u>Q2026</u>                               |
| <b>contract:</b> <u>TULL01</u>                 | <b>lab code:</b> <u>CHEM</u>      | <b>case no.:</b> <u>Q2026</u> <b>sas no.:</b> <u>Q2026</u> |
| <b>matrix:</b> <u>Water</u>                    | <b>sample id:</b> <u>Q2006-02</u> | <b>client id:</b> <u>EFF-WWMS</u>                          |
| <b>Percent Solids for Sample:</b> NA           | <b>Spiked ID:</b> Q2006-02MS      | <b>Percent Solids for Spike Sample:</b> NA                 |

| Analyte | Units | Acceptance<br>Limit %R | Spiked<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|------------------|---|------------------|---|----------------|---------------|------|---|
| Copper  | ug/L  | 75 - 125               | 254              |   | 90.1             |   | 150            | 109           |      | P |
| Iron    | ug/L  | 75 - 125               | 2340             |   | 397              |   | 1500           | 129           | N    | P |
| Lead    | ug/L  | 75 - 125               | 563              |   | 1.77             | J | 500            | 112           |      | P |

**metals**  
**- 5a -**  
**MATRIX SPIKE DUPLICATE SUMMARY**

|                                                |                                   |                                                            |
|------------------------------------------------|-----------------------------------|------------------------------------------------------------|
| <b>client:</b> <u>Tully Environmental, Inc</u> | <b>level:</b> <u>low</u>          | <b>sdg no.:</b> <u>Q2026</u>                               |
| <b>contract:</b> <u>TULL01</u>                 | <b>lab code:</b> <u>CHEM</u>      | <b>case no.:</b> <u>Q2026</u> <b>sas no.:</b> <u>Q2026</u> |
| <b>matrix:</b> <u>Water</u>                    | <b>sample id:</b> <u>Q2006-02</u> | <b>client id:</b> <u>EFF-WWMSD</u>                         |
| <b>Percent Solids for Sample:</b> NA           | <b>Spiked ID:</b> Q2006-02MSD     | <b>Percent Solids for Spike Sample:</b> NA                 |

| Analyte | Units | Acceptance<br>Limit %R | MSD<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|---------------|---|------------------|---|----------------|---------------|------|---|
| Copper  | ug/L  | 75 - 125               | 243           |   | 90.1             |   | 150            | 102           |      | P |
| Iron    | ug/L  | 75 - 125               | 2010          |   | 397              |   | 1500           | 107           |      | P |
| Lead    | ug/L  | 75 - 125               | 529           |   | 1.77             | J | 500            | 105           |      | P |