

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV **SDG No.:** Q2027
Contract: SCAL01 **Lab Code:** CHEM **Case No.:** Q2027 **SAS No.:** Q2027
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV10	Mercury	3.68	4.0	92	90 - 110	CV	05/19/2025	10:17	LB135815

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV **SDG No.:** Q2027

Contract: SCAL01 **Lab Code:** CHEM **Case No.:** Q2027 **SAS No.:** Q2027

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV35	Mercury	4.71	5.0	94	90 - 110	CV	05/19/2025	10:21	LB135815
CCV36	Mercury	4.59	5.0	92	90 - 110	CV	05/19/2025	10:56	LB135815
CCV37	Mercury	4.61	5.0	92	90 - 110	CV	05/19/2025	11:23	LB135815
CCV38	Mercury	4.92	5.0	98	90 - 110	CV	05/19/2025	12:06	LB135815
CCV39	Mercury	4.75	5.0	95	90 - 110	CV	05/19/2025	12:37	LB135815



- 2a -

Client:	<u>Scalamandre – Tully JV</u>	SDG No.:	<u>Q2027</u>	
Contract:	<u>SCAL01</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q2027</u> SAS No.: <u>Q2027</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
ICV01	Arsenic	992	1000	99	90 - 110	P	05/19/2025	12:28	LB135823
	Barium	493	520	95	90 - 110	P	05/19/2025	12:28	LB135823
	Cadmium	508	510	100	90 - 110	P	05/19/2025	12:28	LB135823
	Chromium	519	520	100	90 - 110	P	05/19/2025	12:28	LB135823
	Lead	998	1000	100	90 - 110	P	05/19/2025	12:28	LB135823
	Selenium	1010	1000	101	90 - 110	P	05/19/2025	12:28	LB135823
	Silver	232	250	93	90 - 110	P	05/19/2025	12:28	LB135823

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV **SDG No.:** Q2027

Contract: SCAL01 **Lab Code:** CHEM **Case No.:** Q2027 **SAS No.:** Q2027

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5120	5000	102	90 - 110	P	05/19/2025	13:15	LB135823
	Barium	10100	10000	101	90 - 110	P	05/19/2025	13:15	LB135823
	Cadmium	2540	2500	102	90 - 110	P	05/19/2025	13:15	LB135823
	Chromium	1020	1000	102	90 - 110	P	05/19/2025	13:15	LB135823
	Lead	5090	5000	102	90 - 110	P	05/19/2025	13:15	LB135823
	Selenium	5180	5000	104	90 - 110	P	05/19/2025	13:15	LB135823
	Silver	1280	1250	102	90 - 110	P	05/19/2025	13:15	LB135823
CCV02	Arsenic	5060	5000	101	90 - 110	P	05/19/2025	14:13	LB135823
	Barium	10200	10000	102	90 - 110	P	05/19/2025	14:13	LB135823
	Cadmium	2540	2500	101	90 - 110	P	05/19/2025	14:13	LB135823
	Chromium	1020	1000	102	90 - 110	P	05/19/2025	14:13	LB135823
	Lead	5070	5000	101	90 - 110	P	05/19/2025	14:13	LB135823
	Selenium	5100	5000	102	90 - 110	P	05/19/2025	14:13	LB135823
	Silver	1280	1250	102	90 - 110	P	05/19/2025	14:13	LB135823
CCV03	Arsenic	4900	5000	98	90 - 110	P	05/19/2025	15:14	LB135823
	Barium	9580	10000	96	90 - 110	P	05/19/2025	15:14	LB135823
	Cadmium	2430	2500	97	90 - 110	P	05/19/2025	15:14	LB135823
	Chromium	982	1000	98	90 - 110	P	05/19/2025	15:14	LB135823
	Lead	4880	5000	98	90 - 110	P	05/19/2025	15:14	LB135823
	Selenium	4960	5000	99	90 - 110	P	05/19/2025	15:14	LB135823
	Silver	1230	1250	98	90 - 110	P	05/19/2025	15:14	LB135823
CCV04	Arsenic	4940	5000	99	90 - 110	P	05/19/2025	16:03	LB135823
	Barium	9750	10000	98	90 - 110	P	05/19/2025	16:03	LB135823
	Cadmium	2450	2500	98	90 - 110	P	05/19/2025	16:03	LB135823
	Chromium	989	1000	99	90 - 110	P	05/19/2025	16:03	LB135823
	Lead	4880	5000	98	90 - 110	P	05/19/2025	16:03	LB135823
	Selenium	5030	5000	101	90 - 110	P	05/19/2025	16:03	LB135823
	Silver	1250	1250	100	90 - 110	P	05/19/2025	16:03	LB135823
CCV05	Arsenic	4890	5000	98	90 - 110	P	05/19/2025	16:47	LB135823
	Barium	9850	10000	98	90 - 110	P	05/19/2025	16:47	LB135823
	Cadmium	2440	2500	98	90 - 110	P	05/19/2025	16:47	LB135823
	Chromium	980	1000	98	90 - 110	P	05/19/2025	16:47	LB135823
	Lead	4870	5000	97	90 - 110	P	05/19/2025	16:47	LB135823
	Selenium	4990	5000	100	90 - 110	P	05/19/2025	16:47	LB135823

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV **SDG No.:** Q2027

Contract: SCAL01 **Lab Code:** CHEM **Case No.:** Q2027 **SAS No.:** Q2027

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1220	1250	98	90 - 110	P	05/19/2025	16:47	LB135823
CCV06	Arsenic	4920	5000	98	90 - 110	P	05/19/2025	17:35	LB135823
	Barium	9800	10000	98	90 - 110	P	05/19/2025	17:35	LB135823
	Cadmium	2480	2500	99	90 - 110	P	05/19/2025	17:35	LB135823
	Chromium	1000	1000	100	90 - 110	P	05/19/2025	17:35	LB135823
	Lead	4940	5000	99	90 - 110	P	05/19/2025	17:35	LB135823
	Selenium	5010	5000	100	90 - 110	P	05/19/2025	17:35	LB135823
	Silver	1240	1250	99	90 - 110	P	05/19/2025	17:35	LB135823
CCV07	Arsenic	4960	5000	99	90 - 110	P	05/19/2025	18:21	LB135823
	Barium	9920	10000	99	90 - 110	P	05/19/2025	18:21	LB135823
	Cadmium	2460	2500	99	90 - 110	P	05/19/2025	18:21	LB135823
	Chromium	987	1000	99	90 - 110	P	05/19/2025	18:21	LB135823
	Lead	4910	5000	98	90 - 110	P	05/19/2025	18:21	LB135823
	Selenium	5060	5000	101	90 - 110	P	05/19/2025	18:21	LB135823
	Silver	1240	1250	99	90 - 110	P	05/19/2025	18:21	LB135823
CCV08	Arsenic	4990	5000	100	90 - 110	P	05/19/2025	19:09	LB135823
	Barium	9660	10000	97	90 - 110	P	05/19/2025	19:09	LB135823
	Cadmium	2470	2500	99	90 - 110	P	05/19/2025	19:09	LB135823
	Chromium	985	1000	98	90 - 110	P	05/19/2025	19:09	LB135823
	Lead	4910	5000	98	90 - 110	P	05/19/2025	19:09	LB135823
	Selenium	5060	5000	101	90 - 110	P	05/19/2025	19:09	LB135823
	Silver	1240	1250	99	90 - 110	P	05/19/2025	19:09	LB135823
CCV09	Arsenic	4930	5000	99	90 - 110	P	05/19/2025	19:55	LB135823
	Barium	9570	10000	96	90 - 110	P	05/19/2025	19:55	LB135823
	Cadmium	2430	2500	97	90 - 110	P	05/19/2025	19:55	LB135823
	Chromium	998	1000	100	90 - 110	P	05/19/2025	19:55	LB135823
	Lead	4870	5000	98	90 - 110	P	05/19/2025	19:55	LB135823
	Selenium	4950	5000	99	90 - 110	P	05/19/2025	19:55	LB135823
	Silver	1240	1250	99	90 - 110	P	05/19/2025	19:55	LB135823
CCV10	Arsenic	4950	5000	99	90 - 110	P	05/19/2025	20:11	LB135823
	Barium	9450	10000	94	90 - 110	P	05/19/2025	20:11	LB135823
	Cadmium	2440	2500	98	90 - 110	P	05/19/2025	20:11	LB135823
	Chromium	984	1000	98	90 - 110	P	05/19/2025	20:11	LB135823
	Lead	4880	5000	98	90 - 110	P	05/19/2025	20:11	LB135823

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Scalamandre – Tully JV **SDG No.:** Q2027
Contract: SCAL01 **Lab Code:** CHEM **Case No.:** Q2027 **SAS No.:** Q2027
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Selenium	4990	5000	100	90 - 110	P	05/19/2025	20:11	LB135823
	Silver	1220	1250	98	90 - 110	P	05/19/2025	20:11	LB135823