



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q2027

Contract: SCAL01

Lab Code: CHEM

Case No.: Q2027

SAS No.: Q2027

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB10	Mercury	0.20	+/-0.20	U	0.20	CV	05/19/2025	10:19	LB135815

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q2027

Contract: SCAL01

Lab Code: CHEM

Case No.: Q2027

SAS No.: Q2027

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB35	Mercury	0.20	+/-0.20	U	0.20	CV	05/19/2025	10:24	LB135815
CCB36	Mercury	0.20	+/-0.20	U	0.20	CV	05/19/2025	10:58	LB135815
CCB37	Mercury	0.20	+/-0.20	U	0.20	CV	05/19/2025	11:28	LB135815
CCB38	Mercury	0.20	+/-0.20	U	0.20	CV	05/19/2025	12:11	LB135815
CCB39	Mercury	0.20	+/-0.20	U	0.20	CV	05/19/2025	12:39	LB135815

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Scalamandre – Tully JV</u>		SDG No.: <u>Q2027</u>							
Contract:	<u>SCAL01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2027</u>	SAS No.:	<u>Q2027</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	12:44	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	12:44	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	12:44	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	12:44	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	12:44	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	12:44	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	12:44	LB135823

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV		SDG No.: Q2027							
Contract:	SCAL01	Lab Code:	CHEM	Case No.:	Q2027	SAS No.:	Q2027		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	13:23	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	13:23	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	13:23	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	13:23	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	13:23	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	13:23	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	13:23	LB135823
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	14:24	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	14:24	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	14:24	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	14:24	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	14:24	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	14:24	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	14:24	LB135823
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	15:19	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	15:19	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	15:19	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	15:19	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	15:19	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	15:19	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	15:19	LB135823
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	16:07	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	16:07	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	16:07	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	16:07	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	16:07	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	16:07	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	16:07	LB135823
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	16:51	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	16:51	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	16:51	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	16:51	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	16:51	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	16:51	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	16:51	LB135823
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	17:39	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	17:39	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	17:39	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	17:39	LB135823

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV		SDG No.: Q2027							
Contract:	SCAL01	Lab Code:	CHEM	Case No.:	Q2027	SAS No.:	Q2027		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	17:39	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	17:39	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	17:39	LB135823
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	18:25	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	18:25	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	18:25	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	18:25	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	18:25	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	18:25	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	18:25	LB135823
	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	19:14	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	19:14	LB135823
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	19:14	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	19:14	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	19:14	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	19:14	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	19:14	LB135823
	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	19:59	LB135823
CCB09	Barium	100	+/-100	U	100	P	05/19/2025	19:59	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	19:59	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	19:59	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	19:59	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	19:59	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	19:59	LB135823
CCB10	Arsenic	20.0	+/-20.0	U	20.0	P	05/19/2025	20:15	LB135823
	Barium	100	+/-100	U	100	P	05/19/2025	20:15	LB135823
	Cadmium	6.00	+/-6.00	U	6.00	P	05/19/2025	20:15	LB135823
	Chromium	10.0	+/-10.0	U	10.0	P	05/19/2025	20:15	LB135823
	Lead	12.0	+/-12.0	U	12.0	P	05/19/2025	20:15	LB135823
	Selenium	20.0	+/-20.0	U	20.0	P	05/19/2025	20:15	LB135823
	Silver	10.0	+/-10.0	U	10.0	P	05/19/2025	20:15	LB135823