

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q2027 SAS No.: Q2027 SDG NO.: Q2027

Continuing Calib Date: 05/15/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:32 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1242-2 (2)	5.68	5.69	5.59	5.79	0.01
Aroclor-1242-3 (3)	5.74	5.75	5.65	5.85	0.01
Aroclor-1242-4 (4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1242-5 (5)	6.57	6.58	6.48	6.68	0.01
Tetrachloro-m-xylene	4.51	4.51	4.41	4.61	0.00
Decachlorobiphenyl	10.21	10.24	10.14	10.34	0.03

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q2027 SAS No.: Q2027 SDG NO.: Q2027

Continuing Calib Date: 05/15/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:32 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.88	4.90	4.80	5.00	0.02
Aroclor-1242-2	(2)	4.90	4.92	4.82	5.02	0.02
Aroclor-1242-3	(3)	5.08	5.09	4.99	5.19	0.01
Aroclor-1242-4	(4)	5.16	5.18	5.08	5.28	0.02
Aroclor-1242-5	(5)	5.69	5.70	5.60	5.80	0.01
Tetrachloro-m-xylene		3.80	3.81	3.71	3.91	0.01
Decachlorobiphenyl		8.82	8.85	8.75	8.95	0.03