

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q2027 SAS No.: Q2027 SDG NO.: Q2027

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: AR1660CCC500 Data File : PP072033.D Time Analyzed: 09:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.658	5.571	5.771	471.680	500.000	-5.7
Aroclor-1016-2	5.681	5.593	5.793	494.270	500.000	-1.1
Aroclor-1016-3	5.743	5.654	5.854	486.580	500.000	-2.7
Aroclor-1016-4	5.841	5.752	5.952	484.080	500.000	-3.2
Aroclor-1016-5	6.133	6.045	6.245	471.780	500.000	-5.6
Aroclor-1260-1	7.251	7.165	7.365	460.330	500.000	-7.9
Aroclor-1260-2	7.505	7.418	7.618	471.860	500.000	-5.6
Aroclor-1260-3	7.864	7.777	7.977	498.960	500.000	-0.2
Aroclor-1260-4	8.088	8.002	8.202	480.940	500.000	-3.8
Aroclor-1260-5	8.407	8.322	8.522	510.670	500.000	2.1
Decachlorobiphenyl	10.216	10.137	10.337	52.760	50.000	5.5
Tetrachloro-m-xylene	4.507	4.417	4.617	50.010	50.000	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q2027 SAS No.: Q2027 SDG NO.: Q2027

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: AR1660CCC500 Data File : PP072033.D Time Analyzed: 09:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.884	4.798	4.998	504.730	500.000	0.9
Aroclor-1016-2	4.902	4.816	5.016	523.740	500.000	4.7
Aroclor-1016-3	5.080	4.993	5.193	478.760	500.000	-4.2
Aroclor-1016-4	5.121	5.036	5.236	494.260	500.000	-1.1
Aroclor-1016-5	5.335	5.250	5.450	521.060	500.000	4.2
Aroclor-1260-1	6.368	6.285	6.485	529.420	500.000	5.9
Aroclor-1260-2	6.556	6.473	6.673	530.290	500.000	6.1
Aroclor-1260-3	6.709	6.627	6.827	531.350	500.000	6.3
Aroclor-1260-4	7.179	7.098	7.298	535.400	500.000	7.1
Aroclor-1260-5	7.421	7.339	7.539	536.070	500.000	7.2
Decachlorobiphenyl	8.824	8.749	8.949	55.520	50.000	11.0
Tetrachloro-m-xylene	3.801	3.712	3.912	52.400	50.000	4.8