

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q2027 SAS No.: Q2027 SDG NO.: Q2027

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030288.D Time Analyzed: 21:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.848	8.750	8.950	717.600	712.500	0.7
2,4-D	8.006	7.910	8.110	703.830	705.000	-0.2
2,4-DCAA	6.931	6.834	7.034	740.950	750.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q2027 SAS No.: Q2027 SDG NO.: Q2027

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030288.D Time Analyzed: 21:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.517	9.423	9.623	729.750	712.500	2.4
2,4-D	8.639	8.544	8.744	708.970	705.000	0.6
2,4-DCAA	7.450	7.354	7.554	739.160	750.000	-1.4