

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: SCAL01

Lab Code: CHEM **Case No.:** Q2027 **SAS No.:** Q2027 **SDG NO.:** Q2027

Instrument ID: ECD_D **Date(s) Analyzed:** 05/19/2025 05/19/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.72	4.62	4.82	175277000
		2	5.24	5.14	5.34	179454000
		3	5.95	5.85	6.05	741371000
		4	6.03	5.93	6.13	889415000
		5	6.87	6.77	6.97	150936000
Toxaphene	500	1	6.24	6.14	6.34	26389700
		2	6.44	6.34	6.54	38119600
		3	7.15	7.05	7.25	72288300
		4	7.57	7.47	7.67	91412700
		5	7.93	7.83	8.03	52259600

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: SCAL01

Lab Code: CHEM **Case No.:** Q2027 **SAS No.:** Q2027 **SDG NO.:** Q2027

Instrument ID: ECD_D **Date(s) Analyzed:** 05/19/2025 05/19/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.91	3.81	4.01	813076000
		2	4.49	4.39	4.59	822692000
		3	5.13	5.03	5.23	2571520000
		4	5.19	5.09	5.29	2143050000
		5	6.09	5.99	6.19	983052000
Toxaphene	500	1	5.48	5.38	5.58	139231000
		2	5.65	5.55	5.75	94078800
		3	6.76	6.66	6.86	437083000
		4	7.20	7.10	7.30	311285000
		5	7.33	7.23	7.43	216975000