

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2027

Analytical Method: 8081B

Client: Scalamandre – Tully JV

Datafile : PD088633.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168066BS (Column 1)	gamma-BHC (Lindane)	5	5.30	ug/L	106				82	129		
	Heptachlor	5	5.30	ug/L	106				79	127		
	Heptachlor epoxide	5	5.40	ug/L	108				81	124		
	Endrin	5	5.30	ug/L	106				81	128		
	Methoxychlor	5	4.60	ug/L	92				78	108		
PB168066BS (Column 2)	gamma-BHC (Lindane)	5	5.30	ug/L	106				82	129		
	Heptachlor	5	5.10	ug/L	102				79	127		
	Heptachlor epoxide	5	5.30	ug/L	106				81	124		
	Endrin	5	5.20	ug/L	104				81	128		
	Methoxychlor	5	4.60	ug/L	92				78	108		