

Surrogate Summary

SW-846

SDG No.: Q2027

Client: Scalamandre – Tully JV

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167994TB	PB167994TB	2-Fluorophenol	150	134	89		10	139
		Phenol-d6	150	129	86		10	134
		Nitrobenzene-d5	100	74.2	74		49	133
		2-Fluorobiphenyl	100	72.7	73		52	132
		2,4,6-Tribromophenol	150	135	90		44	137
		Terphenyl-d14	100	83.3	83		48	125
PB168026BL	PB168026BL	2-Fluorophenol	150	132	88		10	139
		Phenol-d6	150	122	82		10	134
		Nitrobenzene-d5	100	76.1	76		49	133
		2-Fluorobiphenyl	100	76.4	76		52	132
		2,4,6-Tribromophenol	150	124	83		44	137
		Terphenyl-d14	100	78.6	79		48	125
PB168026BS	PB168026BS	2-Fluorophenol	150	127	85		10	139
		Phenol-d6	150	117	78		10	134
		Nitrobenzene-d5	100	68.3	68		49	133
		2-Fluorobiphenyl	100	76.9	77		52	132
		2,4,6-Tribromophenol	150	140	93		44	137
		Terphenyl-d14	100	79.2	79		48	125
Q2019-04MS	MH-KMS	2-Fluorophenol	150	127	85		10	139
		Phenol-d6	150	115	77		10	134
		Nitrobenzene-d5	100	76.4	76		49	133
		2-Fluorobiphenyl	100	76.4	76		52	132
		2,4,6-Tribromophenol	150	131	87		44	137
		Terphenyl-d14	100	78.6	79		48	125
Q2019-04MSD	MH-KMSD	2-Fluorophenol	150	128	85		10	139
		Phenol-d6	150	115	77		10	134
		Nitrobenzene-d5	100	78.9	79		49	133
		2-Fluorobiphenyl	100	77.4	77		52	132
		2,4,6-Tribromophenol	150	140	93		44	137
		Terphenyl-d14	100	85.1	85		48	125
Q2027-03	B27-SOIL-SAMPLE	2-Fluorophenol	150	130	87		10	139
		Phenol-d6	150	110	73		10	134
		Nitrobenzene-d5	100	83.6	84		49	133
		2-Fluorobiphenyl	100	81.1	81		52	132
		2,4,6-Tribromophenol	150	157	105		44	137
		Terphenyl-d14	100	88.9	89		48	125
Q2027-04	B28-SOIL-SAMPLE	2-Fluorophenol	150	130	87		10	139
		Phenol-d6	150	113	75		10	134
		Nitrobenzene-d5	100	82.6	83		49	133
		2-Fluorobiphenyl	100	80.9	81		52	132
		2,4,6-Tribromophenol	150	171	114		44	137
		Terphenyl-d14	100	85.8	86		48	125