

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2027

Client: Scalamandre – Tully JV

Analytical Method: 8270E

DataFile: BP024740.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168026BS	Pyridine	50	34.6	ug/L	69				29	97		
	1,4-Dichlorobenzene	50	41.8	ug/L	84				76	103		
	2-Methylphenol	50	40.2	ug/L	80				69	109		
	3+4-Methylphenols	50	39.6	ug/L	79				67	106		
	Hexachloroethane	50	39.0	ug/L	78				76	118		
	Nitrobenzene	50	39.0	ug/L	78				58	106		
	Hexachlorobutadiene	50	47.3	ug/L	95				69	101		
	2,4,6-Trichlorophenol	50	48.5	ug/L	97				61	110		
	2,4,5-Trichlorophenol	50	47.2	ug/L	94				70	106		
	2,4-Dinitrotoluene	50	45.4	ug/L	91				60	115		
	Hexachlorobenzene	50	46.4	ug/L	93				73	106		
	Pentachlorophenol	100	110	ug/L	110				47	114		