

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2027

Client: Scalamandre – Tully JV

Analytical Method: SW8260D

Datafile : VX046232.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0516WBSD01	Vinyl chloride	20	17.4	ug/L	87	6		65	117	20
	1,1-Dichloroethene	20	18.1	ug/L	91	2		74	110	20
	2-Butanone	100	110	ug/L	110	10		65	122	20
	Carbon Tetrachloride	20	19.2	ug/L	96	3		77	113	20
	Chloroform	20	19.8	ug/L	99	2		79	113	20
	Benzene	20	20.0	ug/L	100	1		82	109	20
	1,2-Dichloroethane	20	20.9	ug/L	104	1		80	115	20
	Trichloroethene	20	19.7	ug/L	99	0		77	113	20
	Tetrachloroethene	20	20.2	ug/L	101	3		67	123	20
	Chlorobenzene	20	19.5	ug/L	98	1		82	109	20