

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCAL01
Lab Code: CHEM Case No.: Q2027 SAS No.: Q2027 SDG NO.: Q2027
Lab File ID: VX046228.D Date Analyzed: 05/16/2025
Instrument ID: MSVOA_X Time Analyzed: 09:22
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	94540	5.54	157316	6.75	136812	10.05
UPPER LIMIT	189080	6.038	314632	7.251	273624	10.549
LOWER LIMIT	47270	5.038	78658	6.251	68406	9.549
EPA SAMPLE NO.						
B27-SOIL-SAMPLE	58179	5.55	115702	6.76	111038	10.05
B28-SOIL-SAMPLE	65180	5.54	127503	6.76	120884	10.06
VX0516WBL01	63217	5.54	126396	6.76	119684	10.05
VX0516WBS01	89369	5.54	154326	6.76	137068	10.05
VX0516WBSD01	88353	5.54	151068	6.76	134956	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.