

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2031
Project:	Linden Generating Station - LRSA	Sample ID:	Q2031-05
Client ID:	DSN003MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	71.4		28.6		50.0	1	86		05/14/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2031
Project:	Linden Generating Station - LRSA	Sample ID:	Q2031-05
Client ID:	DSN003MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	76.5		28.6		50.0	1	96		05/14/2025