

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2031
Project:	Linden Generating Station - LRSA	Sample ID:	Q2031-02
Client ID:	DSN002DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Field pH	pH	+/-20	7.24		7.24		1	0		05/13/2025

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2031
Project:	Linden Generating Station - LRSA	Sample ID:	Q2031-04
Client ID:	DSN001DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Field pH	pH	+/-20	6.96		6.95		1	0.14		05/13/2025

Duplicate Sample Summary

Client: Parkway Generation Operating LLC	SDG No.: Q2031
Project: Linden Generating Station - LRSA	Sample ID: Q2031-05
Client ID: DSN003DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	28.6		29.6		1	3.44		05/14/2025
TSS	mg/L	+/-5	5.80		5.90		1	1.71		05/16/2025

Duplicate Sample Summary

Client: Parkway Generation Operating LLC	SDG No.: Q2031
Project: Linden Generating Station - LRSA	Sample ID: Q2031-05
Client ID: DSN003MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	71.4		76.5		1	6.9		05/14/2025

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2031
Project:	Linden Generating Station - LRSA	Sample ID:	Q2031-06
Client ID:	DSN003DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Field pH	pH	+/-20	6.86		6.87		1	0.15		05/13/2025

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2031
Project:	Linden Generating Station - LRSA	Sample ID:	Q2036-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	623		641		1	2.87		05/14/2025