

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2031

Project: Linden Generating Station - LRSA

Run No.: LB135771

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135771BS							
COD	mg/L	50	52.0		104	1	90-110	05/14/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2031

Project: Linden Generating Station - LRSA

Run No.: LB135774

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135774BS							
BOD5	mg/L	198	201		102	1	84.6-115.4	05/14/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2031

Project: Linden Generating Station - LRSA

Run No.: LB135797

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135797BS							
TSS	mg/L	550	532		97	1	90-110	05/16/2025