



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

CDM Smith

Project Name: South River WM Replacement

Project # N/A

Order ID # Q2032

Test Name: Mercury, Metals ICP-TAL

A. Number of Samples and Date of Receipt:

9 Solid samples were received on 05/13/2025.

2 Water samples were received on 05/13/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Diesel Range Organics, Gasoline Range Organics, Herbicide, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, PCB, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for Mercury, Metals ICP-TAL.

C. Analytical Techniques:

The analysis of Metals ICP-TAL was based on method 6010D, digestion based on method 3050 (soils) and 3010 (waters). The analysis and digestion of Mercury was based on method 7470A. The analysis and digestion of Mercury was based on method 7471B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate (3859MSD) analysis met criteria for all samples except for Mercury due to matrix interference.

The Duplicate (TP-24DUP) analysis met criteria for all samples except for Lead due to matrix interference.

The Duplicate (TP-24MSD) analysis met criteria for all samples except for Aluminum, Calcium, Iron, Magnesium, Manganese, Sodium, Vanadium due to matrix interference.

The Matrix Spike (3859MS) analysis met criteria for all samples except for Mercury due to matrix interference.

The Matrix Spike (TP-24MS) analysis met criteria for all samples except for Antimony, Barium, Magnesium, Potassium, Selenium, Silver, Sodium, Vanadium due to matrix interference.



The Matrix Spike (TW-WTS-08MS) analysis met criteria for all samples except for Iron, Silver due to matrix interference.

The Matrix Spike Duplicate (3859MSD) analysis met criteria for all samples except for Mercury due to matrix interference.

The Matrix Spike Duplicate (TP-24MSD) analysis met criteria for all samples except for Antimony, Magnesium, Manganese, Potassium, Vanadium, Zinc due to matrix interference.

The Matrix Spike Duplicate (TW-WTS-08MSD) analysis met criteria for all samples except for Silver due to matrix interference.

The Post Digest Spike (TP-24A) analysis met criteria for all samples except for Magnesium, Manganese and Vanadium due to unknown chemical interferences of matrix with the addition of spike amount after digestion and before analysis , matrix has suppression effect during addition of spike

The Post Digest Spike (TW-WTS-08A) analysis met criteria for all samples except for Iron and Silver due to unknown chemical interferences of matrix with the addition of spike amount after digestion and before analysis , matrix has suppression effect during addition of spike

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met criteria for all samples.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____