

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

LAB CHRONICLE

OrderID:	Q2032	OrderDate:	5/13/2025 4:01:00 PM
Client:	CDM Smith	Project:	South River WM Replacement
Contact:	Marcie Ann Encinas	Location:	L41,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2032-09	COMP-1	SOIL			05/13/25 14:05			05/13/25
			Corrosivity	9045D			05/14/25 15:50	
			Ignitability	1030			05/15/25 11:40	
			Reactive Cyanide	9012B		05/16/25	05/16/25 13:08	
			Reactive Sulfide	9034		05/15/25	05/15/25 11:25	



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	05/13/25 14:05
Project:	South River WM Replacement	Date Received:	05/13/25
Client Sample ID:	COMP-1	SDG No.:	Q2032
Lab Sample ID:	Q2032-09	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	5.12	H	1	0	0	pH		05/14/25 15:50	9045D
Ignitability	NO		1	0	0	oC		05/15/25 11:40	1030
Reactive Cyanide	0.010	J	1	0.0083	0.050	mg/Kg	05/16/25 08:45	05/16/25 13:08	9012B
Reactive Sulfide	6.34	J	1	0.20	10.0	mg/Kg	05/15/25 08:50	05/15/25 11:25	9034

Comments: pH result reported at temperature 21.5 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: CDM Smith

SDG No.: Q2032

Project: South River WM Replacement

RunNo.: LB135777

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Corrosivity		pH	7.01	7	100	90-110	05/14/2025
Sample ID: CCV1 Corrosivity		pH	2.01	2.00	101	90-110	05/14/2025
Sample ID: CCV2 Corrosivity		pH	12.02	12.00	100	90-110	05/14/2025
Sample ID: CCV3 Corrosivity		pH	2.01	2.00	101	90-110	05/14/2025

Initial and Continuing Calibration Verification

Client: CDM Smith

SDG No.: Q2032

Project: South River WM Replacement

RunNo.: LB135800

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.092	0.099	93	85-115	05/16/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/16/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/16/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/16/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/16/2025

Initial and Continuing Calibration Verification

Client: CDM Smith

SDG No.: Q2032

Project: South River WM Replacement

RunNo.: LB135800

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------

Initial and Continuing Calibration Blank Summary

Client: CDM Smith

SDG No.: Q2032

Project: South River WM Replacement

RunNo.: LB135800

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	05/16/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	05/16/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.001	0.0025	J	0.00096	0.005	05/16/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	05/16/2025
Sample ID: CCB4 Reactive Cyanide	mg/L	0.001	0.0025	J	0.00096	0.005	05/16/2025

Initial and Continuing Calibration Blank Summary

Client: CDM Smith

SDG No.: Q2032

Project: South River WM Replacement

RunNo.: LB135800

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------

Preparation Blank Summary

Client: CDM Smith

SDG No.: Q2032

Project: South River WM Replacement

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: PB168011BL							
Reactive Sulfide	mg/Kg	< 5.0000	5.0000	U	0.201	10	05/15/2025
Sample ID: PB168027BL							
Reactive Cyanide	mg/Kg	0.011	0.0250	J	0.0084	0.05	05/16/2025

Duplicate Sample Summary

Client: CDM Smith Project: South River WM Replacement Client ID: MH-IDUP	SDG No.: Q2032 Sample ID: Q2017-04 Percent Solids for Spike Sample: 100
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Reactive Sulfide	mg/Kg	+/-20	3.17	J	3.17	J	1	0		05/15/2025

Duplicate Sample Summary

Client:	CDM Smith	SDG No.:	Q2032
Project:	South River WM Replacement	Sample ID:	Q2019-04
Client ID:	MH-KDUP	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Corrosivity	pH	+/-20	8.10		8.11		1	0.12		05/14/2025

Duplicate Sample Summary

Client: CDM Smith	SDG No.: Q2032
Project: South River WM Replacement	Sample ID: Q2032-09
Client ID: COMP-1DUP	Percent Solids for Spike Sample: 100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		05/15/2025
Reactive Cyanide	mg/Kg	+/-20	0.010	J	0.011	J	1	10		05/16/2025

Duplicate Sample Summary

Client: CDM Smith Project: South River WM Replacement Client ID: L2-WC-4DUP	SDG No.: Q2032 Sample ID: Q2048-16 Percent Solids for Spike Sample: 100
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		05/15/2025



RAW DATA

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135777

Slope : 97.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/14/2025	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	05/14/2025	15:01
3	CAL3	1	Water	NA	NA	20.2	10.02	05/14/2025	15:03
4	ICV	1	Water	NA	NA	20.2	7.01	05/14/2025	15:05
5	CCV1	1	Water	NA	NA	20.2	2.01	05/14/2025	15:10
6	Q2019-04	1	Solid	20.02	20	22.7	8.10	05/14/2025	15:20
7	Q2019-04DUP	1	Solid	20.03	20	22.8	8.11	05/14/2025	15:21
8	Q2020-04	1	Solid	20.04	20	22.7	8.41	05/14/2025	15:25
9	Q2027-03	1	Solid	20.03	20	22.7	10.94	05/14/2025	15:35
10	Q2027-04	1	Solid	20.05	20	21.9	11.52	05/14/2025	15:44
11	Q2032-09	1	Solid	20.02	20	21.5	5.12	05/14/2025	15:50
12	Q2034-04	1	Solid	20.04	20	20.9	5.38	05/14/2025	16:00
13	Q2034-08	1	Solid	20.03	20	21.4	6.94	05/14/2025	16:05
14	Q2034-12	1	Solid	20.04	20	21.4	7.50	05/14/2025	16:11
15	Q2034-16	1	Solid	20.03	20	20.8	5.60	05/14/2025	16:15
16	CCV2	1	Water	NA	NA	20.3	12.02	05/14/2025	16:22
17	Q2034-20	1	Solid	20.04	20	20.7	5.15	05/14/2025	16:30
18	Q2034-24	1	Solid	20.02	20	20.6	7.00	05/14/2025	16:33
19	Q2038-02	1	Solid	20.03	20	21.3	7.23	05/14/2025	16:40
20	Q2048-04	1	Solid	20.04	20	21.8	6.35	05/14/2025	16:44

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	An
21	Q2048-08	1	Solid	20.01	20	21.9	5.98	05/14/2025	16:50
22	Q2048-12	1	Solid	20.03	20	21.8	6.25	05/14/2025	17:00
23	Q2048-16	1	Solid	20.04	20	21.9	6.28	05/14/2025	17:10
24	CCV3	1	Water	NA	NA	20.2	2.01	05/14/2025	17:15

WORKLIST(Hardcopy Internal Chain)

135777

WorkList Name : corrosivity q2032

WorkList ID : 189512

Department : Wet-Chemistry

Date : 05-14-2025 10:43:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2019-04	MH-K	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/13/2025	9045D
Q2020-04	TP-A	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/12/2025	9045D
Q2027-03	B27-SOIL-SAMPLE	Solid	Corrosivity	Cool 4 deg C	SCAL01	L41	05/12/2025	9045D
Q2027-04	B28-SOIL-SAMPLE	Solid	Corrosivity	Cool 4 deg C	SCAL01	L41	05/12/2025	9045D
Q2032-09	COMP-1	Solid	Corrosivity	Cool 4 deg C	CAMP02	L41	05/13/2025	9045D
Q2034-04	L3-WC-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/13/2025	9045D
Q2034-08	L3-WC-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/13/2025	9045D
Q2034-12	L3-WC-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/13/2025	9045D
Q2034-16	L3-WC-4	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/13/2025	9045D
Q2034-20	L3-WC-5	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/13/2025	9045D
Q2034-24	L3-WC-6	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/13/2025	9045D
Q2038-02	72-11991	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	05/14/2025	9045D
Q2048-04	L2-WC-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/14/2025	9045D
Q2048-08	L2-WC-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/14/2025	9045D
Q2048-12	L2-WC-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/14/2025	9045D
Q2048-16	L2-WC-4	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/14/2025	9045D

Date/Time 05-14-25 14:35

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05-14-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Analysis Method: 9034

ANALYST: rubina

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB135785

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3114
Starch Solution, 4L	W3149

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB168011BL		1	5.00	50	2.00	0.00	1.94	1.94	0.06	0.00	0.00	05/15/2025	11:03
2	Q2017-04		1	5.04	50	2.00	0.00	1.90	1.90	0.10	0.04	3.17	05/15/2025	11:06
3	Q2017-04DUP		1	5.04	50	2.00	0.00	1.90	1.90	0.10	0.04	3.17	05/15/2025	11:09
4	Q2017-08		1	5.01	50	2.00	0.00	1.88	1.88	0.12	0.06	4.79	05/15/2025	11:11
5	Q2019-04		1	5.01	50	2.00	0.00	1.90	1.90	0.10	0.04	3.19	05/15/2025	11:14
6	Q2020-04		1	5.04	50	2.00	0.00	1.88	1.88	0.12	0.06	4.76	05/15/2025	11:17
7	Q2027-03		1	5.02	50	2.00	0.00	1.86	1.86	0.14	0.08	6.37	05/15/2025	11:20
8	Q2027-04		1	5.07	50	2.00	0.00	1.90	1.90	0.10	0.04	3.16	05/15/2025	11:22
9	Q2032-09		1	5.05	50	2.00	0.00	1.86	1.86	0.14	0.08	6.34	05/15/2025	11:25
10	Q2034-04		1	5.03	50	2.00	0.00	1.90	1.90	0.10	0.04	3.18	05/15/2025	11:28
11	Q2034-08		1	5.06	50	2.00	0.00	1.86	1.86	0.14	0.08	6.32	05/15/2025	11:30
12	Q2034-12		1	5.02	50	2.00	0.00	1.90	1.90	0.10	0.04	3.19	05/15/2025	11:32

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB135785

ANALYST: rubina

SUPERVISOR REVIEW BY: Iwona

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3114
Starch Solution, 4L	W3149

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
13	Q2034-16		1	5.04	50	2.00	0.00	1.86	1.86	0.14	0.08	6.35	05/15/2025	11:34
14	Q2034-20		1	5.03	50	2.00	0.00	1.88	1.88	0.12	0.06	4.77	05/15/2025	11:36
15	Q2034-24		1	5.01	50	2.00	0.00	1.90	1.90	0.10	0.04	3.19	05/15/2025	11:39
16	Q2038-02		1	5.05	50	2.00	0.00	1.90	1.90	0.10	0.04	3.17	05/15/2025	11:42
17	Q2048-04		1	5.04	50	2.00	0.00	1.88	1.88	0.12	0.06	4.76	05/15/2025	11:45
18	Q2048-08		1	5.01	50	2.00	0.00	1.86	1.86	0.14	0.08	6.39	05/15/2025	11:48
19	Q2048-12		1	5.06	50	2.00	0.00	1.90	1.90	0.10	0.04	3.16	05/15/2025	11:50
20	Q2048-16		1	5.06	50	2.00	0.00	1.90	1.90	0.10	0.04	3.16	05/15/2025	11:53

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume

Analytical Summary Report

Analysis Method: 1030
Parameter: Ignitability
Run Number: LB135790

Reviewed By: Iwona

Supervisor Review By: jignesh

Seq	LabID	ClientID	DF	matrix	Result Status	Burning Rate	Anal Date	Anal Time
1	Q2032-09	COMP-1	1	Solid	NO	0.00	05/15/2025	11:40
2	Q2032-09DUP	COMP-1DUP	1	Solid	NO	0.00	05/15/2025	11:48
3	Q2034-01	L3-WC-1	1	Solid	NO	0.00	05/15/2025	11:55
4	Q2034-04	L3-WC-1	1	Solid	NO	0.00	05/15/2025	12:03
5	Q2034-05	L3-WC-2	1	Solid	NO	0.00	05/15/2025	12:10
6	Q2034-08	L3-WC-2	1	Solid	NO	0.00	05/15/2025	12:17
7	Q2034-09	L3-WC-3	1	Solid	NO	0.00	05/15/2025	12:25
8	Q2034-12	L3-WC-3	1	Solid	NO	0.00	05/15/2025	12:32
9	Q2034-13	L3-WC-4	1	Solid	NO	0.00	05/15/2025	12:40
10	Q2034-16	L3-WC-4	1	Solid	NO	0.00	05/15/2025	12:47
11	Q2034-17	L3-WC-5	1	Solid	NO	0.00	05/15/2025	12:54
12	Q2034-20	L3-WC-5	1	Solid	NO	0.00	05/15/2025	13:00
13	Q2034-21	L3-WC-6	1	Solid	NO	0.00	05/15/2025	13:07
14	Q2034-24	L3-WC-6	1	Solid	NO	0.00	05/15/2025	13:15
15	Q2038-01	72-11991	1	Solid	NO	0.00	05/15/2025	13:22
16	Q2038-02	72-11991	1	Solid	NO	0.00	05/15/2025	13:30
17	Q2048-01	L2-WC-1	1	Solid	NO	0.00	05/15/2025	13:38
18	Q2048-04	L2-WC-1	1	Solid	NO	0.00	05/15/2025	13:45
19	Q2048-05	L2-WC-2	1	Solid	NO	0.00	05/15/2025	13:52
20	Q2048-08	L2-WC-2	1	Solid	NO	0.00	05/15/2025	14:00
21	Q2048-09	L2-WC-3	1	Solid	NO	0.00	05/15/2025	14:07
22	Q2048-12	L2-WC-3	1	Solid	NO	0.00	05/15/2025	14:15
23	Q2048-13	L2-WC-4	1	Solid	NO	0.00	05/15/2025	14:23
24	Q2048-16	L2-WC-4	1	Solid	NO	0.00	05/15/2025	14:30
25	Q2048-16DUP	L2-WC-4DUP	1	Solid	NO	0.00	05/15/2025	14:38

$$\text{Burning Rate} = \frac{\text{Length (mm)}}{\text{Total Time (sec)}}$$

WORKLIST(Hardcopy Internal Chain)

6135796

WorkList Name : ign-05-15

WorkList ID : 189530

Department : Wet-Chemistry

Date : 05-15-2025 08:18:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2032-09	COMP-1	Solid	Ignitability	Cool 4 deg C	CAMP02	L41	05/13/2025	1030
Q2034-01	L3-WC-1	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-04	L3-WC-1	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-05	L3-WC-2	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-08	L3-WC-2	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-09	L3-WC-3	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-12	L3-WC-3	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-13	L3-WC-4	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-16	L3-WC-4	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-17	L3-WC-5	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-20	L3-WC-5	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-21	L3-WC-6	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2034-24	L3-WC-6	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2038-01	72-11991	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/13/2025	1030
Q2038-02	72-11991	Solid	Ignitability	Cool 4 deg C	PSEG03	L31	05/14/2025	1030
Q2048-01	L2-WC-1	Solid	Ignitability	Cool 4 deg C	PSEG03	L31	05/14/2025	1030
Q2048-04	L2-WC-1	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/14/2025	1030
Q2048-05	L2-WC-2	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/14/2025	1030
Q2048-08	L2-WC-2	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/14/2025	1030
Q2048-09	L2-WC-3	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/14/2025	1030
Q2048-12	L2-WC-3	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/14/2025	1030

Date/Time 05/15/25 11:30
Raw Sample Received by: 12(50)
Raw Sample Relinquished by: 20(000)

Date/Time 05/17/25 15:00
Raw Sample Received by: Pcase
Raw Sample Relinquished by: 20(50)

WORKLIST(Hardcopy Internal Chain)

6135790

WorkList Name : ign-05-15

WorkList ID : 189530

Department : Wet-Chemistry

Date : 05-15-2025 08:18:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2048-13	L2-WC-4	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/14/2025	1030
Q2048-16	L2-WC-4	Solid	Ignitability	Cool 4 deg C	PSEG03	L41	05/14/2025	1030

Date/Time 05/15/25 11:30
Raw Sample Received by: 12 (20)
Raw Sample Relinquished by: [Signature]

Date/Time 05/15/25 15:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: 12 (20)

LB1358

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

5/16/2025 13:34

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	92.486	0.0	0.078	
ICB1	1.363	0.0	0.001	
CCV1	241.898	0.0	0.205	
CCB1	1.081	0.0	0.001	
PB168027BL	1.059	0.0	0.001	
Q2032-09	1.011	0.0	0.001	
Q2032-09DUP	1.068	0.0	0.001	
Q2034-04	1.120	0.0	0.001	
Q2034-08	1.124	0.0	0.001	
Q2034-12	1.103	0.0	0.001	
Q2034-16	1.059	0.0	0.001	
Q2034-20	1.114	0.0	0.001	
Q2034-24	1.014	0.0	0.001	
Q2038-02	0.838	0.0	0.001	
CCV2	241.608	0.0	0.204	
CCB2	1.020	0.0	0.001	
Q2048-04	0.834	0.0	0.001	
Q2048-08	0.809	0.0	0.001	
Q2048-12	0.975	0.0	0.001	
Q2048-16	1.075	0.0	0.001	
Q2052-04	1.105	0.0	0.001	
Q2057-04	1.028	0.0	0.001	
Q2062-04	1.241	0.0	0.001	
Q2062-08	1.181	0.0	0.001	
Q2062-12	1.243	0.0	0.001	
Q2062-16	1.179	0.0	0.001	
CCV3	240.005	0.0	0.203	
CCB3	1.063	0.0	0.001	
Q2062-20	1.166	0.0	0.001	
Q2062-24	1.192	0.0	0.001	
CCV4	243.644	0.0	0.206	
CCB4	1.017	0.0	0.001	
N	32			
Mean	34.023			
SD	81.3986			
CV%	239.25			

Aquakem v. 7.2AQ1

Results from time period:

Fri May 16 11:35:26 2025

Fri May 16 13:34:06 2025

Sample Id	Sam/Ctr/cA	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.4134	µg/l	5/16/2025 11:35:26	
5.0PPBCN	A	Total CN	P	5.6364	µg/l	5/16/2025 11:35:27	
10PPBCN	A	Total CN	P	10.8404	µg/l	5/16/2025 11:35:28	
50PPBCN	A	Total CN	P	49.2935	µg/l	5/16/2025 11:35:29	
100PPBCN	A	Total CN	P	97.9209	µg/l	5/16/2025 11:35:30	
250PPBCN	A	Total CN	P	250.8642	µg/l	5/16/2025 11:35:31	
500PPBCN	A	Total CN	P	500.0312	µg/l	5/16/2025 11:35:32	
ICV1	S	Total CN	P	92.4855	µg/l	5/16/2025 13:01:02	
ICB1	S	Total CN	P	1.3634	µg/l	5/16/2025 13:01:03	
CCV1	S	Total CN	P	241.8975	µg/l	5/16/2025 13:01:06	
CCB1	S	Total CN	P	1.0814	µg/l	5/16/2025 13:01:08	
PB168027BL	S	Total CN	P	1.0586	µg/l	5/16/2025 13:01:09	
Q2032-09	S	Total CN	P	1.0106	µg/l	5/16/2025 13:08:37	
Q2032-09DUP	S	Total CN	P	1.0678	µg/l	5/16/2025 13:08:39	
Q2034-04	S	Total CN	P	1.1202	µg/l	5/16/2025 13:08:41	
Q2034-08	S	Total CN	P	1.1237	µg/l	5/16/2025 13:08:42	
Q2034-12	S	Total CN	P	1.1031	µg/l	5/16/2025 13:08:43	
Q2034-16	S	Total CN	P	1.0593	µg/l	5/16/2025 13:08:44	
Q2034-20	S	Total CN	P	1.1138	µg/l	5/16/2025 13:08:45	
Q2034-24	S	Total CN	P	1.0139	µg/l	5/16/2025 13:08:46	
Q2038-02	S	Total CN	P	0.8383	µg/l	5/16/2025 13:16:11	
CCV2	S	Total CN	P	241.6078	µg/l	5/16/2025 13:16:16	
CCB2	S	Total CN	P	1.0202	µg/l	5/16/2025 13:16:18	
Q2048-04	S	Total CN	P	0.8341	µg/l	5/16/2025 13:16:19	
Q2048-08	S	Total CN	P	0.8091	µg/l	5/16/2025 13:16:20	
Q2048-12	S	Total CN	P	0.975	µg/l	5/16/2025 13:16:21	
Q2048-16	S	Total CN	P	1.075	µg/l	5/16/2025 13:23:43	
Q2052-04	S	Total CN	P	1.1047	µg/l	5/16/2025 13:23:44	
Q2057-04	S	Total CN	P	1.0282	µg/l	5/16/2025 13:23:45	
Q2062-04	S	Total CN	P	1.2407	µg/l	5/16/2025 13:23:46	
Q2062-08	S	Total CN	P	1.1812	µg/l	5/16/2025 13:23:47	
Q2062-12	S	Total CN	P	1.2427	µg/l	5/16/2025 13:23:48	
Q2062-16	S	Total CN	P	1.1787	µg/l	5/16/2025 13:23:49	
CCV3	S	Total CN	P	240.0051	µg/l	5/16/2025 13:23:53	
CCB3	S	Total CN	P	1.0628	µg/l	5/16/2025 13:31:19	
Q2062-20	S	Total CN	P	1.1659	µg/l	5/16/2025 13:31:21	
Q2062-24	S	Total CN	P	1.1918	µg/l	5/16/2025 13:31:22	
CCV4	S	Total CN	P	243.6436	µg/l	5/16/2025 13:31:27	
CCB4	S	Total CN	P	1.0172	µg/l	5/16/2025 13:34:06	

Calibration results

Aquakem 7.2AQ1

Page:

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

5/16/2025 11:54


Test Total CN

Accepted 5/16/2025 11:54

Factor 1182
Bias 0

Coeff. of det. 0.999967

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.000	0.4134	0.0000	-
2	5.0PPBCN	0.005	5.6364	5.0000	12.7
3	10PPBCN	0.009	10.8404	10.0000	8.4
4	50PPBCN	0.042	49.2935	50.0000	-1.4
5	100PPBCN	0.083	97.9209	100.0000	-2.1
6	250PPBCN	0.212	250.8642	250.0000	0.3
7	500PPBCN	0.423	500.0312	500.0000	0.0

05/16/2025
RM

SOP ID : M9030B-Sulfide-12

SDG No : N/A

Matrix : SOIL

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#1

Block ID : MC-1, MC-2

Weigh By : RM

Start Digest Date: 05/15/2025 Time : 08:50 Temp : N/A

End Digest Date: 05/15/2025 Time : 10:20 Temp : N/A

Digestion tube ID : M5595

Filter paper ID : N/A

pH Meter ID : N/A

Block Thermometer ID : N/A

Prep Technician Signature: RM

Supervisor Signature: 12

Standard Name	MLS USED	STD REF. # FROM LOG
PBS003	50.0ML	W3112
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.5M ZINC ACETATE	5.0ML	WP113086
FORMALDEHYDE	2.0ML	W2725
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

05/15/2025 RM

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/Nitrite	Comment	Prep Pos
PB168011BL	PBS011	5.00	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2017-04DUP	MH-IDUP	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2017-04	MH-I	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2017-08	MH-J	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2019-04	MH-K	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2020-04	TP-A	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2027-03	B27-SOIL-SAMPLE	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2027-04	B28-SOIL-SAMPLE	5.07	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2032-09	COMP-1	5.05	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-04	L3-WC-1	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-08	L3-WC-2	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-12	L3-WC-3	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-16	L3-WC-4	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-20	L3-WC-5	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-24	L3-WC-6	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2038-02	72-11991	5.05	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2048-04	L2-WC-1	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2048-08	L2-WC-2	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2048-12	L2-WC-3	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2048-16	L2-WC-4	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A

WORKLIST(Hardcopy Internal Chain)

WorkList Name : rsul-05-14 WorkList ID : 189537 Department : Distillation Date : 05-14-2025 14:58:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2017-04	MH-I	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/12/2025	9034
Q2017-08	MH-J	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/12/2025	9034
Q2019-04	MH-K	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/13/2025	9034
Q2020-04	TP-A	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/12/2025	9034
Q2027-03	B27-SOIL-SAMPLE	Solid	Reactive Sulfide	Cool 4 deg C	SCAL01	L41	05/12/2025	9034
Q2027-04	B28-SOIL-SAMPLE	Solid	Reactive Sulfide	Cool 4 deg C	SCAL01	L41	05/12/2025	9034
Q2032-09	COMP-1	Solid	Reactive Sulfide	Cool 4 deg C	CAMP02	L41	05/13/2025	9034
Q2034-04	L3-WC-1	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/13/2025	9034
Q2034-08	L3-WC-2	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/13/2025	9034
Q2034-12	L3-WC-3	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/13/2025	9034
Q2034-16	L3-WC-4	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/13/2025	9034
Q2034-20	L3-WC-5	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/13/2025	9034
Q2034-24	L3-WC-6	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/13/2025	9034
Q2038-02	72-11991	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L31	05/14/2025	9034
Q2048-04	L2-WC-1	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/14/2025	9034
Q2048-08	L2-WC-2	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/14/2025	9034
Q2048-12	L2-WC-3	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/14/2025	9034
Q2048-16	L2-WC-4	Solid	Reactive Sulfide	Cool 4 deg C	PSEG03	L41	05/14/2025	9034

Date/Time 05/15/2025 08:70
 Raw Sample Received by: RM (WCS)
 Raw Sample Relinquished by: JPC (WCS)

Date/Time 05/15/2025 11:00
 Raw Sample Received by: JPC (WCS)
 Raw Sample Relinquished by: RM (WCS)

SOP ID : M9012B-Total, Amenable and Reactive Cyanide-20

SDG No : N/A

Start Digest Date: 05/16/2025 Time : 08:45 Temp : N/A

Matrix : SOIL

End Digest Date: 05/16/2025 Time : 10:20 Temp : N/A

Pipette ID : N/A

1 batch 05/16/2025 10:45
05/16/2025 12:15

Balance ID : WC SC-7

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : N/A

Block ID : MC-1,MC-2

Filter paper ID : N/A

Prep Technician Signature: RIM

Weigh By : RM

pH Meter ID : N/A

Supervisor Signature: 12

Standard Name	MLS USED	STD REF. # FROM LOG
PBS003	50.0ML	W3112
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.25N NaOH	50.0ML	WP111294
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Comment

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
05/16/2025 12:30	RIM (WC)	RIM (WC)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB168027BL	PBS027	5.00	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2032-09DUP	COMP-1DUP	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2032-09	COMP-1	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-04	L3-WC-1	5.07	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-08	L3-WC-2	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-12	L3-WC-3	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-16	L3-WC-4	5.05	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-20	L3-WC-5	5.05	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2034-24	L3-WC-6	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2038-02	72-11991	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2048-04	L2-WC-1	5.05	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2048-08	L2-WC-2	5.07	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2048-12	L2-WC-3	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2048-16	L2-WC-4	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2052-04	TP-B	5.05	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2057-04	MH-L	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2062-04	L1-WC-1	5.07	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2062-08	L1-WC-2	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2062-12	L1-WC-3	5.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2062-16	L1-WC-4	5.06	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2062-20	L1-WC-5	5.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q2062-24	L1-WC-6	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A

Instrument ID: WC PH METER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135777

Review By	jignesh	Review On	5/15/2025 9:09:04 AM
Supervise By	Iwona	Supervise On	5/15/2025 12:05:07 PM
SubDirectory	LB135777	Test	Corrosivity
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3178,W3093,W3191,W3071,W3161,W3072		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	05/14/25 15:00		Jignesh	OK
2	CAL2	CAL2	CAL	05/14/25 15:01		Jignesh	OK
3	CAL3	CAL3	CAL	05/14/25 15:03		Jignesh	OK
4	ICV	ICV	ICV	05/14/25 15:05		Jignesh	OK
5	CCV1	CCV1	CCV	05/14/25 15:10		Jignesh	OK
6	Q2019-04	MH-K	SAM	05/14/25 15:20		Jignesh	OK
7	Q2019-04DUP	MH-KDUP	DUP	05/14/25 15:21		Jignesh	OK
8	Q2020-04	TP-A	SAM	05/14/25 15:25		Jignesh	OK
9	Q2027-03	B27-SOIL-SAMPLE	SAM	05/14/25 15:35		Jignesh	OK
10	Q2027-04	B28-SOIL-SAMPLE	SAM	05/14/25 15:44		Jignesh	OK
11	Q2032-09	COMP-1	SAM	05/14/25 15:50		Jignesh	OK
12	Q2034-04	L3-WC-1	SAM	05/14/25 16:00		Jignesh	OK
13	Q2034-08	L3-WC-2	SAM	05/14/25 16:05		Jignesh	OK
14	Q2034-12	L3-WC-3	SAM	05/14/25 16:11		Jignesh	OK
15	Q2034-16	L3-WC-4	SAM	05/14/25 16:15		Jignesh	OK
16	CCV2	CCV2	CCV	05/14/25 16:22		Jignesh	OK
17	Q2034-20	L3-WC-5	SAM	05/14/25 16:30		Jignesh	OK
18	Q2034-24	L3-WC-6	SAM	05/14/25 16:33		Jignesh	OK

Instrument ID: WC PH METER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135777

Review By	jignesh	Review On	5/15/2025 9:09:04 AM
Supervise By	Iwona	Supervise On	5/15/2025 12:05:07 PM
SubDirectory	LB135777	Test	Corrosivity
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3178,W3093,W3191,W3071,W3161,W3072		

19	Q2038-02	72-11991	SAM	05/14/25 16:40		Jignesh	OK
20	Q2048-04	L2-WC-1	SAM	05/14/25 16:44		Jignesh	OK
21	Q2048-08	L2-WC-2	SAM	05/14/25 16:50		Jignesh	OK
22	Q2048-12	L2-WC-3	SAM	05/14/25 17:00		Jignesh	OK
23	Q2048-16	L2-WC-4	SAM	05/14/25 17:10		Jignesh	OK
24	CCV3	CCV3	CCV	05/14/25 17:15		Jignesh	OK

Instrument ID: TITRAMETRIC

Daily Analysis Runlog For Sequence/QC Batch ID # LB135785

Review By	rubina	Review On	5/15/2025 1:11:12 PM
Supervise By	Iwona	Supervise On	5/15/2025 1:13:03 PM
SubDirectory	LB135785	Test	Reactive Sulfide
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3105,W3114,W3149		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	PB168011BL	PB168011BL	MB	05/15/25 11:03		rubina	OK
2	Q2017-04	MH-I	SAM	05/15/25 11:06		rubina	OK
3	Q2017-04DUP	MH-IDUP	DUP	05/15/25 11:09		rubina	OK
4	Q2017-08	MH-J	SAM	05/15/25 11:11		rubina	OK
5	Q2019-04	MH-K	SAM	05/15/25 11:14		rubina	OK
6	Q2020-04	TP-A	SAM	05/15/25 11:17		rubina	OK
7	Q2027-03	B27-SOIL-SAMPLE	SAM	05/15/25 11:20		rubina	OK
8	Q2027-04	B28-SOIL-SAMPLE	SAM	05/15/25 11:22		rubina	OK
9	Q2032-09	COMP-1	SAM	05/15/25 11:25		rubina	OK
10	Q2034-04	L3-WC-1	SAM	05/15/25 11:28		rubina	OK
11	Q2034-08	L3-WC-2	SAM	05/15/25 11:30		rubina	OK
12	Q2034-12	L3-WC-3	SAM	05/15/25 11:32		rubina	OK
13	Q2034-16	L3-WC-4	SAM	05/15/25 11:34		rubina	OK
14	Q2034-20	L3-WC-5	SAM	05/15/25 11:36		rubina	OK
15	Q2034-24	L3-WC-6	SAM	05/15/25 11:39		rubina	OK
16	Q2038-02	72-11991	SAM	05/15/25 11:42		rubina	OK
17	Q2048-04	L2-WC-1	SAM	05/15/25 11:45		rubina	OK
18	Q2048-08	L2-WC-2	SAM	05/15/25 11:48		rubina	OK

Instrument ID: TITRAMETRIC

Daily Analysis Runlog For Sequence/QC Batch ID # LB135785

Review By	rubina	Review On	5/15/2025 1:11:12 PM
Supervise By	Iwona	Supervise On	5/15/2025 1:13:03 PM
SubDirectory	LB135785	Test	Reactive Sulfide
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3105,W3114,W3149		

19	Q2048-12	L2-WC-3	SAM	05/15/25 11:50		rubina	OK
20	Q2048-16	L2-WC-4	SAM	05/15/25 11:53		rubina	OK

Instrument ID: FLAME

Daily Analysis Runlog For Sequence/QC Batch ID # LB135790

Review By	Iwona	Review On	5/15/2025 3:42:24 PM
Supervise By	jignesh	Supervise On	5/15/2025 3:43:56 PM
SubDirectory	LB135790	Test	Ignitability
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	Q2032-09	COMP-1	SAM	05/15/25 11:40		Iwona	OK
2	Q2032-09DUP	COMP-1DUP	DUP	05/15/25 11:48		Iwona	OK
3	Q2034-01	L3-WC-1	SAM	05/15/25 11:55		Iwona	OK
4	Q2034-04	L3-WC-1	SAM	05/15/25 12:03		Iwona	OK
5	Q2034-05	L3-WC-2	SAM	05/15/25 12:10		Iwona	OK
6	Q2034-08	L3-WC-2	SAM	05/15/25 12:17		Iwona	OK
7	Q2034-09	L3-WC-3	SAM	05/15/25 12:25		Iwona	OK
8	Q2034-12	L3-WC-3	SAM	05/15/25 12:32		Iwona	OK
9	Q2034-13	L3-WC-4	SAM	05/15/25 12:40		Iwona	OK
10	Q2034-16	L3-WC-4	SAM	05/15/25 12:47		Iwona	OK
11	Q2034-17	L3-WC-5	SAM	05/15/25 12:54		Iwona	OK
12	Q2034-20	L3-WC-5	SAM	05/15/25 13:00		Iwona	OK
13	Q2034-21	L3-WC-6	SAM	05/15/25 13:07		Iwona	OK
14	Q2034-24	L3-WC-6	SAM	05/15/25 13:15		Iwona	OK
15	Q2038-01	72-11991	SAM	05/15/25 13:22		Iwona	OK
16	Q2038-02	72-11991	SAM	05/15/25 13:30		Iwona	OK
17	Q2048-01	L2-WC-1	SAM	05/15/25 13:38		Iwona	OK
18	Q2048-04	L2-WC-1	SAM	05/15/25 13:45		Iwona	OK

Instrument ID: FLAME

Daily Analysis Runlog For Sequence/QC Batch ID # LB135790

Review By	Iwona	Review On	5/15/2025 3:42:24 PM
Supervise By	jignesh	Supervise On	5/15/2025 3:43:56 PM
SubDirectory	LB135790	Test	Ignitability
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

19	Q2048-05	L2-WC-2	SAM	05/15/25 13:52		Iwona	OK
20	Q2048-08	L2-WC-2	SAM	05/15/25 14:00		Iwona	OK
21	Q2048-09	L2-WC-3	SAM	05/15/25 14:07		Iwona	OK
22	Q2048-12	L2-WC-3	SAM	05/15/25 14:15		Iwona	OK
23	Q2048-13	L2-WC-4	SAM	05/15/25 14:23		Iwona	OK
24	Q2048-16	L2-WC-4	SAM	05/15/25 14:30		Iwona	OK
25	Q2048-16DUP	L2-WC-4DUP	DUP	05/15/25 14:38		Iwona	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB135800

Review By		Review On	
Supervise By		Supervise On	
SubDirectory	LB135800	Test	Reactive Cyanide
STD. NAME	STD REF.#		
ICAL Standard	WP113092,WP113093,WP113094,WP113095,WP113096,WP113097,WP113098		
ICV Standard	WP113099		
CCV Standard	WP113093		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP112643,WP112900,WP113100		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPBCN	0.0PPBCN	CAL1	05/16/25 11:35		rubina	OK
2	5.0PPBCN	5.0PPBCN	CAL2	05/16/25 11:35		rubina	OK
3	10PPBCN	10PPBCN	CAL3	05/16/25 11:35		rubina	OK
4	50PPBCN	50PPBCN	CAL4	05/16/25 11:35		rubina	OK
5	100PPBCN	100PPBCN	CAL5	05/16/25 11:35		rubina	OK
6	250PPBCN	250PPBCN	CAL6	05/16/25 11:35		rubina	OK
7	500PPBCN	500PPBCN	CAL7	05/16/25 11:35		rubina	OK
8	ICV1	ICV1	ICV	05/16/25 13:01		rubina	OK
9	ICB1	ICB1	ICB	05/16/25 13:01		rubina	OK
10	CCV1	CCV1	CCV	05/16/25 13:01		rubina	OK
11	CCB1	CCB1	CCB	05/16/25 13:01		rubina	OK
12	PB168027BL	PB168027BL	MB	05/16/25 13:01		rubina	OK
13	Q2032-09	COMP-1	SAM	05/16/25 13:08		rubina	OK
14	Q2032-09DUP	COMP-1DUP	DUP	05/16/25 13:08		rubina	OK
15	Q2034-04	L3-WC-1	SAM	05/16/25 13:08		rubina	OK
16	Q2034-08	L3-WC-2	SAM	05/16/25 13:08		rubina	OK
17	Q2034-12	L3-WC-3	SAM	05/16/25 13:08		rubina	OK
18	Q2034-16	L3-WC-4	SAM	05/16/25 13:08		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB135800

Review By	Review On
Supervise By	Supervise On
SubDirectory LB135800	Test Reactive Cyanide
STD. NAME	STD REF.#
ICAL Standard	WP113092,WP113093,WP113094,WP113095,WP113096,WP113097,WP113098
ICV Standard	WP113099
CCV Standard	WP113093
ICSA Standard	N/A
CRI Standard	N/A
LCS Standard	N/A
Chk Standard	WP112643,WP112900,WP113100

19	Q2034-20	L3-WC-5	SAM	05/16/25 13:08		rubina	OK
20	Q2034-24	L3-WC-6	SAM	05/16/25 13:08		rubina	OK
21	Q2038-02	72-11991	SAM	05/16/25 13:16		rubina	OK
22	CCV2	CCV2	CCV	05/16/25 13:16		rubina	OK
23	CCB2	CCB2	CCB	05/16/25 13:16		rubina	OK
24	Q2048-04	L2-WC-1	SAM	05/16/25 13:16		rubina	OK
25	Q2048-08	L2-WC-2	SAM	05/16/25 13:16		rubina	OK
26	Q2048-12	L2-WC-3	SAM	05/16/25 13:16		rubina	OK
27	Q2048-16	L2-WC-4	SAM	05/16/25 13:23		rubina	OK
28	Q2052-04	TP-B	SAM	05/16/25 13:23		rubina	OK
29	Q2057-04	MH-L	SAM	05/16/25 13:23		rubina	OK
30	Q2062-04	L1-WC-1	SAM	05/16/25 13:23		rubina	OK
31	Q2062-08	L1-WC-2	SAM	05/16/25 13:23		rubina	OK
32	Q2062-12	L1-WC-3	SAM	05/16/25 13:23		rubina	OK
33	Q2062-16	L1-WC-4	SAM	05/16/25 13:23		rubina	OK
34	CCV3	CCV3	CCV	05/16/25 13:23		rubina	OK
35	CCB3	CCB3	CCB	05/16/25 13:31		rubina	OK
36	Q2062-20	L1-WC-5	SAM	05/16/25 13:31		rubina	OK
37	Q2062-24	L1-WC-6	SAM	05/16/25 13:31		rubina	OK
38	CCV4	CCV4	CCV	05/16/25 13:31		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB135800

Review By		Review On	
Supervise By		Supervise On	
SubDirectory	LB135800	Test	Reactive Cyanide
STD. NAME	STD REF.#		
ICAL Standard	WP113092,WP113093,WP113094,WP113095,WP113096,WP113097,WP113098		
ICV Standard	WP113099		
CCV Standard	WP113093		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP112643,WP112900,WP113100		

39	CCB4	CCB4	CCB	05/16/25 13:34		rubina	OK
----	------	------	-----	----------------	--	--------	----

Prep Standard - Chemical Standard Summary

Order ID : Q2032

Test : Corrosivity,Ignitability,Percent Solids,Reactive Cyanide,Reactive Sulfide

Prepbatch ID : PB168011,PB168027,

Sequence ID/Qc Batch ID: LB135777,LB135785,LB135790,LB135800,

Standard ID :

WP111294,WP112643,WP112900,WP112995,WP113086,WP113091,WP113092,WP113093,WP113094,WP113095,WP113096,WP113097,WP113098,WP113099,WP113100,

Chemical ID :

M6151,W2668,W2725,W2926,W3019,W3071,W3072,W3093,W3105,W3112,W3113,W3114,W3139,W3149,W3154,W3161,W3173,W3178,W3191,W3203,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
11	Sodium hydroxide absorbing solution 0.25 N	WP111294	01/07/2025	07/07/2025	Niha Farheen Shaik	WETCHEM_SCALE_5 (WC-5)	None	Iwona Zarych 01/07/2025
<u>FROM</u>	21.00000L of W3112 + 210.00000gram of W3113 = Final Quantity: 21.000 L							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
539	CN BUFFER	WP112643	04/09/2025	10/09/2025	Niha Farheen Shaik	WETCHEM_SCALE_5 (WCS-5)	None	Iwona Zarych 04/09/2025
<u>FROM</u>	138.00000gram of W2668 + 862.00000ml of W3112 = Final Quantity: 1000.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
607	PYRIDINE-BARBITURIC ACID	WP112900	05/01/2025	08/18/2025	Rubina Mughal	WETCHEM_S CALE_8 (WC SC-7)	Glass Pipette-A	Iwona Zarych 05/01/2025
<u>FROM</u>	145.00000ml of W3112 + 15.00000gram of W3203 + 15.00000ml of M6151 + 75.00000ml of W3019 = Final Quantity: 250.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3371	Cyanide LCS Spike Solution, 5PPM	WP112995	05/07/2025	07/07/2025	Iwona Zarych	None	WETCHEM_PIPETTE_3 (WC)	Jignesh Parikh 05/07/2025
<u>FROM</u>	1.00000ml of W3173 + 199.00000ml of WP111294 = Final Quantity: 200.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
160	0.5M ZINC ACETATE	WP113086	05/15/2025	08/18/2025	Rubina Mughal	WETCHEM_SCALE_8 (WCS-7)	None	Iwona Zarych 05/15/2025
<u>FROM</u> 0.88900L of W3112 + 1.00000ml of M6151 + 110.00000gram of W2926 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3456	Cyanide Intermediate Working Std, 5PPM	WP113091	05/16/2025	05/17/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3 (WC)	Jignesh Parikh 05/16/2025
<u>FROM</u> 0.25000ml of W3154 + 49.75000ml of WP111294 = Final Quantity: 50.000 ml								

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4	Calibration standard 500 ppb	WP113092	05/16/2025	05/17/2025	Rubina Mughal	None	WETCHEM_F IPETTE_3 (WC)	Jignesh Parikh 05/16/2025
FROM 45.00000ml of WP111294 + 5.00000ml of WP113091 = Final Quantity: 50.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3761	Calibration-CCV CN Standard 250 ppb	WP113093	05/16/2025	05/17/2025	Rubina Mughal	None	WETCHEM_F IPETTE_3 (WC)	Jignesh Parikh 05/16/2025
FROM 2.50000ml of WP113091 + 47.50000ml of WP111294 = Final Quantity: 50.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
6	Calibration Standard 100 ppb	WP113094	05/16/2025	05/17/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3	Jignesh Parikh
(WC) FROM 1.00000ml of WP113091 + 49.00000ml of WP111294 = Final Quantity: 50.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
7	Calibration Standard 50 ppb	WP113095	05/16/2025	05/17/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3	Jignesh Parikh
(WC) FROM 0.50000ml of WP113091 + 49.50000ml of WP111294 = Final Quantity: 50.000 ml								

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
8	Calibration Standard 10 ppb	WP113096	05/16/2025	05/17/2025	Rubina Mughal	None	WETCHEM_F IPETTE_3 (WC)	Jignesh Parikh 05/16/2025

FROM 1.00000ml of WP113092 + 49.00000ml of WP111294 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
9	Calibration Standard 5 ppb	WP113097	05/16/2025	05/17/2025	Rubina Mughal	None	WETCHEM_F IPETTE_3 (WC)	Jignesh Parikh 05/16/2025

FROM 0.50000ml of WP113092 + 49.50000ml of WP111294 = Final Quantity: 50.000 ml

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
167	0 ppb CN calibration std	WP113098	05/16/2025	05/17/2025	Rubina Mughal	None	None	Jignesh Parikh
								05/16/2025

FROM 50.00000ml of WP111294 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2168	RCN ICV STD, 100 PPB	WP113099	05/16/2025	05/17/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3	Jignesh Parikh
							(WC)	05/16/2025

FROM 1.00000ml of WP112995 + 49.00000ml of WP111294 = Final Quantity: 50.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1582	Chloramine T solution, 0.014M	WP113100	05/16/2025	05/17/2025	Rubina Mughal	WETCHEM_SCALE_5 (WCS-5)	Glass Pipette-A	Jignesh Parikh
<u>FROM</u>	0.08000gram of W3139 + 20.00000ml of W3112 = Final Quantity: 20.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L)	22G2862015	08/18/2025	02/18/2025 / Sagar	01/15/2025 / Sagar	M6151

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3818-5 / SODIUM PHOSPHATE, MONOBAS/HYD, CRYST, ACS, 2.5 KG	0000225799	12/03/2025	04/05/2021 / Alexander	02/10/2020 / apatel	W2668

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	EMD-FX0410-5 / FORMALDEHYDE SOLUTION 450ML	60045	06/22/2025	08/19/2024 / lwona	06/22/2020 / apatel	W2725

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J4296-1 / ZINC ACETATE, DIHYD, CRYST, ACS, 500G	383058	07/05/2027	07/05/2022 / ketankumar	07/05/2022 / ketankumar	W2926

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
SIGMA ALDRICH	270970-1L / Pyridine 1L	SHBQ2113	04/03/2028	04/03/2023 / lwona	04/03/2023 / lwona	W3019

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14455-3 / buffer solution pH 7 yellow	4308H30	07/31/2025	01/02/2024 / JIGNESH	12/06/2023 / lwona	W3071

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14940-1 / Buffer Solution, PH12 (500ml)	2310P21	04/30/2025	01/02/2024 / JIGNESH	12/07/2023 / Iwona	W3072

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	566002 / BUFFER PH 7.00 GREEN 1PINT PK6	44001f99	12/31/2025	04/03/2024 / jignesh	04/02/2024 / jignesh	W3093

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL69870-8 / SODIUM THIOSULFATE,0.025N,4LIT RE	4403S13	09/30/2025	04/22/2024 / Iwona	04/22/2024 / Iwona	W3105

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-7 / Sodium Hydroxide Pellets 12 Kg	23B1556310	12/31/2025	07/08/2024 / Iwona	07/08/2024 / Iwona	W3113

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL35830-4 / IODINE SOLUTION .025N 1L	2405D89	05/31/2025	07/10/2024 / Iwona	07/10/2024 / Iwona	W3114

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	JTE494-6 / CHLORAMINE-T BAKER 250GM	10239484	09/09/2029	09/09/2024 / Iwona	09/09/2024 / Iwona	W3139

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL70850-8 / Starch Solution, 4L	4408P62	08/31/2026	10/16/2024 / Iwona	10/16/2024 / Iwona	W3149

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	RC2543-4 / CYANIDE STD 1000PPM 4OZ	1411J58	05/31/2025	12/02/2024 / Iwona	12/02/2024 / Iwona	W3154

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL13850-1 / Buffer Solution, PH2 (500ml)	2411E26	10/31/2026	12/09/2024 / Iwona	12/09/2024 / Iwona	W3161

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	LC135457 / Cyanide Standard, 1000 PPM, Second Source	45010168	07/17/2025	01/24/2025 / Iwona	01/24/2025 / Iwona	W3173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14055-3 / PH 4 BUFFER SOLUTION	2411A93	10/30/2026	04/01/2025 / JIGNESH	01/27/2025 / jignesh	W3178

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	1601-1 / PH 10.01 BUFFER,COLOR CD 475ML	2410F80	03/31/2026	04/01/2025 / JIGNESH	03/13/2025 / jignesh	W3191

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	EM-BX0035-3 / Barbituric Acid, 100 gms	WXBF3271V	05/16/2029	04/21/2025 / Iwona	04/21/2025 / Iwona	W3203

W3071
Rec 12/6/23

Certificate of Analysis 12

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4308H30

Product Number: 1551

Manufacture Date: AUG 09, 2023

Expiration Date: JUL 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Yellow liquid	Passed

*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.002	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (08/09/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

W3019
rec 4/3/23

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.comEmail USA: techserv@sial.comOutside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:

Pyridine - anhydrous, 99.8%

Product Number:

270970

Batch Number:

SHBQ2113

Brand:

SIAL

CAS Number:

110-86-1

MDL Number:

MFCD00011732

Formula:


C₅H₅N

Formula Weight:

79.10 g/mol

Quality Release Date:

15 DEC 2022



Test	Specification	Result
Appearance (Color)	Colorless	Colorless
Appearance (Form)	Liquid	Liquid
Infrared Spectrum	Conforms to Structure	Conforms
Purity (GC)	≥ 99.75 %	99.99 %
Water (by Karl Fischer)	≤ 0.003 %	0.002 %
Residue on Evaporation	≤ 0.0005 %	< 0.0001 %


Larry Coers, Director
Quality Control
Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





RICCA CHEMICAL COMPANY®

1841 Broad Street
Pocomoke City, MD 21851
<http://www.riccachemical.com>
1-888-GO-RICCA
customerservice@riccachemical.com

W 3072
REC. 12/01/23
12

Certificate of Analysis

Buffer, Reference Standard, pH 12.00 ± 0.01 at 25°C

Lot Number: 2310P21

Product Number: 1615

Manufacture Date: OCT 24, 2023

Expiration Date: APR 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

°C	15	20	25	30	35	40
pH	12.35	12.17	11.99	11.78	11.62	11.46

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Colorless liquid	Passed


*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	12.005	0.02	186-I-g, 186-II-g, 191d

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1615-1	4 L natural poly	18 months
1615-16	500 mL clear PET-G	18 months
1615-32	1 L natural poly	18 months
1615-5	20 L Cubitainer®	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Sharon Travers (10/24/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Certificate of Analysis

Date of Release: 2/26/2020

Name: Formaldehyde Solution
GR ACS
Meets ACS Specifications

Item No: FX0410 all size codes

Lot / Batch No: 60045

Country of Origin: USA

Characteristic	Requirement		Results	Units
	Min.	Max.		
Assay	36.5	38.0	36.71	%
Chloride (Cl)		5	<5	ppm
Color (APHA)		10	<10	
Form			Passes test	
Heavy metals (as Pb)		5	<5	ppm
Iron (Fe)		5	0.6	ppm
Residue after ignition		0.005	<0.0050	%
Sulfate (SO ₄)		0.002	<0.0020	%
Titrate acid		0.006	<0.0060	meq/g

Heather Sinn,

Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Millipore Corporation, an affiliate of Merck KGaA, Darmstadt, Germany
290 Concord Road
Billerica, MA 01821
U.S.A

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

 **avantor™**



M6151

R → 11/15/25

Material No.: 9530-33
Batch No.: 22G2862015
Manufactured Date: 2022-06-15
Retest Date: 2027-06-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.9 %
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.191
ACS – Bromide (Br)	≤ 0.005 %	< 0.005 %
ACS – Extractable Organic Substances	≤ 5 ppm	< 1 ppm
ACS – Free Chlorine (as Cl ₂)	≤ 0.5 ppm	< 0.5 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.03 ppm
Sulfate (SO ₄)	≤ 0.5 ppm	< 0.3 ppm
Sulfite (SO ₃)	≤ 0.8 ppm	0.3 ppm
Ammonium (NH ₄)	≤ 3 ppm	< 1 ppm
Trace Impurities – Arsenic (As)	≤ 0.010 ppm	< 0.003 ppm
Trace Impurities – Aluminum (Al)	≤ 10.0 ppb	1.3 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 3.0 ppb
Trace Impurities – Barium (Ba)	≤ 1.0 ppb	0.2 ppb
Trace Impurities – Beryllium (Be)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Bismuth (Bi)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Boron (B)	≤ 20.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Calcium (Ca)	≤ 50.0 ppb	163.0 ppb
Trace Impurities – Chromium (Cr)	≤ 1.0 ppb	0.7 ppb
Trace Impurities – Cobalt (Co)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gallium (Ga)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Germanium (Ge)	≤ 3.0 ppb	< 2.0 ppb
Trace Impurities – Gold (Au)	≤ 4.0 ppb	0.6 ppb
Heavy Metals (as Pb)	≤ 100 ppb	< 50 ppb
Trace Impurities – Iron (Fe)	≤ 15 ppb	6 ppb

>>> Continued on page 2 >>>

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

 **avantorsm**



Material No.: 9530-33
Batch No.: 22G2862015

Test	Specification	Result
Trace Impurities – Lead (Pb)	≤ 1.0 ppb	< 0.5 ppb
Trace Impurities – Lithium (Li)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Magnesium (Mg)	≤ 10.0 ppb	2.9 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	0.1 ppb
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 3.0 ppb
Trace Impurities – Nickel (Ni)	≤ 4.0 ppb	< 0.3 ppb
Trace Impurities – Niobium (Nb)	≤ 1.0 ppb	0.8 ppb
Trace Impurities – Potassium (K)	≤ 9.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se), For Information Only		< 1.0 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	< 10.0 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	0.5 ppb
Trace Impurities – Sodium (Na)	≤ 100.0 ppb	2.3 ppb
Trace Impurities – Strontium (Sr)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Tantalum (Ta)	≤ 1.0 ppb	1.6 ppb
Trace Impurities – Thallium (Tl)	≤ 5.0 ppb	< 2.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	4.0 ppb
Trace Impurities – Titanium (Ti)	≤ 1.0 ppb	1.5 ppb
Trace Impurities – Vanadium (V)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.8 ppb
Trace Impurities – Zirconium (Zr)	≤ 1.0 ppb	0.3 ppb

>>> Continued on page 3 >>>

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA–ANALYZED® Reagent
For Trace Metal Analysis



Material No.: 9530-33
Batch No.: 22G2862015

Test	Specification	Result
------	---------------	--------

For Laboratory, Research, or Manufacturing Use
Product Information (not specifications):
Appearance (clear, fuming liquid)
Meets ACS Specifications
Storage Condition: Store below 25 °C.

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature in cursive script that reads 'Jamie Ethier'.
Jamie Ethier
Vice President Global Quality

Sodium Phosphate, Monobasic, Monohydrate,
Crystal
BAKER ANALYZED® A.C.S. Reagent

(sodium dihydrogen phosphate, monohydrate)



Material No.: 3818-05
Batch No.: 0000225799
Manufactured Date: 2018/12/05
Retest Date: 2025/12/03
Revision No: 1

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay ($\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$)	98.0 – 102.0 %	99.5
pH of 5% Solution at 25°C	4.1 – 4.5	4.3
Insoluble Matter	≤ 0.01 %	< 0.01
Chloride (Cl)	≤ 5 ppm	< 5
ACS – Sulfate (SO_4)	≤ 0.003 %	< 0.003
Calcium (Ca)	≤ 0.005 %	< 0.005
Potassium (K)	≤ 0.01 %	< 0.01
Heavy Metals (as Pb)	≤ 0.001 %	< 0.001
Trace Impurities – Iron (Fe)	≤ 0.001 %	< 0.001

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: IN
Packaging Site: Paris Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality


For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Certificate of Analysis

Product Name:

Zinc acetate dihydrate - ACS reagent, $\geq 98\%$

Product Number: 383058
Batch Number: MKCQ9159
Brand: SIGALD
CAS Number: 5970-45-6
MDL Number: MFCD00066961
Formula: $C_4H_6O_4Zn \cdot 2H_2O$
Formula Weight: 219.51 g/mol
Quality Release Date: 06 JAN 2022



W2926
Open 7/5/22
received
on 7/5/22

Test	Specification	Result
Appearance (Color)	White	White
Appearance (Form)	Powder or Crystal or Chunk(s)	Powder
Infrared Spectrum	Conforms to Structure	Conforms
Insoluble Matter	$\leq 0.005 \%$	0.003 %
Calcium (Ca)	$\leq 0.005 \%$	0.003 %
Chloride (Cl)	$\leq 5 \text{ ppm}$	$< 5 \text{ ppm}$
Iron (Fe)	$\leq 5 \text{ ppm}$	$< 5 \text{ ppm}$
Potassium (K)	$\leq 0.01 \%$	0.00 %
Magnesium (Mg)	$\leq 0.005 \%$	0.003 %
Sodium (Na)	$\leq 0.05 \%$	0.03 %
Lead (Pb)	$\leq 0.002 \%$	$< 0.001 \%$
pH	6.0 - 7.0	6.1
Sulfate (SO ₄)	$\leq 0.005 \%$	$< 0.005 \%$
Complexometric EDTA	98.0 - 101.0 %	100.3 %
Meets ACS Requirements	Meets Requirements	Meets Requirements


Larry Coers, Director
Quality Control
Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



**RICCA CHEMICAL COMPANY®**

1490 Lammers Pike

Batesville, IN 47006

<http://www.riccachemical.com>

1-888-GO-RICCA

customerservice@riccachemical.com

Certificate of Analysis

W3093
004121
04/03/2024
16

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)**Lot Number:** 4401F99**Product Number:** 1551**Manufacture Date:** JAN 08, 2024**Expiration Date:** DEC 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	

Test	Specification	Result
Appearance	Yellow liquid	Passed

*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.004	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-1	4 L natural poly	24 months
1551-1CT	4 L Cubitainer®	24 months
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (01/08/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Certificate of Analysis

Sodium Thiosulfate, 0.0250 Normal (N/40)

Lot Number: 4403S13

Product Number: 7900

Manufacture Date: MAR 29, 2024

Expiration Date: SEP 2025

This product is specially formulated to increase its stability. A preservative is added to prevent bacterial contamination. However, all Sodium Thiosulfate solutions are subject to slow chemical deterioration and should be restandardized periodically.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Thiosulfate Pentahydrate	10102-17-7	ACS
Organic Preservative	Proprietary	
Sodium Carbonate	497-19-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Iodate/Starch)	0.02499-0.02501 N at 20°C	0.02501 N at 20°C	136

Specification	Reference
Standard Sodium Thiosulfate Solution, 0.0250 N	APHA (4500-S2- F)
Standard Sodium Thiosulfate Titrant	APHA (4500-O D)
Standard Sodium Thiosulfate Titrant	APHA (4500-O E)
Standard Sodium Thiosulfate Titrant	APHA (4500-O F)
Standard Sodium Thiosulfate Titrant, 0.025 N	APHA (4500-CI B)
Standard Sodium Thiosulfate Titrant	APHA (4500-O C)
Standard Sodium Thiosulfate Titrant, 0.025 M	APHA (5530 C)
Standard Sodium Thiosulfate Solution (0.025 N)	EPA (SW-846) (9031)
Standard Sodium Thiosulfate solution (0.025 N)	EPA (SW-846) (9034)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7900-1	4 L natural poly	18 months
7900-16	500 mL natural poly	18 months
7900-1CT	4 L Cubitainer®	18 months
7900-32	1 L natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (03/29/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Certificate of Analysis



Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	<= 0.005 %	<0.005 %	PASS
Chloride	<= 0.005 %	0.002 %	PASS
Heavy Metals	<= 0.002 %	<0.002 %	PASS
Iron	<= 0.001 %	<0.001 %	PASS
Magnesium	<= 0.002 %	<0.002 %	PASS
Mercury	<= 0.1 ppm	<0.1 ppm	PASS
Nickel	<= 0.001 %	<0.001 %	PASS
Nitrogen Compounds	<= 0.001 %	<0.001 %	PASS
Phosphate	<= 0.001 %	<0.001 %	PASS
Potassium	<= 0.02 %	<0.02 %	PASS
Purity	>= 97.0 %	99.2 %	PASS
Sodium Carbonate	<= 1.0 %	0.5 %	PASS
Sulfate	<= 0.003 %	<0.003 %	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

Spec Set: 0583ACS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

Certificate of Analysis

Iodine (Iodine-Iodide), 0.0250 Normal (N/40), 1 mL = 0.4008 mg S²⁻

Lot Number: 2405D89

Product Number: 3975

Manufacture Date: MAY 10, 2024

Expiration Date: MAY 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Iodide	7681-11-0	ACS
Iodine	7553-56-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Dark brown liquid	Passed	
Assay (vs. Sodium Thiosulfate/Starch)	0.02498-0.02502 N at 20°C	0.02502 N at 20°C	136

Specification	Reference
Standard Iodine Solution, 0.0250 N	APHA (4500-S2- F)
Iodine Solution (approximately 0.025 N)	EPA (SW-846) (9031)
Standard Iodine Solution, 0.0250 N	EPA (376.1)
Iodine Solution (approximately 0.025 N)	EPA (SW-846) (9034)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3975-1	4 L amber glass	12 months
3975-16	500 mL amber glass	12 months
3975-32	1 L amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Jose Pena (05/10/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

W3139 Received on 9/9/24 by IZ

Product No.: A12044
Product: Chloramine-T trihydrate, 98%
Lot No.: 10239484

Appearance:	White powder
Melting Point:	166°C(dec)
Assay (Iodometric titration):	100.5%
Identification (FTIR):	Conforms

Order our products online thermofisher.com/chemicals

This document has been electronically generated and does not require a signature.

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.



Certificate of Analysis

Starch Indicator, 0.5% (w/v), Mercury Free, for Iodometric Titrations

Lot Number: 4408P62

Product Number: 8000

Manufacture Date: AUG 28, 2024

Expiration Date: AUG 2026

This product is Mercury-free.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Starch, soluble	9005-84-9	ACS
Salicylic Acid	69-72-7	ACS

Test	Specification	Result
Appearance	White translucent liquid	Passed
Suitability for Use	Colorless (Iodine absent) - Blue (Iodine present)	Passed

Specification	Reference
Starch Solution	APHA (4500-S2- F)
Starch Indicator Solution	APHA (4500-CI B)
Starch Indicator	APHA (4500-SO32- B)
Starch indicator solution	APHA (2350 B)
Starch indicator solution	APHA (2350 E)
Starch Solution	APHA (510 B)
Starch Solution	APHA (5530 C)
Starch Indicator	APHA (4500-CI C)
Starch Indicator	EPA (345.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8000-1	4 L natural poly	24 months
8000-16	500 mL natural poly	24 months
8000-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

A handwritten signature in blue ink that reads "Paul Brandon". The signature is fluid and cursive, with the first name "Paul" and last name "Brandon" clearly distinguishable.

Paul Brandon (08/28/2024)
Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Certificate of Analysis

Cyanide Standard, 1000 ppm CN⁻

Lot Number: 1411J58**Product Number:** 2543**Manufacture Date:** NOV 22, 2024**Expiration Date:** MAY 2025

This standard is prepared using accurate volumetric techniques from material that has been assayed against Silver Nitrate solution certified traceable to NIST Standard Reference Material 999. The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is the combined uncertainty based on the stability of the assayed Potassium Cyanide, and the uncertainty in the mass and volume measurements.

Use 0.16% (w/v) (0.04 N) Sodium Hydroxide or 0.225 % (w/v) (0.04 N) Potassium Hydroxide to make dilutions of this standard. Restandardize weekly if extreme accuracy is required.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Cyanide	151-50-8	ACS
Sodium Hydroxide	1310-73-2	Reagent


Test	Specification	Result
Appearance	Colorless liquid	Passed
Cyanide (CN ⁻)	995-1005 ppm	1000 ppm

Specification	Reference
Stock Standard Cyanide Solution	APHA (4500-CN- F)
Stock Cyanide Solution	APHA (4500-CN- E)
Stock Cyanide Solution	APHA (4500-CN- K)
Stock Cyanide Solution	APHA (4500-CN- H)
Cyanide Reference Solution (1000 mg/L)	EPA (SW-846) (7.3.3.2)
Cyanide Calibration Stock Solution (1,000 mg/L CN ⁻)	EPA (SW-846) (9213)
Stock Cyanide Solution	EPA (335.3)
Stock Cyanide Solution	EPA (335.2)
Cyanide Solution Stock	ASTM (D 4282)
Simple Cyanide Solution, Stock (1.0 g/L CN ⁻)	ASTM (D 4374)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2543-16	500 mL amber poly	6 months
2543-32	1 L amber poly	6 months
2543-4	120 mL amber poly	6 months

Recommended Storage: 2°C - 8°C (36°F - 46°F)

A handwritten signature in black ink, appearing to read 'Luis Briceno', is written over a horizontal line.

Luis Briceno (11/22/2024)
Operations Supervisor

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Certificate of Analysis

Buffer, Reference Standard, pH 2.00 ± 0.01 at 25°C**Lot Number:** 2411E26**Product Number:** 1493**Manufacture Date:** NOV 11, 2024**Expiration Date:** OCT 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	10	15	20	25	30	35	40	45	50
pH	1.93	1.98	1.98	2.00	2.01	2.03	2.03	2.04	2.04

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result
Appearance	Colorless liquid	Passed

*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	1.994	0.02	185i, 186-I-g, 186-II-g

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1493-1	4 L natural poly	24 months
1493-16	500 mL natural poly	24 months
1493-1CT	4 L Cubitainer®	24 months
1493-2.5	10 L Cubitainer®	24 months
1493-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Jose Pena (11/11/2024)
Operations Manager

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



Part of TCP Analytical Group

Jackson's Pointe Commerce Park- Building 1000
1010 Jackson's Pointe Court, Zelienople, PA 16063

Certificate of Analysis

Cyanide Standard 1000 ppm (1ml = 1mg CN)

Product Code: **LC13545**

Manufacture Date: January 16, 2025

Lot Number: **45010168**

Expiration Date: July 17, 2025

Test	Specification	Result
Appearance (clarity)	clear solution	clear solution
Appearance (color)	colorless	colorless
Concentration (CN)	0.990 - 1.010mg/mL	1.000mg/mL
Concentration (CN)	990 - 1,010ppm	1,000ppm
Traceable to NIST SRM	Report	999b

Intended Use - Product is intended for use in manufacturing procedures and laboratory procedures and protocols.

Storage Information - Unless noted on the product label, store the product under normal lab conditions in its tightly closed, original container. Do not pipet directly from the container or return unused portions to the container.

Instructions for Handling and Use - Please refer to the associated product label and Safety Data Sheet (SDS) for information regarding safety and handling of this product.

Preparation - All products are manufactured and tested according to established, documented procedures and methodology. Production documentation records manufacturing data, raw material traceability and testing history on a per lot basis. Balances, thermometers, and glassware are calibrated before first use and on a regular schedule with references traceable to NIST

The suffix of the product code may differ from what is on your product label. The suffix will designate the size and be associated with a numeric digit(s). Visit LabChem.com for more information

Suffix	1	2	3/3S/36/36S	4/4C	5	6	7	8	9	20	44	200	246	486
Size	500mL org	1L or 1kg	2.5L/2.5L Coated/6x2.5L/6x2.5L Coated	4L	20L	10L	125mL	25g	100g	20x20mL	4x4L	200L	24x6mL	48x6mL

Michael Monteleone

Michael Monteleone
Chemistry Supervisor - Quality Control
2025011610:36:11bsturges-0-0

ISO9001:2015 Registration #0306-01

**RICCA CHEMICAL COMPANY®**

1841 Broad Street
Pocomoke City, MD 21851
<http://www.riccachemical.com>
1-888-GO-RICCA
customerservice@riccachemical.com

Certificate of Analysis

W21758 58

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number: 2411A93

Product Number: 1501

Manufacture Date: NOV 04, 2024

Expiration Date: OCT 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST Traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	4.00	4.00	4.00	4.00	4.00	4.00	4.01	4.02	4.03	4.04	4.06

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Acid Phthalate	877-24-7	Buffer
Preservative	Proprietary	Commercial
Red Dye	Proprietary	Purified

Test	Specification	Result
Appearance	Red liquid	Passed

*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	4.008	0.02	185i, 186-I-g, 186-II-g

Specification	Reference
Commercial Buffer Solutions	
Buffer B	ASTM (D 1293 B)
Buffer B	ASTM (D 5464)
Buffer B	ASTM (D 5128)

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1501-16	500 mL natural poly	24 months
1501-2.5	10 L Cubitainer®	24 months
1501-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



RICCA CHEMICAL COMPANY®

1841 Broad Street
Pocomoke City, MD 21851
<http://www.riccachemical.com>
1-888-GO-RICCA
customerservice@riccachemical.com

Certificate of Analysis

Buffer, Reference Standard, pH 10.00 ± 0.01 at 25°C (Color Coded Blue)

Lot Number: 2410F80

Product Number: 1601

Manufacture Date: OCT 09, 2024

Expiration Date: MAR 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.
The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	50
pH	10.31	10.23	10.17	10.11	10.05	10.00	9.95	9.91	9.87	9.81

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Carbonate	497-19-8	ACS
Sodium Bicarbonate	144-55-8	ACS
Sodium Hydroxide	1310-73-2	Reagent
Preservative	Proprietary	
Blue Dye	Proprietary	

Test	Specification	Result
Appearance	Blue liquid	Passed

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	10.009	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	
Buffer C	ASTM (D 1293 B)
Buffer C	ASTM (D 5464)
	ASTM (D 5128)

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1601-1	4 L natural poly	18 months
1601-16	500 mL natural poly	18 months
1601-1CT	4 L Cubitainer®	18 months
1601-2.5	10 L Cubitainer®	18 months
1601-32	1 L natural poly	18 months
1601-5	20 L Cubitainer®	18 months

Version: 1.3

Lot Number: 2410F80

Product Number: 1601


Page 1 of 2

Certificate of Analysis

Product Name:

Barbituric acid - ReagentPlus® , 99%

Product Number: 185698
Batch Number: WXBFB3271V
Brand: SIAL
CAS Number: 67-52-7
Formula: C₄H₄N₂O₃
Formula Weight: 128,09 g/mol
Quality Release Date: 16 MAY 2024



Test	Specification	Result
Appearance (Colour)	White to Off-White	White
Appearance (Form)	Powder	Powder
Infrared spectrum	Conforms to Structure	Conforms
Purity (Titration by NaOH)	98.5 - 101.5 %	100.4 %
GC (area %)	≥ 98 %	100 %
VPCT		



Kang Chen
Quality Manager
Wuxi , China CN

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/15/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 05/14/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:28
Out Date: 05/15/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135767

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2021-01	WASTE	1	1.14	10.48	11.62	10.41	88.5	
Q2021-02	VOC	2	1.19	10.25	11.44	9.99	85.9	
Q2021-03	1	3	1.15	10.01	11.16	9.88	87.2	
Q2021-04	2	4	1.18	10.25	11.43	9.88	84.9	
Q2021-05	3	5	1.17	10.09	11.26	10.12	88.7	
Q2021-06	4	6	1.16	10.60	11.76	10.5	88.1	
Q2021-07	5	7	1.18	10.60	11.78	10.36	86.6	
Q2021-08	6	8	1.18	10.21	11.39	10.12	87.6	
Q2032-01	TP-11	9	1.14	10.08	11.22	8.83	76.3	
Q2032-02	TP-29	10	1.18	10.39	11.57	10.13	86.1	
Q2032-03	TP-29-99	11	1.19	10.40	11.59	10.22	86.8	
Q2032-04	TP-24	12	1.15	10.70	11.85	9.09	74.2	
Q2032-05	Q2032-04MS	13	1.15	10.70	11.85	9.09	74.2	
Q2032-06	Q2032-04MSD	14	1.15	10.70	11.85	9.09	74.2	
Q2032-07	TP-37	15	1.19	10.00	11.19	9.44	82.5	
Q2032-08	TP-32	16	1.17	10.08	11.25	9.92	86.8	
Q2034-01	L3-WC-1	17	1.14	10.24	11.38	10.11	87.6	
Q2034-02	L3-WC-1-EPH	18	1.18	10.23	11.41	10.22	88.4	
Q2034-03	L3-WC-1-VOC	19	1.15	10.10	11.25	10.19	89.5	
Q2034-04	L3-WC-1	20	1.14	10.24	11.38	10.11	87.6	
Q2034-05	L3-WC-2	21	1.17	10.18	11.35	9.88	85.6	
Q2034-06	L3-WC-2-EPH	22	1.18	10.71	11.89	10.47	86.7	
Q2034-07	L3-WC-2-VOC	23	1.19	10.52	11.71	10.46	88.1	
Q2034-09	L3-WC-3	24	1.14	10.26	11.4	9.65	82.9	
Q2034-10	L3-WC-3-EPH	25	1.14	10.26	11.4	10.11	87.4	
Q2034-11	L3-WC-3-VOC	26	1.19	10.53	11.72	10.39	87.4	
Q2034-13	L3-WC-4	27	1.18	10.51	11.69	10.26	86.4	
Q2034-14	L3-WC-4-EPH	28	1.15	10.42	11.57	10.23	87.1	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/15/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 05/14/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:28
Out Date: 05/15/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135767

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2034-15	L3-WC-4-VOC	29	1.15	10.37	11.52	10.7	92.1	
Q2034-17	L3-WC-5	30	1.13	10.76	11.89	10.49	87.0	
Q2034-18	L3-WC-5-EPH	31	1.15	10.66	11.81	10.02	83.2	
Q2034-19	L3-WC-5-VOC	32	1.19	10.72	11.91	10.6	87.8	
Q2034-21	L3-WC-6	33	1.18	10.14	11.32	9.92	86.2	
Q2034-22	L3-WC-6-EPH	34	1.14	10.08	11.22	10.6	93.8	
Q2034-23	L3-WC-6-VOC	35	1.11	10.00	11.11	9.3	81.9	
Q2038-01	72-11991	36	1.15	10.56	11.71	10.81	91.5	
Q2039-01	TR-05-05142025	37	1.18	9.92	11.1	9.54	84.3	
Q2039-02	TR-05-05142025-E2	38	1.15	10.23	11.38	10.17	88.2	
Q2040-01	1	39	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2040-02	2	40	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2040-03	3	41	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2040-04	4	42	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2040-05	5	43	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2040-06	6	44	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2040-07	7	45	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2041-01	HIH359A-1-1	46	1.00	1.00	2.00	2.00	100.0	oilc
Q2041-02	HIH359A-1-2	47	1.00	1.00	2.00	2.00	100.0	oilc
Q2041-03	HIA9895-1-1	48	1.00	1.00	2.00	2.00	100.0	oilc
Q2041-04	HIA9895-1-2	49	1.00	1.00	2.00	2.00	100.0	oilc
Q2041-05	250-1-1	50	1.00	1.00	2.00	2.00	100.0	oilc
Q2041-06	250-1-2	51	1.00	1.00	2.00	2.00	100.0	oilc
Q2041-07	250-2-1	52	1.00	1.00	2.00	2.00	100.0	oilc
Q2041-08	250-2-2	53	1.00	1.00	2.00	2.00	100.0	oilc
Q2042-01	HEI570L-1-1	54	1.00	1.00	2.00	2.00	100.0	oilc
Q2042-02	HEI570L-1-2	55	1.00	1.00	2.00	2.00	100.0	oilc
Q2042-04	HEI570L-2-1	56	1.00	1.00	2.00	2.00	100.0	oilc



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/15/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 05/14/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:28
Out Date: 05/15/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135767

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2042-05	HEI570L-2-2	57	1.00	1.00	2.00	2.00	100.0	oilc
Q2043-01	281	58	1.18	10.26	11.44	9.29	79.0	
Q2043-02	281-E2	59	1.16	10.44	11.6	10.6	90.4	
Q2043-03	72-12000	60	1.19	10.35	11.54	8.15	67.2	
Q2043-04	72-12000-E2	61	1.11	10.38	11.49	6.56	52.5	
Q2044-01	3858	62	1.00	1.00	2.00	2.00	100.0	oil sample
Q2046-01	BC136902-1-1	63	1.00	1.00	2.00	2.00	100.0	oilc
Q2046-02	BC136902-1-2	64	1.00	1.00	2.00	2.00	100.0	oilc
Q2047-01	SU-04-051425	65	1.18	10.07	11.25	10.72	94.7	
Q2047-02	SU-04-051425-E2	66	1.18	10.37	11.55	10.9	93.7	
Q2048-01	L2-WC-1	67	1.19	10.47	11.66	10.29	86.9	
Q2048-02	L2-WC-1-EPH	68	1.19	10.30	11.49	9.87	84.3	
Q2048-03	L2-WC-1-VOC	69	1.18	10.31	11.49	10.17	87.2	
Q2048-05	L2-WC-2	70	1.16	10.26	11.42	9.82	84.4	
Q2048-06	L2-WC-2-EPH	71	1.17	10.13	11.3	9.06	77.9	
Q2048-07	L2-WC-2-VOC	72	1.18	10.42	11.6	9.88	83.5	
Q2048-09	L2-WC-3	73	1.17	10.15	11.32	9.02	77.3	
Q2048-10	L2-WC-3-EPH	74	1.18	10.07	11.25	9.17	79.3	
Q2048-11	L2-WC-3-VOC	75	1.17	10.49	11.66	9.94	83.6	
Q2048-13	L2-WC-4	76	1.11	10.64	11.75	10.24	85.8	
Q2048-14	L2-WC-4-EPH	77	1.19	10.13	11.32	10.04	87.4	
Q2048-15	L2-WC-4-VOC	78	1.13	10.07	11.2	9.81	86.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

VB 135767

WorkList Name : %1-051425

WorkList ID : 189504

Department : Wet-Chemistry

Date : 05-14-2025 09:06:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2021-01	WASTE	Solid	Percent Solids	Cool 4 deg C	SCIA01	L41	05/12/2025	Chemtech -SO
Q2021-02	VOC	Solid	Percent Solids	Cool 4 deg C	SCIA01	L41	05/12/2025	Chemtech -SO
Q2021-03	1	Solid	Percent Solids	Cool 4 deg C	SCIA01	L41	05/12/2025	Chemtech -SO
Q2021-04	2	Solid	Percent Solids	Cool 4 deg C	SCIA01	L41	05/12/2025	Chemtech -SO
Q2021-05	3	Solid	Percent Solids	Cool 4 deg C	SCIA01	L41	05/12/2025	Chemtech -SO
Q2021-06	4	Solid	Percent Solids	Cool 4 deg C	SCIA01	L41	05/12/2025	Chemtech -SO
Q2021-07	5	Solid	Percent Solids	Cool 4 deg C	SCIA01	L41	05/12/2025	Chemtech -SO
Q2021-08	6	Solid	Percent Solids	Cool 4 deg C	SCIA01	L41	05/12/2025	Chemtech -SO
Q2032-01	TP-11	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/13/2025	Chemtech -SO
Q2032-02	TP-29	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/13/2025	Chemtech -SO
Q2032-03	TP-29-99	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/13/2025	Chemtech -SO
Q2032-04	TP-24	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/13/2025	Chemtech -SO
Q2032-05	Q2032-04MS	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/13/2025	Chemtech -SO
Q2032-06	Q2032-04MSD	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/13/2025	Chemtech -SO
Q2032-07	TP-37	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/13/2025	Chemtech -SO
Q2032-08	TP-32	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/13/2025	Chemtech -SO
Q2034-01	L3-WC-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-02	L3-WC-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-03	L3-WC-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-04	L3-WC-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-05	L3-WC-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO

Date/Time 05-14-25 15:10

Raw Sample Received by: JB WOC

Raw Sample Relinquished by: [Signature]

Date/Time 05-14-25

Raw Sample Received by: CP SM

Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

B 135767

WorkList Name : %1-051425 WorkList ID : 189504 Department : Wet-Chemistry Date : 05-14-2025 09:06:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2034-06	L3-WC-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-07	L3-WC-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-09	L3-WC-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-10	L3-WC-3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-11	L3-WC-3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-13	L3-WC-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-14	L3-WC-4-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-15	L3-WC-4-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-17	L3-WC-5	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-18	L3-WC-5-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-19	L3-WC-5-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-21	L3-WC-6	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-22	L3-WC-6-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2034-23	L3-WC-6-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/13/2025	Chemtech -SO
Q2038-01	72-11991	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/14/2025	Chemtech -SO
Q2039-01	TR-05-05142025	Solid	Percent Solids	Cool 4 deg C	PSEG05	L21	05/14/2025	Chemtech -SO
Q2039-02	TR-05-05142025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L21	05/14/2025	Chemtech -SO
Q2040-01	1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2040-02	2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2040-03	3	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2040-04	4	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO

Date/Time 05-14-25 15:10 Date/Time 05-14-25 Date/Time 17:30
 Raw Sample Received by: SP WOC Raw Sample Received by: CP SR Raw Sample Received by: CP SR
 Raw Sample Relinquished by: CP SR Raw Sample Relinquished by: SP WOC

WORKLIST(Hardcopy Internal Chain)

23567

WorkList Name : %1-051425 WorkList ID : 189504 Department : Wet-Chemistry Date : 05-14-2025 09:06:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2040-05	5	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2040-06	6	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2040-07	7	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2041-01	HIH359A-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2041-02	HIH359A-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2041-03	HIA9895-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2041-04	HIA9895-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2041-05	250-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2041-06	250-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2041-07	250-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2041-08	250-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2042-01	HEI570L-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2042-02	HEI570L-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2042-04	HEI570L-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2042-05	HEI570L-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2043-01	281	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2043-02	281-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2043-03	72-12000	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2043-04	72-12000-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	05/14/2025	Chemtech -SO
Q2044-01	3858	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/14/2025	Chemtech -SO
Q2046-01	BC136902-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/14/2025	Chemtech -SO

Date/Time 05-14-25 1340
 Raw Sample Received by: cl sm
 Raw Sample Relinquished by: cl sm

WORKLIST(Hardcopy Internal Chain)

87135767

WorkList Name : %1-051425 WorkList ID : 189504 Department : Wet-Chemistry Date : 05-14-2025 09:06:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2046-02	BC136902-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/14/2025	Chemtech -SO
Q2047-01	SU-04-051425	Solid	Percent Solids	Cool 4 deg C	PSEG05	L31	05/14/2025	Chemtech -SO
Q2047-02	SU-04-051425-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L31	05/14/2025	Chemtech -SO
Q2048-01	L2-WC-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-02	L2-WC-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-03	L2-WC-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-05	L2-WC-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-06	L2-WC-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-07	L2-WC-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-09	L2-WC-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-10	L2-WC-3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-11	L2-WC-3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-13	L2-WC-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-14	L2-WC-4-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO
Q2048-15	L2-WC-4-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/14/2025	Chemtech -SO

Date/Time 05-14-25 15:10
Raw Sample Received by: SP WOC
Raw Sample Relinquished by: CL 5m

Date/Time 05-14-25 17:30
Raw Sample Received by: SP WOC
Raw Sample Relinquished by: SP WOC



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: CDM SMITH
ADDRESS: 110 FIELDCREST AVE #8 6TH FLOOR
CITY EDISON STATE: NJ ZIP: 08837
ATTENTION: MARCIE ENCINAS
PHONE: 732-590-4679 FAX: 732-225-7851

CLIENT PROJECT INFORMATION

PROJECT NAME: SOUTH RIVER WM REPLACEMENT
PROJECT NO.: 302781 LOCATION: SOUTH RIVER, NJ
PROJECT MANAGER: MARCIE ENCINAS
e-mail: ENCINASMA@CDMSMITH.COM
PHONE: 732-590-4679 FAX: 732-225-7851

CLIENT BILLING INFORMATION

BILL TO: CDM SMITH PO#:
ADDRESS: 110 FIELDCREST AVE #8 6TH FLOOR
CITY EDISON STATE: NJ ZIP: 08837
ATTENTION: MARCIE ENCINAS PHONE: 732-590-4679

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☒ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT

1. 2. 3. 4. 5. 6. 7. 8. 9.
TCL VOC
TCL SVOC
METALS
PCB
PESTICIDES
HERBICIDES
PAH/CRO
FULL TCLP
RCRA CHARAL

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	TP-11	SOIL		X	5/13/25	0830	6	X	X	X	X	X	X	X			E
2.	TP-29	SOIL		X	5/13/25	1010	6	X	X	X	X	X	X	X			E
3.	TP-29-99	SOIL		X	5/13/25	1010	6	X	X	X	X	X	X	X			E
4.	TP-24	SOIL		X	5/13/25	1130	18	X	X	X	X	X	X	X			E; MS/MSD
5.	TP-37	SOIL		X	5/13/25	1250	6	X	X	X	X	X	X	X			E
6.	TP-32	SOIL		X	5/13/25	1330	6	X	X	X	X	X	X	X			E
7.	COMP-1	SOIL	X		5/13/25	1405	4								X	X	E
8.	FB-05132025	ADUERS		X	5/13/25	1445	10	X	X	X	X	X	X	X			E, A, B
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. [Signature]	DATE/TIME: 5/13/25 1505	RECEIVED BY: [Signature] 1505	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP. 3.2 °C
RELINQUISHED BY SAMPLER: 2. [Signature]	DATE/TIME:	RECEIVED BY:	Comments:
RELINQUISHED BY SAMPLER: 3. [Signature]	DATE/TIME: 5-13-25	RECEIVED BY:	

Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2032	CAMP02	Order Date : 5/13/2025 4:01:00 PM	Project Mgr :
Client Name : CDM Smith		Project Name : South River WM Replacem	Report Type : Level 1
Client Contact : Marcie Ann Encinas		Receive DateTime : 5/13/2025 4:20:00 PM	EDD Type : EXCEL NOCLEANUP
Invoice Name : CDM Smith		Purchase Order :	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2032-01	TP-11	Solid	05/13/2025	08:30	VOC-TCLVOA-10		8260D		10 Bus. Days
Q2032-02	TP-29	Solid	05/13/2025	10:10	VOC-TCLVOA-10		8260D		10 Bus. Days
Q2032-03	TP-29-99	Solid	05/13/2025	10:10	VOC-TCLVOA-10		8260D		10 Bus. Days
Q2032-04	TP-24	Solid	05/13/2025	11:30	VOC-TCLVOA-10		8260D		10 Bus. Days
Q2032-05	Q2032-04MS	Solid	05/13/2025	11:30	VOC-TCLVOA-10		8260D		10 Bus. Days
Q2032-06	Q2032-04MSD	Solid	05/13/2025	11:30	VOC-TCLVOA-10		8260D		10 Bus. Days
Q2032-07	TP-37	Solid	05/13/2025	12:50	VOC-TCLVOA-10		8260D		10 Bus. Days
Q2032-08	TP-32	Solid	05/13/2025	13:30					

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2032	CAMP02	Order Date : 5/13/2025 4:01:00 PM	Project Mgr :
Client Name : CDM Smith		Project Name : South River WM Replacem	Report Type : Level 1
Client Contact : Marcie Ann Encinas		Receive Date/Time : 5/13/2025 4:20:00 PM	EDD Type : EXCEL NOCLEANUP
Invoice Name : CDM Smith		Purchase Order :	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2032-11	FB-05132025	Water	05/13/2025	14:45	VOC-TCLVOA-10		8260D		10 Bus. Days
Q2032-12	TB	Water	05/13/2025	00:00	VOC-TCLVOA-10		8260-Low		10 Bus. Days
					VOC-TCLVOA-10		8260-Low		10 Bus. Days

Relinquished By :

Date / Time : 5/14/25 0940

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room

9:50 Rg # 4
Rg # 6
B22