

Initial and Continuing Calibration Blank Summary

Client: CDM Smith

SDG No.: Q2032

Project: South River WM Replacement

RunNo.: LB135800

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	05/16/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	05/16/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.001	0.0025	J	0.00096	0.005	05/16/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	05/16/2025
Sample ID: CCB4 Reactive Cyanide	mg/L	0.001	0.0025	J	0.00096	0.005	05/16/2025

Initial and Continuing Calibration Blank Summary

Client: CDM Smith

SDG No.: Q2032

Project: South River WM Replacement

RunNo.: LB135800

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------