


LB135784

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.002	-0.002	-41	0.000	-41					4/-1)
0.2	0.200	0.211	0.011	1082	0.0 %	1082					5
2.5	2.500	2.417	-0.083	12710	0.0 %	12710					3
5.0	5.000	5.043	0.043	26558	0.0 %	26558					1
7.5	7.500	7.646	0.146	40281	0.0 %	40281					2
10.0	10.000	9.886	-0.114	52090	0.0 %	52090					1

LB135784 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 May 2025 12:34:44

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000	7471B	PPB	15 May 2025 12:47:50	S	Std
0.2 - 1	50.2	1082	-	0.2000	7471B	PPB	15 May 2025 12:50:08	S	Std
2.5 - 1	52.5	12710	-	2.5000	7471B	PPB	15 May 2025 12:52:25	S	Std
5.0 - 1	55	26558	-	5.0000	7471B	PPB	15 May 2025 12:54:41	S	Std
7.5 - 1	57.5	40281	-	7.5000	7471B	PPB	15 May 2025 12:57:00	S	Std
10.0 - 1	59	52090	-	10.0000	7471B	PPB	15 May 2025 13:03:13	S	Std
ICV09 - 1	ICV09	19870	3.7745	-	7471B	PPB	15 May 2025 13:07:45	U	SMPL
ICB09 - 1	ICB09	-396	-0.0692	-	7471B	PPB	15 May 2025 13:09:59	U	SMPL
CCV29 - 1	CCV29	26987	5.1244	-	7471B	PPB	15 May 2025 13:22:18	U	SMPL
CCB29 - 1	CCB29	-469	-0.0831	-	7471B	PPB	15 May 2025 13:24:35	U	SMPL
CRA - 1	CRA	1023	0.1999	-	7471B	PPB	15 May 2025 13:26:56	U	SMPL
HighStd - 1	HighStd	50256	9.5377	-	7471B	PPB	15 May 2025 13:29:11	U	SMPL
ChkStd - 1	ChkStd	35831	6.8018	-	7471B	PPB	15 May 2025 13:31:26	U	SMPL
PB168020BL - 1	PB168020BL	-555	-0.0994	-	7471B	PPB	15 May 2025 13:33:54	U	SMPL
Q2019-01 - 1	MH-K	261	0.0554	-	7471B	PPB	15 May 2025 13:38:34	U	SMPL
Q2020-01 - 1	TP-A	15480	2.9419	-	7471B	PPB	15 May 2025 13:40:51	U	SMPL
Q2022-01 - 1	336	1662	0.3211	-	7471B	PPB	15 May 2025 13:43:08	U	SMPL
Q2022-02 - 1	COMP-1	1066	0.2081	-	7471B	PPB	15 May 2025 13:45:29	U	SMPL
Q2022-04 - 1	COMP-2	508	0.1022	-	7471B	PPB	15 May 2025 13:47:46	U	SMPL
CCV30 - 1	CCV30	25247	4.7943	-	7471B	PPB	15 May 2025 13:52:38	U	SMPL
CCB30 - 1	CCB30	-260	-0.0434	-	7471B	PPB	15 May 2025 13:54:54	U	SMPL
Q2023-01 - 1	HEATER-PAD	856	0.1682	-	7471B	PPB	15 May 2025 13:57:12	U	SMPL
Q2024-01 - 1	PL-02-051325	781	0.1540	-	7471B	PPB	15 May 2025 13:59:27	U	SMPL
Q2032-01 - 1	TP-11	1972	0.3799	-	7471B	PPB	15 May 2025 14:01:43	U	SMPL
Q2032-02 - 1	TP-29	1338	0.2597	-	7471B	PPB	15 May 2025 14:03:58	U	SMPL
Q2032-03 - 1	TP-29-99	9349	1.7791	-	7471B	PPB	15 May 2025 14:06:16	U	SMPL
Q2032-04 - 1	TP-24	1966	0.3788	-	7471B	PPB	15 May 2025 14:08:33	U	SMPL
Q2032-04DUP - 1	TP-24DUP	2297	0.4415	-	7471B	PPB	15 May 2025 14:10:54	U	SMPL
Q2032-05 - 1	Q2032-04MS	20485	3.8912	-	7471B	PPB	15 May 2025 14:15:34	U	SMPL
Q2032-06 - 1	Q2032-04MSD	20208	3.8386	-	7471B	PPB	15 May 2025 14:17:52	U	SMPL
Q2032-07 - 1	TP-37	1253	0.2435	-	7471B	PPB	15 May 2025 14:20:11	U	SMPL
CCV31 - 1	CCV31	26855	5.0993	-	7471B	PPB	15 May 2025 14:25:30	U	SMPL
CCB31 - 1	CCB31	-361	-0.0626	-	7471B	PPB	15 May 2025 14:27:47	U	SMPL
Q2032-08 - 1	TP-32	4512	0.8617	-	7471B	PPB	15 May 2025 14:30:06	U	SMPL
PB168021BL - 1	PB168021BL	-221	-0.0360	-	7471B	PPB	15 May 2025 14:32:23	U	SMPL
PB168021BS - 1	PB168021BS	21725	4.1263	-	7471B	PPB	15 May 2025 14:40:08	U	SMPL
Q2034-01 - 1	L3-WC-1	2498	0.4797	-	7471B	PPB	15 May 2025 14:42:25	U	SMPL
Q2034-01DUP - 1	L3-WC-1DUP	2262	0.4349	-	7471B	PPB	15 May 2025 14:44:46	U	SMPL
Q2034-01MS - 1	L3-WC-1MS	23294	4.4239	-	7471B	PPB	15 May 2025 14:47:07	U	SMPL
Q2034-01MSD - 1	L3-WC-1MSD	24646	4.6804	-	7471B	PPB	15 May 2025 14:49:27	U	SMPL
Q2034-05 - 1	L3-WC-2	7584	1.4443	-	7471B	PPB	15 May 2025 14:51:49	U	SMPL
Q2034-09 - 1	L3-WC-3	6091	1.1611	-	7471B	PPB	15 May 2025 14:54:12	U	SMPL
Q2034-13 - 1	L3-WC-4	3320	0.6356	-	7471B	PPB	15 May 2025 14:56:31	U	SMPL
CCV32 - 1	CCV32	26155	4.9666	-	7471B	PPB	15 May 2025 14:58:51	U	SMPL
CCB32 - 1	CCB32	-355	-0.0614	-	7471B	PPB	15 May 2025 15:01:10	U	SMPL
Q2034-17 - 1	L3-WC-5	9939	1.8910	-	7471B	PPB	15 May 2025 15:03:29	U	SMPL
Q2034-21 - 1	L3-WC-6	6880	1.3108	-	7471B	PPB	15 May 2025 15:05:47	U	SMPL
Q2038-01 - 1	72-11991	1060	0.2069	-	7471B	PPB	15 May 2025 15:08:06	U	SMPL
Q2039-01 - 1	TR-05-05142025	907	0.1779	-	7471B	PPB	15 May 2025 15:10:25	U	SMPL
Q2043-01 - 1	281	6474	1.2338	-	7471B	PPB	15 May 2025 15:12:44	U	SMPL
Q2043-03 - 1	72-12000	251	0.0535	-	7471B	PPB	15 May 2025 15:15:01	U	SMPL
Q2047-01 - 1	SU-04-051425	1847	0.3562	-	7471B	PPB	15 May 2025 15:17:23	U	SMPL
Q2048-01 - 1	L2-WC-1	1869	0.3604	-	7471B	PPB	15 May 2025 15:19:41	U	SMPL
Q2048-05 - 1	L2-WC-2	7358	1.4014	-	7471B	PPB	15 May 2025 15:21:59	U	SMPL
Q2048-09 - 1	L2-WC-3	10022	1.9067	-	7471B	PPB	15 May 2025 15:24:17	U	SMPL
CCV33 - 1	CCV33	26729	5.0754	-	7471B	PPB	15 May 2025 15:26:35	U	SMPL
CCB33 - 1	CCB33	-296	-0.0502	-	7471B	PPB	15 May 2025 15:28:57	U	SMPL
Q2048-13 - 1	L2-WC-4	2404	0.4618	-	7471B	PPB	15 May 2025 15:31:18	U	SMPL
PB168020BS - 1	LCSS	24124	4.5814	-	7471B	PPB	15 May 2025 15:36:26	U	SMPL
Q2032-04LX5 - 1		13	0.0084	-	7471B	PPB	15 May 2025 15:38:42	U	SMPL
Q2032-04A - 1		23283	4.4218	-	7471B	PPB	15 May 2025 15:41:02	U	SMPL
Q2034-01LX5 - 1		258	0.0548	-	7471B	PPB	15 May 2025 15:43:19	U	SMPL
Q2034-01A - 1		25089	4.7644	-	7471B	PPB	15 May 2025 15:45:40	U	SMPL
CCV34 - 1	CCV34	26409	5.0147	-	7471B	PPB	15 May 2025 15:48:00	U	SMPL
CCB34 - 1	CCB34	-382	-0.0666	-	7471B	PPB	15 May 2025 15:50:24	U	SMPL