


LB135815

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.0443e-004

C= -4.7560e-002

Rho= 0.9993153

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.053	-0.053	-28	0.000	-28	0			
0.2	0.200	0.177	-0.023	1098	0.0 %	1098				
2.5	2.500	2.561	0.061	12760	0.0 %	12760				
5.0	5.000	4.955	-0.045	24473	0.0 %	24473				
7.5	7.500	7.755	0.255	38166	0.0 %	38166				
10.0	10.000	9.805	-0.195	48197	0.0 %	48197				

4/10
-12
2
-1
3
-2

LB135815 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 May 2025 09:26:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	-28	-	0.0000	7470A	PPB	19 May 2025 10:02:41	S	Std
0.2 - 1	SC-2	1098	-	0.2000	7470A	PPB	19 May 2025 10:04:58	S	Std
2.5 - 1	SC-2.5	12760	-	2.5000	7470A	PPB	19 May 2025 10:07:15	S	Std
5.0 - 1	SC-5	24473	-	5.0000	7470A	PPB	19 May 2025 10:09:33	S	Std
7.5 - 1	SC-7.5	38166	-	7.5000	7470A	PPB	19 May 2025 10:11:49	S	Std
10.0 - 1	SC-10	48197	-	10.0000	7470A	PPB	19 May 2025 10:14:07	S	Std
ICV10 - 1	ICV10	18225	3.6782	-	7470A	PPB	19 May 2025 10:17:14	U	SMPL
ICB10 - 1	ICB10	-161	-0.0805	-	7470A	PPB	19 May 2025 10:19:28	U	SMPL
CCV35 - 1	CCV35	23257	4.7068	-	7470A	PPB	19 May 2025 10:21:45	U	SMPL
CCB35 - 1	CCB35	-142	-0.0766	-	7470A	PPB	19 May 2025 10:24:00	U	SMPL
CRA - 1	CRA	1028	0.1626	-	7470A	PPB	19 May 2025 10:26:17	U	SMPL
HighStd - 1	HighStd	48087	9.7828	-	7470A	PPB	19 May 2025 10:28:33	U	SMPL
ChkStd - 1	ChkStd	34522	7.0097	-	7470A	PPB	19 May 2025 10:30:48	U	SMPL
PB168028BL - 1	PBW	80	-0.0312	-	7470A	PPB	19 May 2025 10:35:47	U	SMPL
PB168028BS - 1	LCSW	18339	3.7015	-	7470A	PPB	19 May 2025 10:42:57	U	SMPL
Q2019-04 - 1	MH-K	-113	-0.0707	-	7470A	PPB	19 May 2025 10:45:14	U	SMPL
Q2020-04 - 1	TP-A	70	-0.0333	-	7470A	PPB	19 May 2025 10:47:33	U	SMPL
Q2023-02 - 1	HEATER-PAD	76	-0.0320	-	7470A	PPB	19 May 2025 10:49:51	U	SMPL
Q2027-03 - 1	B27-SOIL-SAMPLE	95	-0.0281	-	7470A	PPB	19 May 2025 10:52:08	U	SMPL
Q2027-04 - 1	B28-SOIL-SAMPLE	46	-0.0382	-	7470A	PPB	19 May 2025 10:54:24	U	SMPL
CCV36 - 1	CCV36	22667	4.5862	-	7470A	PPB	19 May 2025 10:56:39	U	SMPL
CCB36 - 1	CCB36	-156	-0.0795	-	7470A	PPB	19 May 2025 10:58:54	U	SMPL
Q2032-09 - 1	COMP-1	-58	-0.0594	-	7470A	PPB	19 May 2025 11:01:11	U	SMPL
Q2034-04 - 1	L3-WC-1	-4	-0.0484	-	7470A	PPB	19 May 2025 11:03:27	U	SMPL
Q2034-08 - 1	L3-WC-2	2	-0.0472	-	7470A	PPB	19 May 2025 11:05:42	U	SMPL
Q2034-12 - 1	L3-WC-3	72	-0.0328	-	7470A	PPB	19 May 2025 11:07:57	U	SMPL
Q2034-16 - 1	L3-WC-4	129	-0.0212	-	7470A	PPB	19 May 2025 11:10:13	U	SMPL
Q2034-20 - 1	L3-WC-5	103	-0.0265	-	7470A	PPB	19 May 2025 11:12:30	U	SMPL
Q2034-24 - 1	L3-WC-6	107	-0.0257	-	7470A	PPB	19 May 2025 11:14:46	U	SMPL
Q2038-02 - 1	72-11991	-5	-0.0486	-	7470A	PPB	19 May 2025 11:17:04	U	SMPL
Q2048-04 - 1	L2-WC-1	-15	-0.0506	-	7470A	PPB	19 May 2025 11:19:21	U	SMPL
Q2048-08 - 1	L2-WC-2	-15	-0.0506	-	7470A	PPB	19 May 2025 11:21:36	U	SMPL
CCV37 - 1	CCV37	22769	4.6071	-	7470A	PPB	19 May 2025 11:23:52	U	SMPL
CCB37 - 1	CCB37	126	-0.0218	-	7470A	PPB	19 May 2025 11:28:51	U	SMPL
Q2048-12 - 1	L2-WC-3	109	-0.0253	-	7470A	PPB	19 May 2025 11:31:06	U	SMPL
Q2048-16 - 1	L2-WC-4	131	-0.0208	-	7470A	PPB	19 May 2025 11:33:22	U	SMPL
Q2048-16DUP - 1	L2-WC-4DUP	84	-0.0304	-	7470A	PPB	19 May 2025 11:35:37	U	SMPL
Q2048-16MS - 1	L2-WC-4MS	16334	3.2916	-	7470A	PPB	19 May 2025 11:37:53	U	SMPL
Q2048-16MSD - 1	L2-WC-4MSD	17059	3.4398	-	7470A	PPB	19 May 2025 11:40:08	U	SMPL
PB168031BL - 1	PBW	-39	-0.0555	-	7470A	PPB	19 May 2025 11:45:22	U	SMPL
PB168031BS - 1	LCSW	20045	4.0502	-	7470A	PPB	19 May 2025 11:47:38	U	SMPL
Q2032-11 - 1	FB-05132025	-243	-0.0972	-	7470A	PPB	19 May 2025 11:49:56	U	SMPL
Q2043-05 - 1	3139	2151	0.3922	-	7470A	PPB	19 May 2025 11:52:15	U	SMPL
Q2044-02 - 1	3859	-54	-0.0586	-	7470A	PPB	19 May 2025 11:54:30	U	SMPL
CCV38 - 1	CCV38	24306	4.9213	-	7470A	PPB	19 May 2025 12:06:43	U	SMPL
CCB38 - 1	CCB38	-44	-0.0566	-	7470A	PPB	19 May 2025 12:11:43	U	SMPL
Q2044-02DUP - 1	3859DUP	317	0.0172	-	7470A	PPB	19 May 2025 12:13:58	U	SMPL
Q2044-02MS - 1	3859MS	9866	1.9693	-	7470A	PPB	19 May 2025 12:16:14	U	SMPL
Q2044-02MSD - 1	3859MSD	13693	2.7517	-	7470A	PPB	19 May 2025 12:18:29	U	SMPL
PB167994TB - 1	PB167994TB	-107	-0.0694	-	7470A	PPB	19 May 2025 12:20:48	U	SMPL
Q2048-16LX5 - 1		187	-0.0093	-	7470A	PPB	19 May 2025 12:23:05	U	SMPL
Q2048-16A - 1		17710	3.5729	-	7470A	PPB	19 May 2025 12:25:21	U	SMPL
Q2044-02LX5 - 1		325	0.0189	-	7470A	PPB	19 May 2025 12:27:37	U	SMPL
Q2044-02A - 1		18280	3.6894	-	7470A	PPB	19 May 2025 12:29:56	U	SMPL
CCV39 - 1	CCV39	23451	4.7465	-	7470A	PPB	19 May 2025 12:37:34	U	SMPL
CCB39 - 1	CCB39	-233	-0.0952	-	7470A	PPB	19 May 2025 12:39:51	U	SMPL