



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: CDM Smith **SDG No.:** Q2032
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2032 **SAS No.:** Q2032
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.16	0.2	81	70 - 130	CV	05/19/2025	10:26	LB135815
CRI01	Arsenic	19.7	20.0	99	65 - 135	P	05/19/2025	12:48	LB135823
	Barium	96.2	100	96	65 - 135	P	05/19/2025	12:48	LB135823
	Cadmium	6.33	6.0	106	65 - 135	P	05/19/2025	12:48	LB135823
	Chromium	10.1	10.0	101	65 - 135	P	05/19/2025	12:48	LB135823
	Lead	12.8	12.0	107	65 - 135	P	05/19/2025	12:48	LB135823
	Selenium	23.9	20.0	119	65 - 135	P	05/19/2025	12:48	LB135823
	Silver	10.3	10.0	103	65 - 135	P	05/19/2025	12:48	LB135823