

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	<u>CDM Smith</u>	SDG No.:	<u>Q2032</u>
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>
ICS Source:	<u>EPA</u>	Case No.:	<u>Q2032</u>
		SAS No.:	<u>Q2032</u>
		Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	5.21			-20	20	05/19/2025	12:52	LB135823
	Barium	4.26	6.0	71	-94	106	05/19/2025	12:52	LB135823
	Cadmium	-4.22	1.0	422	-5	7	05/19/2025	12:52	LB135823
	Chromium	55.5	52.0	107	42	62	05/19/2025	12:52	LB135823
	Lead	-0.59			-12	12	05/19/2025	12:52	LB135823
	Selenium	4.52			-20	20	05/19/2025	12:52	LB135823
	Silver	-0.75			-10	10	05/19/2025	12:52	LB135823
ICSAB01	Arsenic	115	104	111	88.4	120	05/19/2025	12:57	LB135823
	Barium	481	537	90	437	637	05/19/2025	12:57	LB135823
	Cadmium	1030	972	106	826	1120	05/19/2025	12:57	LB135823
	Chromium	571	542	105	460	624	05/19/2025	12:57	LB135823
	Lead	49.5	49.0	101	37	61	05/19/2025	12:57	LB135823
	Selenium	58.0	46.0	126	26	66	05/19/2025	12:57	LB135823
	Silver	213	201	106	170	232	05/19/2025	12:57	LB135823