


Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2032
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2032 **SAS No.:** Q2032

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168025BS							
Arsenic	ug/L	4000	3830		96	80 - 120	P
Barium	ug/L	1000	906		91	80 - 120	P
Cadmium	ug/L	1000	962		96	80 - 120	P
Chromium	ug/L	2000	1920		96	80 - 120	P
Lead	ug/L	5000	4700		94	80 - 120	P
Selenium	ug/L	10000	9740		97	80 - 120	P
Silver	ug/L	380	361		95	80 - 120	P



- 7 -

Client:	<u>CDM Smith</u>	SDG No.:	<u>Q2032</u>				
Contract:	CAMP02	Lab Code:	CHEM	Case No.:	Q2032	SAS No.:	Q2032

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168028BS Mercury	ug/L	4.0	3.70		92	80 - 120	CV