

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CAMP02
 ProjectID: South River WM Replacement
 Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG No.: Q2032

Calibration Sequence : FF042225		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	116059922	116060	FF015786.D
500	58079559	116159	FF015787.D
200	21235975	106180	FF015788.D
100	11342548	113425	FF015789.D
50	7274526	145491	FF015790.D
AVG RF : 119463		% RSD : 12.646	AVG RT : 15.02