

**DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY**

Lab Name: Chemtech Contract: CAMP02  
 ProjectID: South River WM Replacement  
 Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG No.: Q2032

| Calibration Sequence : FG042425 |            | Test : Diesel Range Organics |                  |
|---------------------------------|------------|------------------------------|------------------|
| Concentration (PPM)             | Area Count | Reference Factor             | File ID          |
| 1000                            | 122641169  | 122641                       | FG015756.D       |
| 500                             | 64139521   | 128279                       | FG015757.D       |
| 200                             | 25210755   | 126054                       | FG015758.D       |
| 100                             | 13317775   | 133178                       | FG015759.D       |
| 50                              | 6223650    | 124473                       | FG015760.D       |
| AVG RF : 126925                 |            | % RSD : 3.202                | AVG RT : 15.0012 |