

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CAMP02
ProjectID: South River WM Replacement
Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG No.: Q2032
DataFile: FF015870.D Analyst Name: YP\AJ Analyst Date: 05-19-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	64441382	128883	119463	7.885