

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Chemtech **Client:** CDM Smith
Lab Code: CHEM **Cas No:** Q2032 **SAS No :** Q2032 **SDG No:** Q2032
Matrix Spike - EPA Sample No : PB168064BSD **Datafile:** FF015867.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
DRO	200	0	196	98	78-117

LCS/LCSD % Recovery RPD : 3.0