

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS030202.D Time Analyzed: 22:42

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 9.132  | 9.033     | 9.233  | 701.260            | 712.500           | -1.6 |
| 2,4,5-TP(Silvex) | 8.850  | 8.750     | 8.950  | 699.420            | 712.500           | -1.8 |
| 2,4-D            | 8.010  | 7.910     | 8.110  | 682.710            | 705.000           | -3.2 |
| 2,4-DB           | 9.690  | 9.589     | 9.789  | 733.930            | 712.500           | 3.0  |
| 2,4-DCAA         | 6.933  | 6.834     | 7.034  | 725.760            | 750.000           | -3.2 |
| DICAMBA          | 7.110  | 7.010     | 7.210  | 691.030            | 705.000           | -2.0 |
| DICHLORPROP      | 7.789  | 7.689     | 7.889  | 680.280            | 705.000           | -3.5 |
| Dinoseb          | 10.850 | 10.751    | 10.951 | 683.670            | 705.000           | -3.0 |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS030202.D Time Analyzed: 22:42

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 9.925  | 9.826     | 10.026 | 705.510            | 712.500           | -1.0 |
| 2,4,5-TP(Silvex) | 9.521  | 9.423     | 9.623  | 710.430            | 712.500           | -0.3 |
| 2,4-D            | 8.643  | 8.544     | 8.744  | 702.650            | 705.000           | -0.3 |
| 2,4-DB           | 10.483 | 10.385    | 10.585 | 666.360            | 712.500           | -6.5 |
| 2,4-DCAA         | 7.453  | 7.354     | 7.554  | 737.130            | 750.000           | -1.7 |
| DICAMBA          | 7.638  | 7.539     | 7.739  | 706.190            | 705.000           | 0.2  |
| DICHLORPROP      | 8.330  | 8.232     | 8.432  | 680.660            | 705.000           | -3.5 |
| Dinoseb          | 10.856 | 10.758    | 10.958 | 689.570            | 705.000           | -2.2 |