

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030214.D Time Analyzed: 05:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.132	9.033	9.233	710.470	712.500	-0.3
2,4,5-TP(Silvex)	8.850	8.750	8.950	709.200	712.500	-0.5
2,4-D	8.009	7.910	8.110	691.270	705.000	-1.9
2,4-DB	9.690	9.589	9.789	743.140	712.500	4.3
2,4-DCAA	6.933	6.834	7.034	737.200	750.000	-1.7
DICAMBA	7.109	7.010	7.210	702.480	705.000	-0.4
DICHLORPROP	7.789	7.689	7.889	691.550	705.000	-1.9
Dinoseb	10.850	10.751	10.951	692.450	705.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030214.D Time Analyzed: 05:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.925	9.826	10.026	712.150	712.500	0.0
2,4,5-TP(Silvex)	9.521	9.423	9.623	719.650	712.500	1.0
2,4-D	8.643	8.544	8.744	710.310	705.000	0.8
2,4-DB	10.483	10.385	10.585	636.260	712.500	-10.7
2,4-DCAA	7.452	7.354	7.554	745.760	750.000	-0.6
DICAMBA	7.638	7.539	7.739	715.570	705.000	1.5
DICHLORPROP	8.330	8.232	8.432	691.480	705.000	-1.9
Dinoseb	10.856	10.758	10.958	698.230	705.000	-1.0