

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030259.D Time Analyzed: 09:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.131	9.033	9.233	741.330	712.500	4.0
2,4,5-TP(Silvex)	8.849	8.750	8.950	741.520	712.500	4.1
2,4-D	8.008	7.910	8.110	721.550	705.000	2.3
2,4-DB	9.688	9.589	9.789	790.740	712.500	11.0
2,4-DCAA	6.932	6.834	7.034	764.820	750.000	2.0
DICAMBA	7.108	7.010	7.210	731.410	705.000	3.7
DICHLORPROP	7.788	7.689	7.889	718.270	705.000	1.9
Dinoseb	10.849	10.751	10.951	725.310	705.000	2.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030259.D Time Analyzed: 09:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.922	9.826	10.026	721.940	712.500	1.3
2,4,5-TP(Silvex)	9.519	9.423	9.623	731.420	712.500	2.7
2,4-D	8.641	8.544	8.744	711.450	705.000	0.9
2,4-DB	10.480	10.385	10.585	634.210	712.500	-11.0
2,4-DCAA	7.451	7.354	7.554	742.740	750.000	-1.0
DICAMBA	7.637	7.539	7.739	715.420	705.000	1.5
DICHLORPROP	8.329	8.232	8.432	691.050	705.000	-2.0
Dinoseb	10.854	10.758	10.958	713.860	705.000	1.3