

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030264.D Time Analyzed: 11:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.130	9.033	9.233	733.740	712.500	3.0
2,4,5-TP(Silvex)	8.849	8.750	8.950	731.940	712.500	2.7
2,4-D	8.008	7.910	8.110	715.780	705.000	1.5
2,4-DB	9.688	9.589	9.789	785.740	712.500	10.3
2,4-DCAA	6.932	6.834	7.034	756.570	750.000	0.9
DICAMBA	7.108	7.010	7.210	721.170	705.000	2.3
DICHLORPROP	7.788	7.689	7.889	707.740	705.000	0.4
Dinoseb	10.849	10.751	10.951	714.260	705.000	1.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030264.D Time Analyzed: 11:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.922	9.826	10.026	721.700	712.500	1.3
2,4,5-TP(Silvex)	9.518	9.423	9.623	727.120	712.500	2.1
2,4-D	8.641	8.544	8.744	705.740	705.000	0.1
2,4-DB	10.480	10.385	10.585	661.390	712.500	-7.2
2,4-DCAA	7.451	7.354	7.554	737.310	750.000	-1.7
DICAMBA	7.636	7.539	7.739	715.630	705.000	1.5
DICHLORPROP	8.329	8.232	8.432	687.950	705.000	-2.4
Dinoseb	10.854	10.758	10.958	711.010	705.000	0.9