

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL07 Date Analyzed: 05/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS030300.D Time Analyzed: 02:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.128	9.033	9.233	747.950	712.500	5.0
2,4,5-TP(Silvex)	8.847	8.750	8.950	740.100	712.500	3.9
2,4-D	8.006	7.910	8.110	727.670	705.000	3.2
2,4-DB	9.686	9.589	9.789	797.900	712.500	12.0
2,4-DCAA	6.930	6.834	7.034	766.930	750.000	2.3
DICAMBA	7.107	7.010	7.210	731.780	705.000	3.8
DICHLORPROP	7.786	7.689	7.889	718.720	705.000	1.9
Dinoseb	10.846	10.751	10.951	723.570	705.000	2.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL07 Date Analyzed: 05/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS030300.D Time Analyzed: 02:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.920	9.826	10.026	738.440	712.500	3.6
2,4,5-TP(Silvex)	9.516	9.423	9.623	744.460	712.500	4.5
2,4-D	8.638	8.544	8.744	726.120	705.000	3.0
2,4-DB	10.478	10.385	10.585	649.170	712.500	-8.9
2,4-DCAA	7.449	7.354	7.554	755.260	750.000	0.7
DICAMBA	7.635	7.539	7.739	731.940	705.000	3.8
DICHLORPROP	8.326	8.232	8.432	705.910	705.000	0.1
Dinoseb	10.852	10.758	10.958	722.680	705.000	2.5