

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2032</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	Datafile :	<u>PS030210.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168003BSD (Column 1)	DICAMBA	5	4.80	ug/L	96	2			67	136	20
	DICHLORPROP	5	4.70	ug/L	94	2			88	119	20
	2,4-D	5	4.70	ug/L	94	2			83	130	20
	2,4,5-TP(Silvex)	5	4.90	ug/L	98	2			78	127	20
	2,4,5-T	5	4.90	ug/L	98	2			74	129	20
	2,4-DB	5	5.00	ug/L	100	2			53	149	20
	Dinoseb	5	4.70	ug/L	94	2			72	131	20
	DICAMBA	5	4.70	ug/L	94	2			67	136	20
PB168003BSD (Column 2)	DICHLORPROP	5	4.70	ug/L	94	2			88	119	20
	2,4-D	5	4.80	ug/L	96	2			83	130	20
	2,4,5-TP(Silvex)	5	4.80	ug/L	96	2			78	127	20
	2,4,5-T	5	4.80	ug/L	96	2			74	129	20
	2,4-DB	5	4.50	ug/L	90	2			53	149	20
	Dinoseb	5	4.60	ug/L	92	2			72	131	20