

**SW-846**

Datafile : PS030261.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB168030BS (Column 1)	DICAMBA	166.5	172	ug/Kg	103				72	129
	DICHLORPROP	166.5	171	ug/Kg	103				77	135
	2,4-D	166.5	171	ug/Kg	103				65	144
	2,4,5-TP(Silvex)	166.5	176	ug/Kg	106				74	146
	2,4,5-T	166.5	176	ug/Kg	106				77	134
	2,4-DB	166.5	184	ug/Kg	111				72	122
	Dinoseb	166.5	171	ug/Kg	103				74	132
PB168030BS (Column 2)	DICAMBA	166.5	164	ug/Kg	98				72	129
	DICHLORPROP	166.5	165	ug/Kg	99				77	135
	2,4-D	166.5	167	ug/Kg	100				65	144
	2,4,5-TP(Silvex)	166.5	171	ug/Kg	103				74	146
	2,4,5-T	166.5	169	ug/Kg	102				77	134
	2,4-DB	166.5	147	ug/Kg	88				72	122
	Dinoseb	166.5	165	ug/Kg	99				74	132