

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

Continuing Calib Date: 05/16/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 09:35 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.77       | 4.77      | 4.67      | 4.87 | 0.00       |
| Aroclor-1016-2       | (2) | 4.79       | 4.79      | 4.69      | 4.89 | 0.00       |
| Aroclor-1016-3       | (3) | 4.85       | 4.85      | 4.75      | 4.95 | 0.00       |
| Aroclor-1016-4       | (4) | 4.97       | 4.97      | 4.87      | 5.07 | 0.00       |
| Aroclor-1016-5       | (5) | 5.23       | 5.23      | 5.13      | 5.33 | 0.01       |
| Aroclor-1260-1       | (1) | 6.27       | 6.27      | 6.17      | 6.37 | 0.00       |
| Aroclor-1260-2       | (2) | 6.46       | 6.46      | 6.36      | 6.56 | 0.01       |
| Aroclor-1260-3       | (3) | 6.82       | 6.82      | 6.72      | 6.92 | 0.00       |
| Aroclor-1260-4       | (4) | 7.08       | 7.08      | 6.98      | 7.18 | 0.00       |
| Aroclor-1260-5       | (5) | 7.32       | 7.33      | 7.23      | 7.43 | 0.01       |
| Tetrachloro-m-xylene |     | 3.68       | 3.68      | 3.58      | 3.78 | 0.00       |
| Decachlorobiphenyl   |     | 8.72       | 8.72      | 8.62      | 8.82 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

Continuing Calib Date: 05/16/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 09:35 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.76       | 4.76      | 4.66                   | 4.86 | 0.00       |
| Aroclor-1016-2       | (2) | 4.78       | 4.78      | 4.68                   | 4.88 | 0.00       |
| Aroclor-1016-3       | (3) | 4.95       | 4.95      | 4.85                   | 5.05 | 0.00       |
| Aroclor-1016-4       | (4) | 4.99       | 4.99      | 4.89                   | 5.09 | 0.00       |
| Aroclor-1016-5       | (5) | 5.21       | 5.21      | 5.11                   | 5.31 | 0.00       |
| Aroclor-1260-1       | (1) | 6.24       | 6.24      | 6.14                   | 6.34 | 0.00       |
| Aroclor-1260-2       | (2) | 6.43       | 6.43      | 6.33                   | 6.53 | 0.00       |
| Aroclor-1260-3       | (3) | 6.58       | 6.58      | 6.48                   | 6.68 | 0.00       |
| Aroclor-1260-4       | (4) | 7.05       | 7.05      | 6.95                   | 7.15 | 0.00       |
| Aroclor-1260-5       | (5) | 7.29       | 7.29      | 7.19                   | 7.39 | 0.00       |
| Tetrachloro-m-xylene |     | 3.68       | 3.68      | 3.58                   | 3.78 | 0.00       |
| Decachlorobiphenyl   |     | 8.67       | 8.67      | 8.57                   | 8.77 | 0.00       |